

Surrogate Summary

SDG No.: **Q1574**

Client: **G Environmental**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO109425.D	PIBLK-PO109425.D	Tetrachloro-m-xylene	1	20	25.3	126		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	25.3	126		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	24.1	121		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	25.5	127		70 (60)	130 (140)
I.BLK-PO109927.D	PIBLK-PO109927.D	Tetrachloro-m-xylene	1	20	20.0	100		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	17.1	86		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.4	102		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	17.4	87		70 (60)	130 (140)
Q1574-01	WC1	Tetrachloro-m-xylene	1	20	18.4	92		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	27.8	139		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.8	99		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	23.6	118		30 (32)	150 (175)
Q1574-01DL	WC1DL	Tetrachloro-m-xylene	1	20	20.4	102		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	32.6	163	*	30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.6	93		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	23.8	119		30 (32)	150 (175)
I.BLK-PO109939.D	PIBLK-PO109939.D	Tetrachloro-m-xylene	1	20	20.4	102		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	17.6	88		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.5	103		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	16.5	83		70 (60)	130 (140)
I.BLK-PP070418.D	PIBLK-PP070418.D	Tetrachloro-m-xylene	1	20	22.0	110		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	23.1	116		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	23.0	115		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	23.5	117		70 (60)	130 (140)
I.BLK-PP070547.D	PIBLK-PP070547.D	Tetrachloro-m-xylene	1	20	19.5	97		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	20.3	101		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.5	102		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.4	102		70 (60)	130 (140)
PB167143BL	PB167143BL	Tetrachloro-m-xylene	1	20	20.4	102		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	21.2	106		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.4	107		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	21.0	105		30 (32)	150 (175)
PB167143BS	PB167143BS	Tetrachloro-m-xylene	1	20	21.4	107		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	21.8	109		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.7	108		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	20.0	100		30 (32)	150 (175)
I.BLK-PP070562.D	PIBLK-PP070562.D	Tetrachloro-m-xylene	1	20	20.1	100		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	20.9	104		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	21.2	106		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.2	101		70 (60)	130 (140)
Q1572-06MS	C0018MS	Tetrachloro-m-xylene	1	20	20.3	101		30 (32)	150 (144)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q1574

Client: G Environmental

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1572-06MS	C0018MS	Decachlorobiphenyl	1	20	21.8	109		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.0	95		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	34.0	170	*	30 (32)	150 (175)
Q1572-06MSD	C0018MSD	Tetrachloro-m-xylene	1	20	20.8	104		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	22.3	112		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.6	98		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	33.2	166	*	30 (32)	150 (175)
I.BLK-PP070572.D	PIBLK-PP070572.D	Tetrachloro-m-xylene	1	20	20.2	101		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	21.0	105		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	21.1	105		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.5	102		70 (60)	130 (140)