

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1574 SAS No.: Q1574 SDG No.: Q1574
Instrument ID: BNA_F Calibration Date(s): 03/10/2025 03/10/2025
Calibration Time(s): 11:01 15:20

LAB FILE ID:		RRF2.5 = BF141897.D			RRF005 = BF141898.D		RRF010 = BF141899.D	
		RRF020 = BF141900.D			RRF040 = BF141901.D		RRF060 = BF141903.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF060	RRF	% RSD
2-Fluorophenol		1.267	1.243	1.298	1.148	1.146	1.198	5.8
Benzaldehyde			1.021	0.975	0.778	0.716	0.853	15.8
Phenol-d6		1.623	1.593	1.644	1.442	1.471	1.526	6.0
Phenol		1.708	1.675	1.747	1.536	1.532	1.605	6.3
bis(2-Chloroethyl) ether		1.287	1.246	1.281	1.144	1.176	1.205	5.4
2-Chlorophenol		1.421	1.354	1.434	1.259	1.284	1.329	5.6
2-Methylphenol		1.080	1.050	1.122	1.008	1.079	1.057	3.7
2,2-oxybis(1-Chloropropane)		1.591	1.525	1.534	1.363	1.472	1.455	7.1
Acetophenone		0.514	0.493	0.513	0.460	0.469	0.479	5.8
3+4-Methylphenols		1.449	1.394	1.470	1.294	1.334	1.354	6.5
n-Nitroso-di-n-propylamine	1.020	1.024	1.000	1.043	0.913	0.991	0.980	5.3
Nitrobenzene-d5		0.345	0.356	0.379	0.348	0.363	0.355	3.5
Hexachloroethane		0.545	0.551	0.576	0.517	0.560	0.544	3.8
Nitrobenzene		0.347	0.351	0.375	0.346	0.359	0.353	3.1
Isophorone		0.647	0.620	0.655	0.600	0.648	0.628	3.5
2-Nitrophenol		0.150	0.159	0.182	0.175	0.188	0.173	7.9
2,4-Dimethylphenol		0.247	0.234	0.251	0.231	0.244	0.239	3.5
bis(2-Chloroethoxy) methane		0.411	0.406	0.416	0.372	0.397	0.394	4.6
2,4-Dichlorophenol		0.292	0.295	0.309	0.286	0.303	0.296	2.5
Naphthalene		1.104	1.075	1.101	0.984	0.973	1.027	6.2
4-Chloroaniline		0.383	0.378	0.393	0.351	0.352	0.363	6.1
Hexachlorobutadiene		0.211	0.211	0.221	0.203	0.214	0.213	2.5
Caprolactam		0.092	0.089	0.094	0.084	0.090	0.090	4.0
4-Chloro-3-methylphenol		0.334	0.331	0.349	0.318	0.334	0.331	3.1
2-Methylnaphthalene		0.746	0.709	0.733	0.653	0.667	0.687	6.1
Hexachlorocyclopentadiene		0.204	0.221	0.249	0.250	0.252	0.238	8.0
2,4,6-Trichlorophenol		0.388	0.389	0.402	0.398	0.401	0.398	1.9
2-Fluorobiphenyl		1.430	1.379	1.379	1.254	1.246	1.315	5.9
2,4,5-Trichlorophenol		0.396	0.391	0.422	0.389	0.415	0.403	3.0
1,1-Biphenyl		1.636	1.573	1.591	1.454	1.431	1.520	5.3
2-Chloronaphthalene		1.197	1.141	1.181	1.091	1.090	1.133	3.8
2-Nitroaniline		0.296	0.314	0.346	0.322	0.332	0.328	6.1
Dimethylphthalate		1.481	1.414	1.469	1.329	1.345	1.401	4.5
Acenaphthylene		1.765	1.740	1.778	1.635	1.595	1.681	4.7

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1574 SAS No.: Q1574 SDG No.: Q1574
Instrument ID: BNA_F Calibration Date(s): 03/10/2025 03/10/2025
Calibration Time(s): 11:01 15:20

LAB FILE ID:		RRF2.5 = BF141897.D			RRF005 = BF141898.D		RRF010 = BF141899.D	
		RRF020 = BF141900.D			RRF040 = BF141901.D		RRF060 = BF141903.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF060	RRF	% RSD
2,6-Dinitrotoluene		0.277	0.279	0.305	0.285	0.296	0.292	4.1
3-Nitroaniline		0.291	0.295	0.316	0.293	0.301	0.296	3.6
Acenaphthene		1.236	1.195	1.231	1.146	1.165	1.181	3.4
2,4-Dinitrophenol			0.085	0.124	0.142	0.161	0.139	22.0
4-Nitrophenol		0.192	0.216	0.234	0.230	0.230	0.223	6.6
Dibenzofuran		1.849	1.787	1.791	1.614	1.596	1.688	6.9
2,4-Dinitrotoluene		0.365	0.382	0.402	0.375	0.383	0.381	2.9
Diethylphthalate		1.475	1.471	1.497	1.336	1.338	1.397	5.8
4-Chlorophenyl-phenylether		0.730	0.688	0.702	0.643	0.656	0.679	4.4
Fluorene		1.443	1.361	1.381	1.229	1.234	1.302	7.0
4-Nitroaniline		0.284	0.288	0.304	0.288	0.286	0.288	2.7
4,6-Dinitro-2-methylphenol			0.083	0.111	0.121	0.133	0.119	16.6
n-Nitrosodiphenylamine		0.675	0.651	0.671	0.620	0.638	0.647	3.0
2,4,6-Tribromophenol		0.234	0.236	0.259	0.250	0.265	0.254	5.7
4-Bromophenyl-phenylether		0.240	0.249	0.255	0.244	0.245	0.251	3.6
Hexachlorobenzene		0.262	0.267	0.284	0.269	0.271	0.275	3.9
Atrazine		0.216	0.213	0.193	0.164		0.198	10.7
Pentachlorophenol		0.130	0.151	0.172	0.177	0.186	0.170	12.7
Phenanthrene		1.155	1.151	1.138	1.038	1.032	1.080	5.9
Anthracene		1.173	1.139	1.139	1.046	1.036	1.084	5.9
Carbazole		0.996	0.994	0.985	0.908	0.892	0.935	5.9
Di-n-butylphthalate		1.320	1.321	1.319	1.177	1.175	1.240	6.5
Fluoranthene		1.234	1.226	1.228	1.108	1.059	1.146	7.7
Pyrene		1.695	1.651	1.727	1.611	1.823	1.730	5.4
Terphenyl-d14		1.343	1.309	1.360	1.276	1.417	1.353	3.6
Butylbenzylphthalate		0.629	0.656	0.702	0.655	0.717	0.677	4.7
3,3-Dichlorobenzidine		0.370	0.389	0.380	0.372	0.394	0.379	2.8
Benzo (a) anthracene		1.370	1.304	1.334	1.273	1.317	1.312	2.4
Chrysene		1.143	1.194	1.260	1.136	1.213	1.190	3.5
Bis (2-ethylhexyl) phthalate		0.897	0.912	0.966	0.904	0.952	0.934	3.4
Di-n-octyl phthalate		1.145	1.219	1.338	1.265	1.390	1.299	7.1
Benzo (b) fluoranthene		1.420	1.310	1.485	1.210	1.390	1.371	6.6
Benzo (k) fluoranthene		1.162	1.268	1.197	1.205	1.210	1.173	6.2
Benzo (a) pyrene		1.040	1.042	1.123	1.036	1.083	1.073	3.2
Indeno (1,2,3-cd) pyrene		1.097	1.142	1.308	1.287	1.472	1.294	10.6

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1574 SAS No.: Q1574 SDG No.: Q1574
Instrument ID: BNA_F Calibration Date(s): 03/10/2025 03/10/2025
Calibration Time(s): 11:01 15:20

LAB FILE ID:		RRF2.5 = BF141897.D		RRF005 = BF141898.D		RRF010 = BF141899.D		
		RRF020 = BF141900.D		RRF040 = BF141901.D		RRF060 = BF141903.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF060	RRF	% RSD
Dibenzo (a,h) anthracene		0.910	0.953	1.073	1.068	1.195	1.067	9.9
Benzo (g,h,i) perylene		0.912	0.949	1.061	1.065	1.167	1.055	9.2
1,2,4,5-Tetrachlorobenzene		0.615	0.604	0.626	0.592	0.606	0.609	2.1
1,4-Dioxane		0.490	0.482	0.536	0.504	0.501	0.502	3.4
2,3,4,6-Tetrachlorophenol		0.355	0.365	0.390	0.359	0.366	0.367	3.0