

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bp031825

SequenceID :	<u>bp031825</u>	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)		_____	_____	_____✓_____
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)		_____	_____	_____✓_____
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series		_____	_____	_____✓_____
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series		_____	_____	_____✓_____
5. GC/MS Calibration Met:		_____	_____	_____✓_____
a. Initial calibration Meet Criteria		_____	_____	_____✓_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.		_____	_____	_____
b. Continuous Calibration(CCC) Meet Criteria		_____	_____✓_____	_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.		_____	_____	_____
Compound #77 is biased high in the CCC, if any samples are found with hit of this compound they will be re-analyzed with a properly passing CCC.		_____	_____	_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:		_____	_____✓_____	_____
a. B/N Fraction		_____	_____	_____
d. Acid Fraction		_____	_____	_____

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

____ ✓ ____

a. B/N Fraction

d. Acid Fraction Recoveries of 2-Fluorophenol and Phenol-d6 failed in the sample Q1583-02 due to matrix interference, as seen in the chromatogram.
This analysis is final.

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

✓ ____ ____

a. B/N Fraction

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

____ ____ ✓

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

____ ____ ✓

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

____ ✓ ____

ADDITIONAL COMMENTS:

Terphenyl-d14 was marginally biased high in PB167157BL. The data will be used for the hard copies.

Jagrut
Analyst

03/18/2025
Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bp031825

SequenceID : bp031825 NA NO YES