

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF031025

Review By	anahy	Review On	3/11/2025 9:11:47 AM
Supervise By	Jagrut	Supervise On	3/11/2025 10:18:40 AM
SubDirectory	BF031025	HP Acquire Method	BNA_F
		HP Processing Method	bf031025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12653,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141896.D	10 Mar 2025 10:31		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF141897.D	10 Mar 2025 11:01		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF141898.D	10 Mar 2025 11:30	Compound#9,32,54,65 removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF141899.D	10 Mar 2025 12:00		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF141900.D	10 Mar 2025 12:29	Compound #54 Kept on LR, Method is good for DOD . Method failed for compound #77.	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF141901.D	10 Mar 2025 12:58		RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF141902.D	10 Mar 2025 13:28	Not used	RC/JU	Not Ok
8	SSTDICC060	SSTDICC060	BF141903.D	10 Mar 2025 13:57	Compound#69 Removed from 60 ppm	RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF141904.D	10 Mar 2025 14:27	Compound#9,69 Removed from 80 ppm	RC/JU	Ok,M
10	SSTDICC050	SSTDICC050	BF141905.D	10 Mar 2025 15:20		RC/JU	Ok
11	SSTDICV040	ICVBF031025	BF141906.D	10 Mar 2025 15:53		RC/JU	Ok
12	PB167012BL	PB167012BL	BF141907.D	10 Mar 2025 17:09		RC/JU	Ok
13	SP6752	SP6752	BF141908.D	10 Mar 2025 17:39	8270 SPIKE-SP6752	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF031725

Review By	anahy	Review On	3/18/2025 9:44:22 AM
Supervise By	Jagrut	Supervise On	3/18/2025 11:41:13 AM
SubDirectory	BF031725	HP Acquire Method	BNA_F
		HP Processing Method	bf031025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12655,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141972.D	17 Mar 2025 09:10		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF141973.D	17 Mar 2025 09:39		RC/JU	Ok,M
3	PB167138BL	PB167138BL	BF141974.D	17 Mar 2025 10:09		RC/JU	Ok
4	PB167138BS	PB167138BS	BF141975.D	17 Mar 2025 10:38		RC/JU	Ok,M
5	Q1549-01	72-11978	BF141976.D	17 Mar 2025 11:12		RC/JU	Ok,M
6	Q1585-01	OK-02-03142025	BF141977.D	17 Mar 2025 13:01		RC/JU	Ok,M
7	Q1574-01	WC1	BF141978.D	17 Mar 2025 13:31	Poor internal standard recovery, Analyze with straight 5X	RC/JU	Not Ok
8	Q1581-01	TR-06-031425	BF141979.D	17 Mar 2025 14:00		RC/JU	Ok,M
9	Q1585-01MS	OK-02-03142025MS	BF141980.D	17 Mar 2025 14:30		RC/JU	Ok
10	Q1585-01MSD	OK-02-03142025MSD	BF141981.D	17 Mar 2025 15:00		RC/JU	Ok,M
11	Q1574-01	WC1	BF141982.D	17 Mar 2025 15:29	Need further 5X Dilution	RC/JU	Dilution
12	Q1568-04	JC-03-03132025	BF141983.D	17 Mar 2025 16:01		RC/JU	Ok
13	Q1574-01DL	WC1DL	BF141984.D	17 Mar 2025 17:00		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP031225

Review By	anahy	Review On	3/13/2025 11:41:15 AM
Supervise By	Jagrut	Supervise On	3/17/2025 4:44:19 PM
SubDirectory	BP031225	HP Acquire Method	BNA_P
		HP Processing Method	bp031225

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12653,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024159.D	12 Mar 2025 15:36		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP024160.D	12 Mar 2025 16:17		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP024161.D	12 Mar 2025 16:57	Compound #32,54,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP024162.D	12 Mar 2025 17:38		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BP024163.D	12 Mar 2025 18:19		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BP024164.D	12 Mar 2025 19:00	This calibration is good for 8270E and 8270E DOD methods	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BP024165.D	12 Mar 2025 19:41		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP024166.D	12 Mar 2025 20:21	Compound #69 Removed from 60PPM	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP024167.D	12 Mar 2025 21:02	Compound #32,54,69 Removed from 80PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBP031225	BP024168.D	12 Mar 2025 21:43		RC/JU	Ok,M
11	PB167078BL	PB167078BL	BP024169.D	12 Mar 2025 22:24		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP031825

Review By	anahy	Review On	3/19/2025 9:19:50 AM
Supervise By	mohammad	Supervise On	3/24/2025 3:03:08 AM
SubDirectory	BP031825	HP Acquire Method	BNA_P
		HP Processing Method	bp031225

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12655,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024178.D	18 Mar 2025 09:05		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024179.D	18 Mar 2025 09:46		RC/JU	Ok
3	PB167157BL	PB167157BL	BP024180.D	18 Mar 2025 10:26		RC/JU	Ok
4	PB167157BS	PB167157BS	BP024181.D	18 Mar 2025 11:07		RC/JU	Ok
5	PB167171BS	PB167171BS	BP024182.D	18 Mar 2025 11:48		RC/JU	Ok,M
6	PB167171BSD	PB167171BSD	BP024183.D	18 Mar 2025 12:29		RC/JU	Ok,M
7	PB167171BL	PB167171BL	BP024184.D	18 Mar 2025 13:10		RC/JU	Ok
8	Q1583-02	EFF-WASTE WATER	BP024185.D	18 Mar 2025 13:57	Surrogate Fail	RC/JU	Ok,M

M : Manual Integration