

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030425

Review By	Mahesh Dadoda	Review On	3/5/2025 12:14:39 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/5/2025 12:18:08 PM		
SubDirectory	VY030425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y030425s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP133216			
Initial Calibration Stds		VP133207,VP133208,VP133209,VP133210,VP133211,VP133212			
CCC		VP133214,VP133215			
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP133213			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021395.D	04 Mar 2025 08:52		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY021396.D	04 Mar 2025 09:23	not used	SY/MD	Not Ok
3	VSTDICC010	VSTDICC010	VY021397.D	04 Mar 2025 09:46		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY021398.D	04 Mar 2025 10:09	Comp.#11 is on Linear Regression	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021399.D	04 Mar 2025 10:30	Comp.#95 is on Quadratic Regression	SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY021400.D	04 Mar 2025 11:06		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY021401.D	04 Mar 2025 11:29	Method fail for comp.#13	SY/MD	Ok,M
8	VIBLK	VIBLK	VY021402.D	04 Mar 2025 11:52		SY/MD	Ok
9	VSTDICC005	VSTDICC005	VY021403.D	04 Mar 2025 12:15		SY/MD	Ok,M
10	VSTDICV050	ICVVY030425	VY021404.D	04 Mar 2025 13:12		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY031725

Review By	Maresh Dadoda	Review On	3/18/2025 1:56:06 PM
Supervise By	Semsettin Yesilyurt	Supervise On	3/18/2025 1:58:06 PM
SubDirectory	VY031725	HP Acquire Method	HP Processing Method 82y030425s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133309
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133310,VP133311 VP131783

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021542.D	17 Mar 2025 09:26		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021543.D	17 Mar 2025 10:04		SY/MD	Ok,M
3	VY0317SBL01	VY0317SBL01	VY021544.D	17 Mar 2025 10:39		SY/MD	Ok
4	VY0317SBS01	VY0317SBS01	VY021545.D	17 Mar 2025 12:49		SY/MD	Ok,M
5	VY0317SBSD01	VY0317SBSD01	VY021546.D	17 Mar 2025 13:12		SY/MD	Ok,M
6	Q1585-01	OK-02-03142025	VY021547.D	17 Mar 2025 13:39	Internal Standard Fail;Surrogate Fail	SY/MD	ReRun
7	Q1581-01	TR-06-031425	VY021548.D	17 Mar 2025 14:03	Not purge	SY/MD	Not Ok
8	Q1574-01	WC1	VY021549.D	17 Mar 2025 14:26	Internal Standard Fail	SY/MD	ReRun
9	Q1585-01RE	OK-02-03142025RE	VY021550.D	17 Mar 2025 15:08	Internal Standard Fail	SY/MD	Confirms
10	Q1581-01	TR-06-031425	VY021551.D	17 Mar 2025 15:32	Not purge	SY/MD	Not Ok
11	Q1574-01RE	WC1RE	VY021552.D	17 Mar 2025 15:55	Internal Standard Fail	SY/MD	Confirms
12	VY0317SBS02	VY0317SBS02	VY021553.D	17 Mar 2025 16:18	Not Required	SY/MD	Not Ok
13	VSTDCCC050	VSTDCCC050EC	VY021554.D	17 Mar 2025 16:40		SY/MD	Ok,M

M : Manual Integration