

Surrogate Summary

SDG No.: **Q1574**

Client: **G Environmental**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1574-01	WC1	1,2-Dichloroethane-d4	50	49.0	98	70 (63)	130 (155)
		Dibromofluoromethane	50	53.6	107	70 (70)	130 (134)
		Toluene-d8	50	45.9	92	70 (74)	130 (123)
		4-Bromofluorobenzene	50	30.1	60 *	70 (38)	130 (136)
Q1574-01RE	WC1RE	1,2-Dichloroethane-d4	50	54.7	109	70 (63)	130 (155)
		Dibromofluoromethane	50	53.9	108	70 (70)	130 (134)
		Toluene-d8	50	46.9	94	70 (74)	130 (123)
		4-Bromofluorobenzene	50	32.8	66 *	70 (38)	130 (136)
VY0317SBL01	VY0317SBL01	1,2-Dichloroethane-d4	50	52.5	105	70 (63)	130 (155)
		Dibromofluoromethane	50	50.1	100	70 (70)	130 (134)
		Toluene-d8	50	48.2	96	70 (74)	130 (123)
		4-Bromofluorobenzene	50	41.1	82	70 (38)	130 (136)
VY0317SBS01	VY0317SBS01	1,2-Dichloroethane-d4	50	52.0	104	70 (63)	130 (155)
		Dibromofluoromethane	50	51.5	103	70 (70)	130 (134)
		Toluene-d8	50	50.2	100	70 (74)	130 (123)
		4-Bromofluorobenzene	50	50.7	101	70 (38)	130 (136)
VY0317SBSD01	VY0317SBSD01	1,2-Dichloroethane-d4	50	51.5	103	70 (63)	130 (155)
		Dibromofluoromethane	50	50.8	102	70 (70)	130 (134)
		Toluene-d8	50	50.3	101	70 (74)	130 (123)
		4-Bromofluorobenzene	50	50.3	101	70 (38)	130 (136)

