

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1574 SAS No.: Q1574 SDG NO.: Q1574
Lab File ID: VY021543.D Date Analyzed: 03/17/2025
Instrument ID: MSVOA_Y Time Analyzed: 10:04
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	202760	7.71	316994	8.62	284592	11.42
UPPER LIMIT	405520	8.207	633988	9.116	569184	11.92
LOWER LIMIT	101380	7.207	158497	8.116	142296	10.92
EPA SAMPLE NO.						
WC1	98896 *	7.71	155221 *	8.62	116517 *	11.41
WC1RE	140963	7.71	240910	8.62	190816	11.41
VY0317SBL01	272314	7.71	493002	8.62	420924	11.41
VY0317SBS01	209203	7.71	320346	8.62	284395	11.42
VY0317SBSD01	219903	7.71	338794	8.62	299964	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.