

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1574 SAS No.: Q1574 SDG NO.: Q1574
 Lab File ID: VY021543.D Date Analyzed: 03/17/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:04
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	146377	13.346				
UPPER LIMIT	292754	13.846				
LOWER LIMIT	73188.5	12.846				
EPA SAMPLE NO.						
WC1	33476 *	13.35				
WC1RE	59425 *	13.35				
VY0317SBL01	158203	13.35				
VY0317SBS01	146283	13.35				
VY0317SBSD01	155756	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.