

Surrogate Summary

SDG No.: Q1575

Client: G Environmental

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1575-01	MW3	1,2-Dichloroethane-d4	50	52.2	104	70 (74)	130 (125)
		Dibromofluoromethane	50	48.5	97	70 (75)	130 (124)
		Toluene-d8	50	52.4	105	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.9	110	70 (77)	130 (121)
Q1575-02	MW7	1,2-Dichloroethane-d4	50	55.0	110	70 (74)	130 (125)
		Dibromofluoromethane	50	52.9	106	70 (75)	130 (124)
		Toluene-d8	50	49.8	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.1	100	70 (77)	130 (121)
VN0320WBL01	VN0320WBL01	1,2-Dichloroethane-d4	50	58.2	116	70 (74)	130 (125)
		Dibromofluoromethane	50	54.8	110	70 (75)	130 (124)
		Toluene-d8	50	48.6	97	70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.0	86	70 (77)	130 (121)
VN0320WBS01	VN0320WBS01	1,2-Dichloroethane-d4	50	48.7	97	70 (74)	130 (125)
		Dibromofluoromethane	50	49.0	98	70 (75)	130 (124)
		Toluene-d8	50	49.3	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.6	95	70 (77)	130 (121)
VN0320WBSD0	VN0320WBSD01	1,2-Dichloroethane-d4	50	49.0	98	70 (74)	130 (125)
		Dibromofluoromethane	50	47.5	95	70 (75)	130 (124)
		Toluene-d8	50	48.9	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.2	96	70 (77)	130 (121)

