

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1575 SAS No.: Q1575 SDG NO.: Q1575
Lab File ID: VN086025.D Date Analyzed: 03/20/2025
Instrument ID: MSVOA_N Time Analyzed: 15:13
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	154803	8.22	247467	9.10	235468	11.87
UPPER LIMIT	309606	8.724	494934	9.6	470936	12.365
LOWER LIMIT	77401.5	7.724	123734	8.6	117734	11.365
EPA SAMPLE NO.						
MW3	157894	8.22	282837	9.10	271249	11.87
MW7	138616	8.22	242981	9.10	226759	11.87
VN0320WBL01	133645	8.22	242254	9.10	218780	11.87
VN0320WBS01	171117	8.22	271859	9.10	243713	11.87
VN0320WBSD01	163068	8.22	259777	9.10	237082	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.