

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1575 SAS No.: Q1575 SDG NO.: Q1575
 Lab File ID: VN086025.D Date Analyzed: 03/20/2025
 Instrument ID: MSVOA_N Time Analyzed: 15:13
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	132733	13.788				
UPPER LIMIT	265466	14.288				
LOWER LIMIT	66366.5	13.288				
EPA SAMPLE NO.						
MW3	117699	13.79				
MW7	92285	13.79				
VN0320WBL01	87559	13.79				
VN0320WBS01	126819	13.79				
VN0320WBSD01	122440	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.