

INSTRUMENT IN: Cx1



Accept=Accepted

[illegible]

LB135071 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Mar 2025 12:48:46

Sample ID	Extended ID	μ Abs	Conc	Std Cond Method	Units	Date	Type	Type
0.0 - 1	50	-180	-	0.00007470A	PPB	18 Mar 2025 13:17:39	S	Std
0.2 - 1	50.2	849	-	0.20007470A	PPB	18 Mar 2025 13:19:56	S	Std
2.5 - 1	50.5	10573	-	2.50007470A	PPB	18 Mar 2025 13:22:13	S	Std
5.0 - 1	55	22798	-	5.00007470A	PPB	18 Mar 2025 13:32:20	S	Std
7.5 - 1	57.1	34120	-	7.50007470A	PPB	18 Mar 2025 13:34:36	S	Std
10.0 - 1	57.1	42791	-	10.00007470A	PPB	18 Mar 2025 13:39:34	S	Std
ICV47 - 1	ICV47	18511	4.2044	-7470A	PPB	18 Mar 2025 13:42:31	U	SMPL
ICB47 - 1	ICB47	188	0.0370	-7470A	PPB	18 Mar 2025 13:44:46	U	SMPL
CCV23 - 1	CCV23	22809	5.1819	-7470A	PPB	18 Mar 2025 13:47:03	U	SMPL
CCB23 - 1	CCB23	71	0.0104	-7470A	PPB	18 Mar 2025 13:49:18	U	SMPL
CRA - 1	CRA	880	0.1944	-7470A	PPB	18 Mar 2025 13:51:35	U	SMPL
HighStd - 1	HighStd	42251	9.6037	-7470A	PPB	18 Mar 2025 13:53:51	U	SMPL
ChkStd - 1	ChkStd	30740	6.9857	-7470A	PPB	18 Mar 2025 13:58:46	U	SMPL
PB167196BL - 1	PBW	-286	-0.0708	-7470A	PPB	18 Mar 2025 14:01:01	U	SMPL
PB167196BS - 1	LCSW	17100	3.8834	-7470A	PPB	18 Mar 2025 14:03:20	U	SMPL
Q1553-03 - 1	FRAC-TANK-FM1120	1245	0.2774	-7470A	PPB	18 Mar 2025 14:05:36	U	SMPL
Q1553-03DUP - 1	FRAC-TANK-FM1120DUP	1114	0.2476	-7470A	PPB	18 Mar 2025 14:07:54	U	SMPL
Q1553-03MS - 1	FRAC-TANK-FM1120MS	12244	2.7790	-7470A	PPB	18 Mar 2025 14:10:11	U	SMPL
Q1553-03MSD - 1	FRAC-TANK-FM1120MSD	12644	2.8700	-7470A	PPB	18 Mar 2025 14:12:28	U	SMPL
Q1563-01 - 1	437	2063	0.4635	-7470A	PPB	18 Mar 2025 14:14:46	U	SMPL
CCV24 - 1	CCV24	20755	4.7147	-7470A	PPB	18 Mar 2025 14:17:04	U	SMPL
CCB24 - 1	CCB24	36	0.0024	-7470A	PPB	18 Mar 2025 14:19:21	U	SMPL
Q1563-02 - 1	FERNOT-WATER	372	0.0789	-7470A	PPB	18 Mar 2025 14:21:39	U	SMPL
Q1576-01 - 1	MW14	48064	10.9258	-7470A	PPB	18 Mar 2025 14:23:54	U	SMPL
Q1577-03 - 1	MOO-25-0072	-613	-0.1452	-7470A	PPB	18 Mar 2025 14:26:10	U	SMPL
Q1588-01 - 1	TAPFTA-MW01D-031425-00-T1	92	0.0152	-7470A	PPB	18 Mar 2025 14:28:32	U	SMPL
Q1590-04 - 1	3799	144	0.0270	-7470A	PPB	18 Mar 2025 14:30:48	U	SMPL
Q1553-03LX5 - 1		376	0.0798	-7470A	PPB	18 Mar 2025 14:44:49	U	SMPL
Q1553-03A - 1		17027	3.8668	-7470A	PPB	18 Mar 2025 14:47:05	U	SMPL
Q1576-01DLX2 - 1		22721	5.1619	-7470A	PPB	18 Mar 2025 14:49:21	U	SMPL
CCV25 - 1	CCV25	21064	4.7850	-7470A	PPB	18 Mar 2025 14:54:41	U	SMPL
CCB25 - 1	CCB25	-31	-0.0128	-7470A	PPB	18 Mar 2025 14:56:57	U	SMPL