



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	G Environmental				SDG No.:	Q1576				
Contract:	GENV01	Lab Code:	CHEM		Case No.:	Q1576	SAS No.:	Q1576		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB47	Mercury	0.20	+/-0.20	U	0.20	CV	03/18/2025	13:44	LB135071	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q1576

Contract: GENV01

Lab Code: CHEM

Case No.: Q1576

SAS No.: Q1576

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB23	Mercury	0.20	+/-0.20	U	0.20	CV	03/18/2025	13:49	LB135071
CCB24	Mercury	0.20	+/-0.20	U	0.20	CV	03/18/2025	14:19	LB135071
CCB25	Mercury	0.20	+/-0.20	U	0.20	CV	03/18/2025	14:56	LB135071

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1576</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1576</u>	SAS No.: <u>Q1576</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	03/21/2025	10:11	LB135126
	Antimony	50.0	+/-50.0	U	50.0	P	03/21/2025	10:11	LB135126
	Arsenic	20.0	+/-20.0	U	20.0	P	03/21/2025	10:11	LB135126
	Barium	100	+/-100	U	100	P	03/21/2025	10:11	LB135126
	Beryllium	6.00	+/-6.00	U	6.00	P	03/21/2025	10:11	LB135126
	Cadmium	6.00	+/-6.00	U	6.00	P	03/21/2025	10:11	LB135126
	Calcium	2000	+/-2000	U	2000	P	03/21/2025	10:11	LB135126
	Chromium	10.0	+/-10.0	U	10.0	P	03/21/2025	10:11	LB135126
	Cobalt	30.0	+/-30.0	U	30.0	P	03/21/2025	10:11	LB135126
	Copper	20.0	+/-20.0	U	20.0	P	03/21/2025	10:11	LB135126
	Iron	100	+/-100	U	100	P	03/21/2025	10:11	LB135126
	Lead	12.0	+/-12.0	U	12.0	P	03/21/2025	10:11	LB135126
	Magnesium	2000	+/-2000	U	2000	P	03/21/2025	10:11	LB135126
	Manganese	20.0	+/-20.0	U	20.0	P	03/21/2025	10:11	LB135126
	Nickel	40.0	+/-40.0	U	40.0	P	03/21/2025	10:11	LB135126
	Potassium	2000	+/-2000	U	2000	P	03/21/2025	10:11	LB135126
	Selenium	20.0	+/-20.0	U	20.0	P	03/21/2025	10:11	LB135126
	Silver	10.0	+/-10.0	U	10.0	P	03/21/2025	10:11	LB135126
	Sodium	2000	+/-2000	U	2000	P	03/21/2025	10:11	LB135126
	Thallium	40.0	+/-40.0	U	40.0	P	03/21/2025	10:11	LB135126
	Vanadium	40.0	+/-40.0	U	40.0	P	03/21/2025	10:11	LB135126
	Zinc	40.0	+/-40.0	U	40.0	P	03/21/2025	10:11	LB135126

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental SDG No.: Q1576
Contract: GENV01 Lab Code: CHEM Case No.: Q1576 SAS No.: Q1576

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	03/21/2025	10:46	LB135126
	Antimony	50.0	+/-50.0	U	50.0	P	03/21/2025	10:46	LB135126
	Arsenic	20.0	+/-20.0	U	20.0	P	03/21/2025	10:46	LB135126
	Barium	100	+/-100	U	100	P	03/21/2025	10:46	LB135126
	Beryllium	6.00	+/-6.00	U	6.00	P	03/21/2025	10:46	LB135126
	Cadmium	6.00	+/-6.00	U	6.00	P	03/21/2025	10:46	LB135126
	Calcium	2000	+/-2000	U	2000	P	03/21/2025	10:46	LB135126
	Chromium	10.0	+/-10.0	U	10.0	P	03/21/2025	10:46	LB135126
	Cobalt	30.0	+/-30.0	U	30.0	P	03/21/2025	10:46	LB135126
	Copper	20.0	+/-20.0	U	20.0	P	03/21/2025	10:46	LB135126
	Iron	100	+/-100	U	100	P	03/21/2025	10:46	LB135126
	Lead	12.0	+/-12.0	U	12.0	P	03/21/2025	10:46	LB135126
	Magnesium	2000	+/-2000	U	2000	P	03/21/2025	10:46	LB135126
	Manganese	20.0	+/-20.0	U	20.0	P	03/21/2025	10:46	LB135126
	Nickel	40.0	+/-40.0	U	40.0	P	03/21/2025	10:46	LB135126
	Potassium	2000	+/-2000	U	2000	P	03/21/2025	10:46	LB135126
	Selenium	20.0	+/-20.0	U	20.0	P	03/21/2025	10:46	LB135126
	Silver	10.0	+/-10.0	U	10.0	P	03/21/2025	10:46	LB135126
	Sodium	2000	+/-2000	U	2000	P	03/21/2025	10:46	LB135126
	Thallium	40.0	+/-40.0	U	40.0	P	03/21/2025	10:46	LB135126
	Vanadium	40.0	+/-40.0	U	40.0	P	03/21/2025	10:46	LB135126
	Zinc	40.0	+/-40.0	U	40.0	P	03/21/2025	10:46	LB135126
CCB02	Aluminum	100	+/-100	U	100	P	03/21/2025	11:37	LB135126
	Antimony	50.0	+/-50.0	U	50.0	P	03/21/2025	11:37	LB135126
	Arsenic	20.0	+/-20.0	U	20.0	P	03/21/2025	11:37	LB135126
	Barium	100	+/-100	U	100	P	03/21/2025	11:37	LB135126
	Beryllium	6.00	+/-6.00	U	6.00	P	03/21/2025	11:37	LB135126
	Cadmium	6.00	+/-6.00	U	6.00	P	03/21/2025	11:37	LB135126
	Calcium	2000	+/-2000	U	2000	P	03/21/2025	11:37	LB135126
	Chromium	10.0	+/-10.0	U	10.0	P	03/21/2025	11:37	LB135126
	Cobalt	30.0	+/-30.0	U	30.0	P	03/21/2025	11:37	LB135126
	Copper	20.0	+/-20.0	U	20.0	P	03/21/2025	11:37	LB135126
	Iron	100	+/-100	U	100	P	03/21/2025	11:37	LB135126
	Lead	12.0	+/-12.0	U	12.0	P	03/21/2025	11:37	LB135126
	Magnesium	2000	+/-2000	U	2000	P	03/21/2025	11:37	LB135126
	Manganese	20.0	+/-20.0	U	20.0	P	03/21/2025	11:37	LB135126
	Nickel	40.0	+/-40.0	U	40.0	P	03/21/2025	11:37	LB135126
	Potassium	2000	+/-2000	U	2000	P	03/21/2025	11:37	LB135126
	Selenium	20.0	+/-20.0	U	20.0	P	03/21/2025	11:37	LB135126

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q1576							
Contract: GENV01	Lab Code: CHEM	Case No.: Q1576	SAS No.: Q1576						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	03/21/2025	11:37	LB135126
	Sodium	2000	+/-2000	U	2000	P	03/21/2025	11:37	LB135126
	Thallium	40.0	+/-40.0	U	40.0	P	03/21/2025	11:37	LB135126
	Vanadium	40.0	+/-40.0	U	40.0	P	03/21/2025	11:37	LB135126
	Zinc	40.0	+/-40.0	U	40.0	P	03/21/2025	11:37	LB135126
CCB03	Aluminum	100	+/-100	U	100	P	03/21/2025	12:31	LB135126
	Antimony	50.0	+/-50.0	U	50.0	P	03/21/2025	12:31	LB135126
	Arsenic	20.0	+/-20.0	U	20.0	P	03/21/2025	12:31	LB135126
	Barium	100	+/-100	U	100	P	03/21/2025	12:31	LB135126
	Beryllium	6.00	+/-6.00	U	6.00	P	03/21/2025	12:31	LB135126
	Cadmium	6.00	+/-6.00	U	6.00	P	03/21/2025	12:31	LB135126
	Calcium	2000	+/-2000	U	2000	P	03/21/2025	12:31	LB135126
	Chromium	10.0	+/-10.0	U	10.0	P	03/21/2025	12:31	LB135126
	Cobalt	30.0	+/-30.0	U	30.0	P	03/21/2025	12:31	LB135126
	Copper	20.0	+/-20.0	U	20.0	P	03/21/2025	12:31	LB135126
	Iron	100	+/-100	U	100	P	03/21/2025	12:31	LB135126
	Lead	12.0	+/-12.0	U	12.0	P	03/21/2025	12:31	LB135126
	Magnesium	2000	+/-2000	U	2000	P	03/21/2025	12:31	LB135126
	Manganese	20.0	+/-20.0	U	20.0	P	03/21/2025	12:31	LB135126
	Nickel	40.0	+/-40.0	U	40.0	P	03/21/2025	12:31	LB135126
	Potassium	2000	+/-2000	U	2000	P	03/21/2025	12:31	LB135126
	Selenium	20.0	+/-20.0	U	20.0	P	03/21/2025	12:31	LB135126
	Silver	10.0	+/-10.0	U	10.0	P	03/21/2025	12:31	LB135126
	Sodium	2000	+/-2000	U	2000	P	03/21/2025	12:31	LB135126
	Thallium	40.0	+/-40.0	U	40.0	P	03/21/2025	12:31	LB135126
	Vanadium	40.0	+/-40.0	U	40.0	P	03/21/2025	12:31	LB135126
	Zinc	40.0	+/-40.0	U	40.0	P	03/21/2025	12:31	LB135126
CCB04	Aluminum	100	+/-100	U	100	P	03/21/2025	13:21	LB135126
	Antimony	50.0	+/-50.0	U	50.0	P	03/21/2025	13:21	LB135126
	Arsenic	20.0	+/-20.0	U	20.0	P	03/21/2025	13:21	LB135126
	Barium	100	+/-100	U	100	P	03/21/2025	13:21	LB135126
	Beryllium	6.00	+/-6.00	U	6.00	P	03/21/2025	13:21	LB135126
	Cadmium	6.00	+/-6.00	U	6.00	P	03/21/2025	13:21	LB135126
	Calcium	2000	+/-2000	U	2000	P	03/21/2025	13:21	LB135126
	Chromium	10.0	+/-10.0	U	10.0	P	03/21/2025	13:21	LB135126
	Cobalt	30.0	+/-30.0	U	30.0	P	03/21/2025	13:21	LB135126
	Copper	20.0	+/-20.0	U	20.0	P	03/21/2025	13:21	LB135126
	Iron	100	+/-100	U	100	P	03/21/2025	13:21	LB135126
	Lead	12.0	+/-12.0	U	12.0	P	03/21/2025	13:21	LB135126

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1576</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1576</u>	SAS No.: <u>Q1576</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	03/21/2025	13:21	LB135126
	Manganese	20.0	+/-20.0	U	20.0	P	03/21/2025	13:21	LB135126
	Nickel	40.0	+/-40.0	U	40.0	P	03/21/2025	13:21	LB135126
	Potassium	2000	+/-2000	U	2000	P	03/21/2025	13:21	LB135126
	Selenium	20.0	+/-20.0	U	20.0	P	03/21/2025	13:21	LB135126
	Silver	10.0	+/-10.0	U	10.0	P	03/21/2025	13:21	LB135126
	Sodium	2000	+/-2000	U	2000	P	03/21/2025	13:21	LB135126
	Thallium	40.0	+/-40.0	U	40.0	P	03/21/2025	13:21	LB135126
	Vanadium	40.0	+/-40.0	U	40.0	P	03/21/2025	13:21	LB135126
	Zinc	40.0	+/-40.0	U	40.0	P	03/21/2025	13:21	LB135126
CCB05	Aluminum	100	+/-100	U	100	P	03/21/2025	14:11	LB135126
	Antimony	50.0	+/-50.0	U	50.0	P	03/21/2025	14:11	LB135126
	Arsenic	20.0	+/-20.0	U	20.0	P	03/21/2025	14:11	LB135126
	Barium	100	+/-100	U	100	P	03/21/2025	14:11	LB135126
	Beryllium	6.00	+/-6.00	U	6.00	P	03/21/2025	14:11	LB135126
	Cadmium	6.00	+/-6.00	U	6.00	P	03/21/2025	14:11	LB135126
	Calcium	2000	+/-2000	U	2000	P	03/21/2025	14:11	LB135126
	Chromium	10.0	+/-10.0	U	10.0	P	03/21/2025	14:11	LB135126
	Cobalt	30.0	+/-30.0	U	30.0	P	03/21/2025	14:11	LB135126
	Copper	20.0	+/-20.0	U	20.0	P	03/21/2025	14:11	LB135126
	Iron	100	+/-100	U	100	P	03/21/2025	14:11	LB135126
	Lead	12.0	+/-12.0	U	12.0	P	03/21/2025	14:11	LB135126
	Magnesium	2000	+/-2000	U	2000	P	03/21/2025	14:11	LB135126
	Manganese	20.0	+/-20.0	U	20.0	P	03/21/2025	14:11	LB135126
	Nickel	40.0	+/-40.0	U	40.0	P	03/21/2025	14:11	LB135126
	Potassium	2000	+/-2000	U	2000	P	03/21/2025	14:11	LB135126
	Selenium	20.0	+/-20.0	U	20.0	P	03/21/2025	14:11	LB135126
	Silver	10.0	+/-10.0	U	10.0	P	03/21/2025	14:11	LB135126
	Sodium	2000	+/-2000	U	2000	P	03/21/2025	14:11	LB135126
	Thallium	40.0	+/-40.0	U	40.0	P	03/21/2025	14:11	LB135126
	Vanadium	40.0	+/-40.0	U	40.0	P	03/21/2025	14:11	LB135126
	Zinc	40.0	+/-40.0	U	40.0	P	03/21/2025	14:11	LB135126
CCB06	Aluminum	100	+/-100	U	100	P	03/21/2025	15:02	LB135126
	Antimony	50.0	+/-50.0	U	50.0	P	03/21/2025	15:02	LB135126
	Arsenic	20.0	+/-20.0	U	20.0	P	03/21/2025	15:02	LB135126
	Barium	100	+/-100	U	100	P	03/21/2025	15:02	LB135126
	Beryllium	6.00	+/-6.00	U	6.00	P	03/21/2025	15:02	LB135126
	Cadmium	6.00	+/-6.00	U	6.00	P	03/21/2025	15:02	LB135126
	Calcium	2000	+/-2000	U	2000	P	03/21/2025	15:02	LB135126

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1576</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1576</u>	SAS No.: <u>Q1576</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	03/21/2025	15:02	LB135126
	Cobalt	30.0	+/-30.0	U	30.0	P	03/21/2025	15:02	LB135126
	Copper	20.0	+/-20.0	U	20.0	P	03/21/2025	15:02	LB135126
	Iron	100	+/-100	U	100	P	03/21/2025	15:02	LB135126
	Lead	12.0	+/-12.0	U	12.0	P	03/21/2025	15:02	LB135126
	Magnesium	2000	+/-2000	U	2000	P	03/21/2025	15:02	LB135126
	Manganese	20.0	+/-20.0	U	20.0	P	03/21/2025	15:02	LB135126
	Nickel	40.0	+/-40.0	U	40.0	P	03/21/2025	15:02	LB135126
	Potassium	2000	+/-2000	U	2000	P	03/21/2025	15:02	LB135126
	Selenium	20.0	+/-20.0	U	20.0	P	03/21/2025	15:02	LB135126
	Silver	10.0	+/-10.0	U	10.0	P	03/21/2025	15:02	LB135126
	Sodium	2000	+/-2000	U	2000	P	03/21/2025	15:02	LB135126
	Thallium	40.0	+/-40.0	U	40.0	P	03/21/2025	15:02	LB135126
	Vanadium	40.0	+/-40.0	U	40.0	P	03/21/2025	15:02	LB135126
	Zinc	40.0	+/-40.0	U	40.0	P	03/21/2025	15:02	LB135126
CCB07	Aluminum	100	+/-100	U	100	P	03/21/2025	15:54	LB135126
	Antimony	50.0	+/-50.0	U	50.0	P	03/21/2025	15:54	LB135126
	Arsenic	20.0	+/-20.0	U	20.0	P	03/21/2025	15:54	LB135126
	Barium	100	+/-100	U	100	P	03/21/2025	15:54	LB135126
	Beryllium	6.00	+/-6.00	U	6.00	P	03/21/2025	15:54	LB135126
	Cadmium	6.00	+/-6.00	U	6.00	P	03/21/2025	15:54	LB135126
	Calcium	2000	+/-2000	U	2000	P	03/21/2025	15:54	LB135126
	Chromium	10.0	+/-10.0	U	10.0	P	03/21/2025	15:54	LB135126
	Cobalt	30.0	+/-30.0	U	30.0	P	03/21/2025	15:54	LB135126
	Copper	20.0	+/-20.0	U	20.0	P	03/21/2025	15:54	LB135126
	Iron	100	+/-100	U	100	P	03/21/2025	15:54	LB135126
	Lead	12.0	+/-12.0	U	12.0	P	03/21/2025	15:54	LB135126
	Magnesium	2000	+/-2000	U	2000	P	03/21/2025	15:54	LB135126
	Manganese	20.0	+/-20.0	U	20.0	P	03/21/2025	15:54	LB135126
	Nickel	40.0	+/-40.0	U	40.0	P	03/21/2025	15:54	LB135126
	Potassium	2000	+/-2000	U	2000	P	03/21/2025	15:54	LB135126
	Selenium	20.0	+/-20.0	U	20.0	P	03/21/2025	15:54	LB135126
	Silver	10.0	+/-10.0	U	10.0	P	03/21/2025	15:54	LB135126
	Sodium	2000	+/-2000	U	2000	P	03/21/2025	15:54	LB135126
	Thallium	40.0	+/-40.0	U	40.0	P	03/21/2025	15:54	LB135126
	Vanadium	40.0	+/-40.0	U	40.0	P	03/21/2025	15:54	LB135126
	Zinc	40.0	+/-40.0	U	40.0	P	03/21/2025	15:54	LB135126
CCB08	Aluminum	100	+/-100	U	100	P	03/21/2025	16:48	LB135126
	Antimony	50.0	+/-50.0	U	50.0	P	03/21/2025	16:48	LB135126

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1576</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1576</u>	SAS No.: <u>Q1576</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	03/21/2025	16:48	LB135126
	Barium	100	+/-100	U	100	P	03/21/2025	16:48	LB135126
	Beryllium	6.00	+/-6.00	U	6.00	P	03/21/2025	16:48	LB135126
	Cadmium	6.00	+/-6.00	U	6.00	P	03/21/2025	16:48	LB135126
	Calcium	2000	+/-2000	U	2000	P	03/21/2025	16:48	LB135126
	Chromium	10.0	+/-10.0	U	10.0	P	03/21/2025	16:48	LB135126
	Cobalt	30.0	+/-30.0	U	30.0	P	03/21/2025	16:48	LB135126
	Copper	20.0	+/-20.0	U	20.0	P	03/21/2025	16:48	LB135126
	Iron	100	+/-100	U	100	P	03/21/2025	16:48	LB135126
	Lead	12.0	+/-12.0	U	12.0	P	03/21/2025	16:48	LB135126
	Magnesium	2000	+/-2000	U	2000	P	03/21/2025	16:48	LB135126
	Manganese	20.0	+/-20.0	U	20.0	P	03/21/2025	16:48	LB135126
	Nickel	40.0	+/-40.0	U	40.0	P	03/21/2025	16:48	LB135126
	Potassium	2000	+/-2000	U	2000	P	03/21/2025	16:48	LB135126
	Selenium	20.0	+/-20.0	U	20.0	P	03/21/2025	16:48	LB135126
	Silver	10.0	+/-10.0	U	10.0	P	03/21/2025	16:48	LB135126
	Sodium	2000	+/-2000	U	2000	P	03/21/2025	16:48	LB135126
	Thallium	40.0	+/-40.0	U	40.0	P	03/21/2025	16:48	LB135126
	Vanadium	40.0	+/-40.0	U	40.0	P	03/21/2025	16:48	LB135126
	Zinc	40.0	+/-40.0	U	40.0	P	03/21/2025	16:48	LB135126
CCB09	Aluminum	100	+/-100	U	100	P	03/21/2025	17:30	LB135126
	Antimony	50.0	+/-50.0	U	50.0	P	03/21/2025	17:30	LB135126
	Arsenic	20.0	+/-20.0	U	20.0	P	03/21/2025	17:30	LB135126
	Barium	100	+/-100	U	100	P	03/21/2025	17:30	LB135126
	Beryllium	6.00	+/-6.00	U	6.00	P	03/21/2025	17:30	LB135126
	Cadmium	6.00	+/-6.00	U	6.00	P	03/21/2025	17:30	LB135126
	Calcium	2000	+/-2000	U	2000	P	03/21/2025	17:30	LB135126
	Chromium	10.0	+/-10.0	U	10.0	P	03/21/2025	17:30	LB135126
	Cobalt	30.0	+/-30.0	U	30.0	P	03/21/2025	17:30	LB135126
	Copper	20.0	+/-20.0	U	20.0	P	03/21/2025	17:30	LB135126
	Iron	100	+/-100	U	100	P	03/21/2025	17:30	LB135126
	Lead	12.0	+/-12.0	U	12.0	P	03/21/2025	17:30	LB135126
	Magnesium	2000	+/-2000	U	2000	P	03/21/2025	17:30	LB135126
	Manganese	20.0	+/-20.0	U	20.0	P	03/21/2025	17:30	LB135126
	Nickel	40.0	+/-40.0	U	40.0	P	03/21/2025	17:30	LB135126
	Potassium	2000	+/-2000	U	2000	P	03/21/2025	17:30	LB135126
	Selenium	20.0	+/-20.0	U	20.0	P	03/21/2025	17:30	LB135126
	Silver	10.0	+/-10.0	U	10.0	P	03/21/2025	17:30	LB135126
	Sodium	2000	+/-2000	U	2000	P	03/21/2025	17:30	LB135126

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1576</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1576</u>	SAS No.: <u>Q1576</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	03/21/2025	17:30	LB135126
	Vanadium	40.0	+/-40.0	U	40.0	P	03/21/2025	17:30	LB135126
	Zinc	40.0	+/-40.0	U	40.0	P	03/21/2025	17:30	LB135126
CCB10	Aluminum	100	+/-100	U	100	P	03/21/2025	17:45	LB135126
	Antimony	50.0	+/-50.0	U	50.0	P	03/21/2025	17:45	LB135126
	Arsenic	20.0	+/-20.0	U	20.0	P	03/21/2025	17:45	LB135126
	Barium	100	+/-100	U	100	P	03/21/2025	17:45	LB135126
	Beryllium	6.00	+/-6.00	U	6.00	P	03/21/2025	17:45	LB135126
	Cadmium	6.00	+/-6.00	U	6.00	P	03/21/2025	17:45	LB135126
	Calcium	2000	+/-2000	U	2000	P	03/21/2025	17:45	LB135126
	Chromium	10.0	+/-10.0	U	10.0	P	03/21/2025	17:45	LB135126
	Cobalt	30.0	+/-30.0	U	30.0	P	03/21/2025	17:45	LB135126
	Copper	20.0	+/-20.0	U	20.0	P	03/21/2025	17:45	LB135126
	Iron	100	+/-100	U	100	P	03/21/2025	17:45	LB135126
	Lead	12.0	+/-12.0	U	12.0	P	03/21/2025	17:45	LB135126
	Magnesium	2000	+/-2000	U	2000	P	03/21/2025	17:45	LB135126
	Manganese	20.0	+/-20.0	U	20.0	P	03/21/2025	17:45	LB135126
	Nickel	40.0	+/-40.0	U	40.0	P	03/21/2025	17:45	LB135126
	Potassium	2000	+/-2000	U	2000	P	03/21/2025	17:45	LB135126
	Selenium	20.0	+/-20.0	U	20.0	P	03/21/2025	17:45	LB135126
	Silver	10.0	+/-10.0	U	10.0	P	03/21/2025	17:45	LB135126
	Sodium	2000	+/-2000	U	2000	P	03/21/2025	17:45	LB135126
	Thallium	40.0	+/-40.0	U	40.0	P	03/21/2025	17:45	LB135126
	Vanadium	40.0	+/-40.0	U	40.0	P	03/21/2025	17:45	LB135126
	Zinc	40.0	+/-40.0	U	40.0	P	03/21/2025	17:45	LB135126

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1576</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1576</u>	SAS No.: <u>Q1576</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	03/24/2025	10:49	LB135158
	Antimony	50.0	+/-50.0	U	50.0	P	03/24/2025	10:49	LB135158
	Arsenic	20.0	+/-20.0	U	20.0	P	03/24/2025	10:49	LB135158
	Barium	100	+/-100	U	100	P	03/24/2025	10:49	LB135158
	Beryllium	6.00	+/-6.00	U	6.00	P	03/24/2025	10:49	LB135158
	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	10:49	LB135158
	Calcium	2000	+/-2000	U	2000	P	03/24/2025	10:49	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	10:49	LB135158
	Cobalt	30.0	+/-30.0	U	30.0	P	03/24/2025	10:49	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	10:49	LB135158
	Iron	100	+/-100	U	100	P	03/24/2025	10:49	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	10:49	LB135158
	Magnesium	2000	+/-2000	U	2000	P	03/24/2025	10:49	LB135158
	Manganese	20.0	+/-20.0	U	20.0	P	03/24/2025	10:49	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	10:49	LB135158
	Potassium	2000	+/-2000	U	2000	P	03/24/2025	10:49	LB135158
	Selenium	20.0	+/-20.0	U	20.0	P	03/24/2025	10:49	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	10:49	LB135158
	Sodium	2000	+/-2000	U	2000	P	03/24/2025	10:49	LB135158
	Thallium	40.0	+/-40.0	U	40.0	P	03/24/2025	10:49	LB135158
	Vanadium	40.0	+/-40.0	U	40.0	P	03/24/2025	10:49	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	10:49	LB135158

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental **SDG No.:** Q1576
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1576 **SAS No.:** Q1576

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	03/24/2025	11:26	LB135158
	Antimony	50.0	+/-50.0	U	50.0	P	03/24/2025	11:26	LB135158
	Arsenic	20.0	+/-20.0	U	20.0	P	03/24/2025	11:26	LB135158
	Barium	100	+/-100	U	100	P	03/24/2025	11:26	LB135158
	Beryllium	6.00	+/-6.00	U	6.00	P	03/24/2025	11:26	LB135158
	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	11:26	LB135158
	Calcium	2000	+/-2000	U	2000	P	03/24/2025	11:26	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	11:26	LB135158
	Cobalt	30.0	+/-30.0	U	30.0	P	03/24/2025	11:26	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	11:26	LB135158
	Iron	100	+/-100	U	100	P	03/24/2025	11:26	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	11:26	LB135158
	Magnesium	2000	+/-2000	U	2000	P	03/24/2025	11:26	LB135158
	Manganese	20.0	+/-20.0	U	20.0	P	03/24/2025	11:26	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	11:26	LB135158
	Potassium	2000	+/-2000	U	2000	P	03/24/2025	11:26	LB135158
	Selenium	20.0	+/-20.0	U	20.0	P	03/24/2025	11:26	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	11:26	LB135158
	Sodium	2000	+/-2000	U	2000	P	03/24/2025	11:26	LB135158
	Thallium	40.0	+/-40.0	U	40.0	P	03/24/2025	11:26	LB135158
	Vanadium	40.0	+/-40.0	U	40.0	P	03/24/2025	11:26	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	11:26	LB135158
CCB02	Aluminum	100	+/-100	U	100	P	03/24/2025	12:17	LB135158
	Antimony	50.0	+/-50.0	U	50.0	P	03/24/2025	12:17	LB135158
	Arsenic	20.0	+/-20.0	U	20.0	P	03/24/2025	12:17	LB135158
	Barium	100	+/-100	U	100	P	03/24/2025	12:17	LB135158
	Beryllium	6.00	+/-6.00	U	6.00	P	03/24/2025	12:17	LB135158
	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	12:17	LB135158
	Calcium	2000	+/-2000	U	2000	P	03/24/2025	12:17	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	12:17	LB135158
	Cobalt	30.0	+/-30.0	U	30.0	P	03/24/2025	12:17	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	12:17	LB135158
	Iron	100	+/-100	U	100	P	03/24/2025	12:17	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	12:17	LB135158
	Magnesium	2000	+/-2000	U	2000	P	03/24/2025	12:17	LB135158
	Manganese	20.0	+/-20.0	U	20.0	P	03/24/2025	12:17	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	12:17	LB135158
	Potassium	2000	+/-2000	U	2000	P	03/24/2025	12:17	LB135158
	Selenium	20.0	+/-20.0	U	20.0	P	03/24/2025	12:17	LB135158

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1576</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1576</u>	SAS No.: <u>Q1576</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	12:17	LB135158
	Sodium	2000	+/-2000	U	2000	P	03/24/2025	12:17	LB135158
	Thallium	40.0	+/-40.0	U	40.0	P	03/24/2025	12:17	LB135158
	Vanadium	40.0	+/-40.0	U	40.0	P	03/24/2025	12:17	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	12:17	LB135158
CCB03	Aluminum	100	+/-100	U	100	P	03/24/2025	13:11	LB135158
	Antimony	50.0	+/-50.0	U	50.0	P	03/24/2025	13:11	LB135158
	Arsenic	20.0	+/-20.0	U	20.0	P	03/24/2025	13:11	LB135158
	Barium	100	+/-100	U	100	P	03/24/2025	13:11	LB135158
	Beryllium	6.00	+/-6.00	U	6.00	P	03/24/2025	13:11	LB135158
	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	13:11	LB135158
	Calcium	2000	+/-2000	U	2000	P	03/24/2025	13:11	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	13:11	LB135158
	Cobalt	30.0	+/-30.0	U	30.0	P	03/24/2025	13:11	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	13:11	LB135158
	Iron	100	+/-100	U	100	P	03/24/2025	13:11	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	13:11	LB135158
	Magnesium	2000	+/-2000	U	2000	P	03/24/2025	13:11	LB135158
	Manganese	20.0	+/-20.0	U	20.0	P	03/24/2025	13:11	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	13:11	LB135158
	Potassium	2000	+/-2000	U	2000	P	03/24/2025	13:11	LB135158
	Selenium	20.0	+/-20.0	U	20.0	P	03/24/2025	13:11	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	13:11	LB135158
	Sodium	2000	+/-2000	U	2000	P	03/24/2025	13:11	LB135158
	Thallium	40.0	+/-40.0	U	40.0	P	03/24/2025	13:11	LB135158
	Vanadium	40.0	+/-40.0	U	40.0	P	03/24/2025	13:11	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	13:11	LB135158
CCB04	Aluminum	100	+/-100	U	100	P	03/24/2025	14:11	LB135158
	Antimony	50.0	+/-50.0	U	50.0	P	03/24/2025	14:11	LB135158
	Arsenic	20.0	+/-20.0	U	20.0	P	03/24/2025	14:11	LB135158
	Barium	100	+/-100	U	100	P	03/24/2025	14:11	LB135158
	Beryllium	6.00	+/-6.00	U	6.00	P	03/24/2025	14:11	LB135158
	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	14:11	LB135158
	Calcium	2000	+/-2000	U	2000	P	03/24/2025	14:11	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	14:11	LB135158
	Cobalt	30.0	+/-30.0	U	30.0	P	03/24/2025	14:11	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	14:11	LB135158
	Iron	100	+/-100	U	100	P	03/24/2025	14:11	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	14:11	LB135158

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q1576							
Contract: GENV01	Lab Code: CHEM	Case No.: Q1576	SAS No.: Q1576						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	03/24/2025	14:11	LB135158
	Manganese	20.0	+/-20.0	U	20.0	P	03/24/2025	14:11	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	14:11	LB135158
	Potassium	2000	+/-2000	U	2000	P	03/24/2025	14:11	LB135158
	Selenium	20.0	+/-20.0	U	20.0	P	03/24/2025	14:11	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	14:11	LB135158
	Sodium	2000	+/-2000	U	2000	P	03/24/2025	14:11	LB135158
	Thallium	40.0	+/-40.0	U	40.0	P	03/24/2025	14:11	LB135158
	Vanadium	40.0	+/-40.0	U	40.0	P	03/24/2025	14:11	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	14:11	LB135158
CCB05	Aluminum	100	+/-100	U	100	P	03/24/2025	15:34	LB135158
	Antimony	50.0	+/-50.0	U	50.0	P	03/24/2025	15:34	LB135158
	Arsenic	20.0	+/-20.0	U	20.0	P	03/24/2025	15:34	LB135158
	Barium	100	+/-100	U	100	P	03/24/2025	15:34	LB135158
	Beryllium	6.00	+/-6.00	U	6.00	P	03/24/2025	15:34	LB135158
	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	15:34	LB135158
	Calcium	2000	+/-2000	U	2000	P	03/24/2025	15:34	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	15:34	LB135158
	Cobalt	30.0	+/-30.0	U	30.0	P	03/24/2025	15:34	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	15:34	LB135158
	Iron	100	+/-100	U	100	P	03/24/2025	15:34	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	15:34	LB135158
	Magnesium	2000	+/-2000	U	2000	P	03/24/2025	15:34	LB135158
	Manganese	20.0	+/-20.0	U	20.0	P	03/24/2025	15:34	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	15:34	LB135158
	Potassium	2000	+/-2000	U	2000	P	03/24/2025	15:34	LB135158
	Selenium	20.0	+/-20.0	U	20.0	P	03/24/2025	15:34	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	15:34	LB135158
	Sodium	2000	+/-2000	U	2000	P	03/24/2025	15:34	LB135158
	Thallium	40.0	+/-40.0	U	40.0	P	03/24/2025	15:34	LB135158
	Vanadium	40.0	+/-40.0	U	40.0	P	03/24/2025	15:34	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	15:34	LB135158
CCB06	Aluminum	100	+/-100	U	100	P	03/24/2025	16:43	LB135158
	Antimony	50.0	+/-50.0	U	50.0	P	03/24/2025	16:43	LB135158
	Arsenic	20.0	+/-20.0	U	20.0	P	03/24/2025	16:43	LB135158
	Barium	100	+/-100	U	100	P	03/24/2025	16:43	LB135158
	Beryllium	6.00	+/-6.00	U	6.00	P	03/24/2025	16:43	LB135158
	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	16:43	LB135158
	Calcium	2000	+/-2000	U	2000	P	03/24/2025	16:43	LB135158

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q1576							
Contract: GENV01	Lab Code: CHEM	Case No.: Q1576	SAS No.: Q1576						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	16:43	LB135158
	Cobalt	30.0	+/-30.0	U	30.0	P	03/24/2025	16:43	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	16:43	LB135158
	Iron	100	+/-100	U	100	P	03/24/2025	16:43	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	16:43	LB135158
	Magnesium	2000	+/-2000	U	2000	P	03/24/2025	16:43	LB135158
	Manganese	20.0	+/-20.0	U	20.0	P	03/24/2025	16:43	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	16:43	LB135158
	Potassium	2000	+/-2000	U	2000	P	03/24/2025	16:43	LB135158
	Selenium	20.0	+/-20.0	U	20.0	P	03/24/2025	16:43	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	16:43	LB135158
	Sodium	2000	+/-2000	U	2000	P	03/24/2025	16:43	LB135158
	Thallium	40.0	+/-40.0	U	40.0	P	03/24/2025	16:43	LB135158
	Vanadium	40.0	+/-40.0	U	40.0	P	03/24/2025	16:43	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	16:43	LB135158
CCB07	Aluminum	100	+/-100	U	100	P	03/24/2025	17:33	LB135158
	Antimony	50.0	+/-50.0	U	50.0	P	03/24/2025	17:33	LB135158
	Arsenic	20.0	+/-20.0	U	20.0	P	03/24/2025	17:33	LB135158
	Barium	100	+/-100	U	100	P	03/24/2025	17:33	LB135158
	Beryllium	6.00	+/-6.00	U	6.00	P	03/24/2025	17:33	LB135158
	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	17:33	LB135158
	Calcium	2000	+/-2000	U	2000	P	03/24/2025	17:33	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	17:33	LB135158
	Cobalt	30.0	+/-30.0	U	30.0	P	03/24/2025	17:33	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	17:33	LB135158
	Iron	100	+/-100	U	100	P	03/24/2025	17:33	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	17:33	LB135158
	Magnesium	2000	+/-2000	U	2000	P	03/24/2025	17:33	LB135158
	Manganese	20.0	+/-20.0	U	20.0	P	03/24/2025	17:33	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	17:33	LB135158
	Potassium	2000	+/-2000	U	2000	P	03/24/2025	17:33	LB135158
	Selenium	20.0	+/-20.0	U	20.0	P	03/24/2025	17:33	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	17:33	LB135158
	Sodium	2000	+/-2000	U	2000	P	03/24/2025	17:33	LB135158
	Thallium	40.0	+/-40.0	U	40.0	P	03/24/2025	17:33	LB135158
	Vanadium	40.0	+/-40.0	U	40.0	P	03/24/2025	17:33	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	17:33	LB135158
CCB08	Aluminum	100	+/-100	U	100	P	03/24/2025	18:08	LB135158
	Antimony	50.0	+/-50.0	U	50.0	P	03/24/2025	18:08	LB135158

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q1576

Contract: GENV01

Lab Code: CHEM

Case No.: Q1576

SAS No.: Q1576

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	03/24/2025	18:08	LB135158
	Barium	100	+/-100	U	100	P	03/24/2025	18:08	LB135158
	Beryllium	6.00	+/-6.00	U	6.00	P	03/24/2025	18:08	LB135158
	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	18:08	LB135158
	Calcium	2000	+/-2000	U	2000	P	03/24/2025	18:08	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	18:08	LB135158
	Cobalt	30.0	+/-30.0	U	30.0	P	03/24/2025	18:08	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	18:08	LB135158
	Iron	100	+/-100	U	100	P	03/24/2025	18:08	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	18:08	LB135158
	Magnesium	2000	+/-2000	U	2000	P	03/24/2025	18:08	LB135158
	Manganese	20.0	+/-20.0	U	20.0	P	03/24/2025	18:08	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	18:08	LB135158
	Potassium	2000	+/-2000	U	2000	P	03/24/2025	18:08	LB135158
	Selenium	20.0	+/-20.0	U	20.0	P	03/24/2025	18:08	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	18:08	LB135158
	Sodium	2000	+/-2000	U	2000	P	03/24/2025	18:08	LB135158
	Thallium	40.0	+/-40.0	U	40.0	P	03/24/2025	18:08	LB135158
	Vanadium	40.0	+/-40.0	U	40.0	P	03/24/2025	18:08	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	18:08	LB135158

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q1576

Contract: GENV01

Lab Code: CHEM

Case No.: Q1576

SAS No.: Q1576

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------