

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental **SDG No.:** Q1576
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1576 **SAS No.:** Q1576

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167177BS							
Aluminum	ug/L	1000	938		94	80 - 120	P
Antimony	ug/L	400	410		102	80 - 120	P
Arsenic	ug/L	400	401		100	80 - 120	P
Barium	ug/L	100	86.1		86	80 - 120	P
Beryllium	ug/L	100	92.5		92	80 - 120	P
Cadmium	ug/L	100	93.5		94	80 - 120	P
Calcium	ug/L	500	464	J	93	80 - 120	P
Chromium	ug/L	200	202		101	80 - 120	P
Cobalt	ug/L	100	94.5		94	80 - 120	P
Copper	ug/L	150	156		104	80 - 120	P
Iron	ug/L	1500	1430		95	80 - 120	P
Lead	ug/L	500	467		93	80 - 120	P
Magnesium	ug/L	1000	926	J	93	80 - 120	P
Manganese	ug/L	100	87.3		87	80 - 120	P
Nickel	ug/L	250	237		95	80 - 120	P
Potassium	ug/L	5000	4650		93	80 - 120	P
Selenium	ug/L	1000	995		100	80 - 120	P
Silver	ug/L	37.5	37.5		100	80 - 120	P
Sodium	ug/L	1500	1300		87	80 - 120	P
Thallium	ug/L	1000	967		97	80 - 120	P
Vanadium	ug/L	150	139		93	80 - 120	P
Zinc	ug/L	100	100		100	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q1576

Contract: GENV01

Lab Code: CHEM

Case No.: Q1576

SAS No.: Q1576

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167196BS Mercury	ug/L	4.0	3.88		97	82 - 119	CV