

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1576 SAS No.: Q1576 SDG NO.: Q1576
 Lab File ID: VX045312.D Date Analyzed: 03/18/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:02
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	99818	5.54	175808	6.76	151375	10.05
UPPER LIMIT	199636	6.044	351616	7.257	302750	10.549
LOWER LIMIT	49909	5.044	87904	6.257	75687.5	9.549
EPA SAMPLE NO.						
MW14	70033	5.54	137531	6.76	128480	10.05
VX0318WBL01	66984	5.54	130552	6.76	117786	10.05
VX0318WBS01	101216	5.55	179978	6.76	159085	10.05
VX0318WBSD01	94187	5.54	168890	6.76	150181	10.05

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.