

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031725\
 Data File : BN036681.D
 Acq On : 18 Mar 2025 04:52
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 18 06:01:22 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 16:06:28 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	107	-0.01
2	1,4-Dioxane	0.400	0.419	-4.7	100	0.00
3	n-Nitrosodimethylamine	0.400	0.404	-1.0	104	0.00
4 S	2-Fluorophenol	0.400	0.363	9.3	92	0.00
5 S	Phenol-d6	0.400	0.343	14.2	94	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.336	16.0	91	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	113	-0.01
8 S	Nitrobenzene-d5	0.400	0.341	14.8	101	-0.01
9	Naphthalene	0.400	0.355	11.3	97	-0.01
10	Hexachlorobutadiene	0.400	0.363	9.3	96	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.350	12.5	97	-0.02
12	2-Methylnaphthalene	0.400	0.352	12.0	97	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	109	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.349	12.8	92	-0.01
15 S	2-Fluorobiphenyl	0.400	0.371	7.3	98	0.00
16	Acenaphthylene	0.400	0.356	11.0	95	-0.01
17	Acenaphthene	0.400	0.361	9.8	95	-0.02
18	Fluorene	0.400	0.357	10.8	92	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	101	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.411	-2.7	89	0.00
21	4-Bromophenyl-phenylether	0.400	0.385	3.8	89	-0.01
22	Hexachlorobenzene	0.400	0.392	2.0	90	0.00
23	Atrazine	0.400	0.377	5.8	90	0.00
24	Pentachlorophenol	0.400	0.336	16.0	86	-0.01
25	Phenanthrene	0.400	0.381	4.8	89	-0.01
26	Anthracene	0.400	0.376	6.0	90	-0.01
27 SURR	Fluoranthene-d10	0.400	0.359	10.3	84	0.00
28	Fluoranthene	0.400	0.367	8.3	86	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	98	0.00
30	Pyrene	0.400	0.380	5.0	86	0.00
31 S	Terphenyl-d14	0.400	0.360	10.0	82	0.00
32	Benzo(a)anthracene	0.400	0.349	12.8	83	0.00
33	Chrysene	0.400	0.382	4.5	88	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.364	9.0	85	0.00
35 I	Perylene-d12	0.400	0.400	0.0	103	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.383	4.3	92	-0.02
37	Benzo(b)fluoranthene	0.400	0.368	8.0	89	0.00
38	Benzo(k)fluoranthene	0.400	0.359	10.3	87	0.00
39 C	Benzo(a)pyrene	0.400	0.380	5.0	92	0.00
40	Dibenzo(a,h)anthracene	0.400	0.390	2.5	97	-0.02
41	Benzo(g,h,i)perylene	0.400	0.392	2.0	94	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4