

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031825\
 Data File : BN036687.D
 Acq On : 18 Mar 2025 12:35
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 18 12:58:38 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 16:06:28 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	-0.01
2	1,4-Dioxane	0.444	0.434	2.3	70	0.00
3	n-Nitrosodimethylamine	0.898	0.916	-2.0	78	0.00
4 S	2-Fluorophenol	0.932	0.829	11.1	67	-0.01
5 S	Phenol-d6	1.152	1.015	11.9	72	-0.02
6	bis(2-Chloroethyl)ether	1.190	1.033	13.2	70	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	89	-0.01
8 S	Nitrobenzene-d5	0.435	0.372	14.5	80	-0.02
9	Naphthalene	1.177	1.056	10.3	78	-0.02
10	Hexachlorobutadiene	0.277	0.250	9.7	76	-0.01
11 SURR	2-Methylnaphthalene-d10	0.595	0.532	10.6	78	-0.02
12	2-Methylnaphthalene	0.749	0.685	8.5	80	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	89	-0.01
14 S	2,4,6-Tribromophenol	0.182	0.166	8.8	79	-0.01
15 S	2-Fluorobiphenyl	2.327	2.097	9.9	78	0.00
16	Acenaphthylene	1.888	1.715	9.2	79	-0.01
17	Acenaphthene	1.236	1.133	8.3	79	-0.02
18	Fluorene	1.672	1.550	7.3	78	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	-0.01
20	4,6-Dinitro-2-methylphenol	0.086	0.064	25.6#	76	0.00
21	4-Bromophenyl-phenylether	0.251	0.237	5.6	79	-0.01
22	Hexachlorobenzene	0.303	0.276	8.9	75	-0.01
23	Atrazine	0.201	0.187	7.0	80	0.00
24	Pentachlorophenol	0.138	0.115	16.7	77	-0.01
25	Phenanthrene	1.200	1.127	6.1	80	-0.01
26	Anthracene	1.083	0.998	7.8	80	-0.01
27 SURR	Fluoranthene-d10	1.025	0.964	6.0	79	0.00
28	Fluoranthene	1.348	1.279	5.1	81	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
30	Pyrene	1.956	1.806	7.7	81	0.00
31 S	Terphenyl-d14	0.958	0.858	10.4	80	0.00
32	Benzo(a)anthracene	1.391	1.209	13.1	80	0.00
33	Chrysene	1.520	1.454	4.3	86	0.00
34	Bis(2-ethylhexyl)phthalate	0.990	0.885	10.6	81	0.00
35 I	Perylene-d12	1.000	1.000	0.0	96	0.00
36	Indeno(1,2,3-cd)pyrene	1.444	1.248	13.6	78	-0.01
37	Benzo(b)fluoranthene	1.456	1.348	7.4	84	0.00
38	Benzo(k)fluoranthene	1.527	1.414	7.4	84	0.00
39 C	Benzo(a)pyrene	1.226	1.127	8.1	83	0.00
40	Dibenzo(a,h)anthracene	1.124	0.949	15.6	79	0.00
41	Benzo(g,h,i)perylene	1.286	1.121	12.8	78	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4