

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX022825

Review By	Mahesh Dadoda	Review On	2/28/2025 11:09:50 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:09 AM
SubDirectory	VX022825	HP Acquire Method	HP Processing Method 82X022825W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133187,VP133189
Initial Calibration Stds	VP133194,VP133195,VP133196,VP133197,VP133198,VP133199
CCC	VP133188,VP133190
Internal Standard/PEM	
ICV/I.BLK	VP133200
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045067.D	28 Feb 2025 01:03		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX045068.D	28 Feb 2025 01:27		JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX045069.D	28 Feb 2025 02:13		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX045070.D	28 Feb 2025 02:37		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX045071.D	28 Feb 2025 03:00		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX045072.D	28 Feb 2025 03:23		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX045073.D	28 Feb 2025 03:47		JC/MD	Ok,M
8	IBLK	IBLK	VX045074.D	28 Feb 2025 04:10		JC/MD	Ok
9	VSTDICV050	ICVVX022825	VX045075.D	28 Feb 2025 04:33		JC/MD	Ok,M
10	BFB	BFB	VX045076.D	28 Feb 2025 10:03		JC/MD	Ok
11	VSTDIC050	VSTDIC050	VX045077.D	28 Feb 2025 10:32	V12668	JC/MD	Ok,M
12	VX0228MBL01	VX0228MBL01	VX045078.D	28 Feb 2025 11:00		JC/MD	Ok
13	VX0228WBL01	VX0228WBL01	VX045079.D	28 Feb 2025 11:23		JC/MD	Ok
14	VX0228WBS01	VX0228WBS01	VX045080.D	28 Feb 2025 11:46		JC/MD	Ok,M
15	VX0228WBSD01	VX0228WBSD01	VX045081.D	28 Feb 2025 12:13		JC/MD	Ok,M
16	Q1401-03	BP-VPB-192-GW-840-8	VX045082.D	28 Feb 2025 12:37	vial B pH<2	JC/MD	Ok
17	Q1401-06	BP-VPB-192-GW-900-9	VX045083.D	28 Feb 2025 13:00	vial B pH<2	JC/MD	Ok
18	Q1423-01	BP-VPB-192-TB-20250	VX045084.D	28 Feb 2025 13:23	vial B pH<2 TB	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX022825

Review By	Maresh Dadoda	Review On	2/28/2025 11:09:50 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:09 AM
SubDirectory	VX022825	HP Acquire Method	HP Processing Method 82X022825W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133187,VP133189		
Initial Calibration Stds	VP133194,VP133195,VP133196,VP133197,VP133198,VP133199		
CCC	VP133188,VP133190		
Internal Standard/PEM			
ICV/I.BLK	VP133200		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1423-03	BP-VPB-192-EB-20250	VX045085.D	28 Feb 2025 13:47	vial B pH<2 EB	JC/MD	Ok
20	Q1403-01	Storage-Blank-SOIL-RE	VX045086.D	28 Feb 2025 14:10	vial B pH<2	JC/MD	Ok
21	Q1403-02	Storage-Blank-WATER-	VX045087.D	28 Feb 2025 14:33	vial B pH<2	JC/MD	Ok
22	Q1435-01	286107	VX045088.D	28 Feb 2025 14:57	vial A pH<2	JC/MD	Ok
23	Q1403-04	Storage-Blank-SAMPLE	VX045089.D	28 Feb 2025 15:20	vial B pH<2	JC/MD	Ok
24	Q1462-02	FB	VX045090.D	28 Feb 2025 15:44	vial A pH<2 FB	JC/MD	Ok
25	Q1469-01	Storage-Blank-SOIL-RE	VX045091.D	28 Feb 2025 16:07	vial A pH<2	JC/MD	Ok
26	Q1469-02	Storage-Blank-WATER-	VX045092.D	28 Feb 2025 16:31	vial A pH<2	JC/MD	Ok
27	Q1469-03	Storage-Blank-WATER-	VX045093.D	28 Feb 2025 16:54	vial A pH<2	JC/MD	Ok
28	Q1469-04	Storage-Blank-SAMPLE	VX045094.D	28 Feb 2025 17:17	vial A pH<2	JC/MD	Ok
29	Q1403-03	Storage-Blank-WATER-	VX045095.D	28 Feb 2025 17:41	vial B pH<2	JC/MD	Ok
30	Q1462-01	MW2	VX045096.D	28 Feb 2025 18:04	vial A pH<2	JC/MD	Ok,M
31	IBLK	IBLK	VX045097.D	28 Feb 2025 18:27		JC/MD	Ok
32	VSTDCCC050	VSTDCCC050EC	VX045098.D	28 Feb 2025 18:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX032025

Review By	John Carlone	Review On	3/21/2025 10:02:10 AM
Supervise By	Maresh Dadoda	Supervise On	3/21/2025 10:05:58 AM
SubDirectory	VX032025	HP Acquire Method	HP Processing Method 82X022825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133422		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133423,VP133424		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045346.D	20 Mar 2025 09:09		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045347.D	20 Mar 2025 09:38	pH#Lot#V12668	JC/MD	Ok,M
3	VX0320MBL01	VX0320MBL01	VX045348.D	20 Mar 2025 10:06		JC/MD	Ok
4	VX0320WBL01	VX0320WBL01	VX045349.D	20 Mar 2025 10:29		JC/MD	Ok
5	VX0320WBS01	VX0320WBS01	VX045350.D	20 Mar 2025 10:53		JC/MD	Ok,M
6	VX0320WBS02	VX0320WBS02	VX045351.D	20 Mar 2025 11:20	Surrogate Fail for DOD	JC/MD	Not Ok
7	Q1559-09DL	DUP03-20250310DL	VX045352.D	20 Mar 2025 11:44	vial B pH<2	JC/MD	Ok
8	Q1559-22DL	RE120D2-20250312DL	VX045353.D	20 Mar 2025 12:07	vial B pH<2	JC/MD	Ok
9	Q1559-21DL	RE120D3-20250312DL	VX045354.D	20 Mar 2025 12:30	vial B pH<2	JC/MD	Ok
10	VX0320WBSD01	VX0320WBSD01	VX045355.D	20 Mar 2025 12:53	BSD failed low for comp. #11,13	JC/MD	Ok,M
11	Q1559-08	DUP02-20250310	VX045356.D	20 Mar 2025 13:17	vial B pH<2	JC/MD	Ok
12	Q1584-02	Storage-Blank-WATER	VX045357.D	20 Mar 2025 13:40	vial A pH<2	JC/MD	Ok
13	Q1584-03	Storage-Blank-WATER	VX045358.D	20 Mar 2025 14:03	vial A pH<2	JC/MD	Ok
14	Q1584-04	Storage-Blank-SAMPLE	VX045359.D	20 Mar 2025 14:26	vial A pH<2	JC/MD	Ok
15	Q1582-16	FB	VX045360.D	20 Mar 2025 14:50	vial A pH<2 FB	JC/MD	Ok
16	Q1582-05	FB02-20250313	VX045361.D	20 Mar 2025 15:13	vial A pH<2 FB;Hit of comp.#16	JC/MD	ReRun
17	Q1584-01	Storage-Blank-SOIL-RE	VX045362.D	20 Mar 2025 15:36	vial A pH<2	JC/MD	Ok
18	Q1582-02	TT180D1-20250313	VX045363.D	20 Mar 2025 15:59	vial A pH<2 Need 4X	JC/MD	Dilution

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX032025

Review By	John Carlone	Review On	3/21/2025 10:02:10 AM
Supervise By	Maresh Dadoda	Supervise On	3/21/2025 10:05:58 AM
SubDirectory	VX032025	HP Acquire Method	HP Processing Method 82X022825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133422		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133423,VP133424		

19	Q1582-03	TT180D2-20250313	VX045364.D	20 Mar 2025 16:22	vial A pH<2 Need 10X	JC/MD	Dilution
20	Q1582-04	TT180D-20250313	VX045365.D	20 Mar 2025 16:46	vial A pH<2 E flage in previous sample	JC/MD	ReRun
21	Q1582-06	RE131D1-20250313	VX045366.D	20 Mar 2025 17:09	vial A pH<2 Need 5X	JC/MD	Dilution
22	Q1582-07	RE131D2-20250313	VX045367.D	20 Mar 2025 17:32	vial A pH<2 Need 5X	JC/MD	Dilution
23	Q1582-08	RE131D3-20250313	VX045368.D	20 Mar 2025 17:55	vial A pH<2 Need 20X	JC/MD	Dilution
24	Q1582-09	TT119S1-20250313	VX045369.D	20 Mar 2025 18:19	vial A pH<2 E flage in previous sample	JC/MD	ReRun
25	Q1582-10	RE114D1-20250313	VX045370.D	20 Mar 2025 18:42	vial A pH<2 Need 10X	JC/MD	Dilution
26	Q1582-11	BPOW1-7-20250314	VX045371.D	20 Mar 2025 19:05	vial A pH<2 E flage in previous sample	JC/MD	ReRun
27	Q1582-12	BPOW1-8-20250314	VX045372.D	20 Mar 2025 19:28	vial A pH<2	JC/MD	Ok
28	VSTDCCC050	VSTDCCC050EC	VX045373.D	20 Mar 2025 19:51		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX032125

Review By	Semsettin Yesilyurt	Review On	3/24/2025 2:03:22 AM
Supervise By	Mahesh Dadoda	Supervise On	3/24/2025 2:04:46 AM
SubDirectory	VX032125	HP Acquire Method	MSVOA_X HP Processing Method 82X022825W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133432
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133433,VP133434

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045374.D	21 Mar 2025 09:29		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045375.D	21 Mar 2025 10:00		JC/MD	Ok,M
3	VX0321MBL01	VX0321MBL01	VX045376.D	21 Mar 2025 10:29		JC/MD	Ok
4	VX0321WBL01	VX0321WBL01	VX045377.D	21 Mar 2025 10:52		JC/MD	Ok
5	VX0321WBS01	VX0321WBS01	VX045378.D	21 Mar 2025 11:15		JC/MD	Ok,M
6	VX0321WBSD01	VX0321WBSD01	VX045379.D	21 Mar 2025 11:44		JC/MD	Ok,M
7	Q1582-02DL	TT180D1-20250313DL	VX045380.D	21 Mar 2025 12:08	vial B pH<2	JC/MD	Ok
8	Q1582-06DL	RE131D1-20250313DL	VX045381.D	21 Mar 2025 12:31	vial B pH<2	JC/MD	Ok
9	Q1582-07DL	RE131D2-20250313DL	VX045382.D	21 Mar 2025 12:54	vial B pH<2	JC/MD	Ok
10	Q1582-03DL	TT180D2-20250313DL	VX045383.D	21 Mar 2025 13:18	vial B pH<2	JC/MD	Ok
11	Q1582-10DL	RE114D1-20250313DL	VX045384.D	21 Mar 2025 13:41	vial B pH<2	JC/MD	Ok
12	Q1582-08DL	RE131D3-20250313DL	VX045385.D	21 Mar 2025 14:04	vial B pH<2	JC/MD	Ok
13	Q1582-04	TT180D-20250313	VX045386.D	21 Mar 2025 14:28	vial B pH<2	JC/MD	Ok
14	Q1582-05	FB02-20250313	VX045387.D	21 Mar 2025 14:51	vial B pH<2,FB	JC/MD	Ok
15	Q1582-09	TT119S1-20250313	VX045388.D	21 Mar 2025 15:14	vial B pH<2	JC/MD	Ok
16	Q1582-11	BPOW1-7-20250314	VX045389.D	21 Mar 2025 15:38	vial B pH<2	JC/MD	Ok
17	Q1582-01	RE117D1-20250313	VX045390.D	21 Mar 2025 16:01	vial A pH<2	JC/MD	Ok
18	Q1582-13	TT191D1-20250314	VX045391.D	21 Mar 2025 16:25	vial A pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX032125

Review By	Semsettin Yesilyurt	Review On	3/24/2025 2:03:22 AM		
Supervise By	Mahesh Dadoda	Supervise On	3/24/2025 2:04:46 AM		
SubDirectory	VX032125	HP Acquire Method	MSVOA_X	HP Processing Method	82X022825W.M
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133432			
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		VP133433,VP133434			

19	Q1582-14	TT191D2-20250314	VX045392.D	21 Mar 2025 16:48	vial A pH<2	JC/MD	Ok
20	Q1582-15	DUP06-20250314	VX045393.D	21 Mar 2025 17:11	vial A pH<2	JC/MD	Ok
21	Q1608-01	TT172S1-20250317	VX045394.D	21 Mar 2025 17:35		JC/MD	Ok
22	Q1608-02	RE139D1-20250317	VX045395.D	21 Mar 2025 17:58	Need 4X	JC/MD	Dilution
23	Q1608-03	RE139D2-20250317	VX045396.D	21 Mar 2025 18:21	Surrogate Fail; Need 4X	JC/MD	Dilution
24	Q1608-04	TT205S1-20250317	VX045397.D	21 Mar 2025 18:44	vial B pH<2	JC/MD	Ok
25	Q1608-05	DUP07-20250317	VX045398.D	21 Mar 2025 19:08		JC/MD	Ok
26	Q1608-06	TT163S1-20250317	VX045399.D	21 Mar 2025 19:31		JC/MD	Ok
27	Q1608-07	RW11-MW01S-202503	VX045400.D	21 Mar 2025 19:54		JC/MD	Ok
28	VSTDCCC050	VSTDCCC050EC	VX045401.D	21 Mar 2025 20:41		JC/MD	Ok,M

M : Manual Integration