

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1582 SAS No.: Q1582 SDG NO.: Q1582
 Lab File ID: VX045375.D Date Analyzed: 03/21/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:00
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	92791	5.54	160605	6.75	137508	10.05
UPPER LIMIT	185582	6.043	321210	7.25	275016	10.549
LOWER LIMIT	46395.5	5.043	80302.5	6.25	68754	9.549
EPA SAMPLE NO.						
RE117D1-20250313	63852	5.55	125973	6.76	116271	10.05
TT180D1-20250313DL	65215	5.54	127269	6.76	114057	10.05
TT180D2-20250313DL	67762	5.54	132441	6.76	123000	10.06
TT180D-20250313	70002	5.55	139085	6.76	124948	10.06
FB02-20250313	68479	5.55	130752	6.76	122006	10.06
RE131D1-20250313DL	70298	5.55	138928	6.76	125925	10.06
RE131D2-20250313DL	66532	5.55	130418	6.76	120413	10.06
RE131D3-20250313DL	66174	5.55	129977	6.76	120263	10.05
TT119S1-20250313	66994	5.55	129665	6.76	114806	10.06
RE114D1-20250313DL	63633	5.54	122958	6.76	113115	10.06
BPOW1-7-20250314	67762	5.55	132174	6.76	120805	10.06
TT191D1-20250314	69371	5.55	133271	6.76	121816	10.05
TT191D2-20250314	68899	5.54	133744	6.76	124586	10.05
DUP06-20250314	64262	5.54	125895	6.76	116324	10.06
VX0321WBL01	67403	5.54	133286	6.76	125274	10.05
VX0321WBS01	96755	5.54	172413	6.76	151575	10.05

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1582 SAS No.: Q1582 SDG NO.: Q1582
 Lab File ID: VX045375.D Date Analyzed: 03/21/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:00
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	92791	5.54	160605	6.75	137508	10.05
UPPER LIMIT	185582	6.043	321210	7.25	275016	10.549
LOWER LIMIT	46395.5	5.043	80302.5	6.25	68754	9.549
EPA SAMPLE NO.						

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.