

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1582 SAS No.: Q1582 SDG NO.: Q1582
Lab File ID: VX045347.D Date Analyzed: 03/20/2025
Instrument ID: MSVOA_X Time Analyzed: 09:38
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	64140	12.018				
UPPER LIMIT	128280	12.518				
LOWER LIMIT	32070	11.518				
EPA SAMPLE NO.						
TT180D1-20250313	58076	12.02				
TT180D2-20250313	53957	12.02				
RE131D1-20250313	42757	12.02				
RE131D2-20250313	47211	12.02				
RE131D3-20250313	48478	12.02				
RE114D1-20250313	40317	12.02				
BPOW1-8-20250314	51214	12.02				
FB	44156	12.02				
VX0320WBL01	47437	12.02				
VX0320WBS01	65210	12.02				
VX0320WBSD01	66247	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.