

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1582 SAS No.: Q1582 SDG NO.: Q1582
 Lab File ID: VX045375.D Date Analyzed: 03/21/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:00
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	63153	12.018				
UPPER LIMIT	126306	12.518				
LOWER LIMIT	31576.5	11.518				
EPA SAMPLE NO.						
RE117D1-20250313	46151	12.02				
TT180D1-20250313DL	41111	12.02				
TT180D2-20250313DL	48525	12.02				
TT180D-20250313	48597	12.02				
FB02-20250313	47392	12.02				
RE131D1-20250313DL	52363	12.02				
RE131D2-20250313DL	50604	12.02				
RE131D3-20250313DL	48451	12.02				
TT119S1-20250313	41321	12.02				
RE114D1-20250313DL	43844	12.02				
BPOW1-7-20250314	48608	12.02				
TT191D1-20250314	46699	12.02				
TT191D2-20250314	49466	12.02				
DUP06-20250314	41891	12.02				
VX0321WBL01	52894	12.02				
VX0321WBS01	67294	12.02				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1582 SAS No.: Q1582 SDG NO.: Q1582
 Lab File ID: VX045375.D Date Analyzed: 03/21/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:00
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	63153	12.018				
UPPER LIMIT	126306	12.518				
LOWER LIMIT	31576.5	11.518				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.