

Initial and Continuing Calibration Verification

Client: Ardmore Chemical

SDG No.: Q1583

Project: PVSC Monthly 2025

RunNo.: LB135075

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.092	0.099	93	85-115	03/18/2025
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	03/18/2025
Sample ID: Cyanide	CCV2	mg/L	0.25	0.25	100	90-110	03/18/2025