

Duplicate Sample Summary

Client:	Ardmore Chemical	SDG No.:	Q1583
Project:	PVSC Monthly 2025	Sample ID:	Q1551-02
Client ID:	COMPDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	438		440		1	0.46		03/17/2025

Duplicate Sample Summary

Client:	Ardmore Chemical	SDG No.:	Q1583
Project:	PVSC Monthly 2025	Sample ID:	Q1583-01
Client ID:	EFF-WASTE WATERDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.0012	U	0.0012	U	1	0		03/18/2025

Duplicate Sample Summary

Client: Ardmore Chemical Project: PVSC Monthly 2025 Client ID: EFF-WASTE WATERMSD	SDG No.: Q1583 Sample ID: Q1583-01 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.036		0.036		1	0		03/18/2025

Duplicate Sample Summary

Client:	Ardmore Chemical	SDG No.:	Q1583
Project:	PVSC Monthly 2025	Sample ID:	Q1583-02
Client ID:	EFF-WASTE WATERDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	242		247		1	2.15		03/14/2025