

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q1583

Project: PVSC Monthly 2025

Run No.: LB135038

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135038BS							
BOD5	mg/L	198	190		96	1	84.6-115.4	03/14/2025

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q1583

Project: PVSC Monthly 2025

Run No.: LB135048

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135048BS							
TSS	mg/L	550	533		97	1	90-110	03/17/2025

Laboratory Control Sample Summary

Client: Ardmore Chemical

SDG No.: Q1583

Project: PVSC Monthly 2025

Run No.: LB135075

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167192BS							
Cyanide	mg/L	0.1	0.093		93	1	85-115	03/18/2025