



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Ardmore Chemical **SDG No.:** Q1583
Contract: ARDM01 **Lab Code:** CHEM **Case No.:** Q1583 **SAS No.:** Q1583
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.57	6.0	93	40 - 160	P	03/24/2025	10:54	LB135158
	Copper	22.5	20.0	113	40 - 160	P	03/24/2025	10:54	LB135158
	Lead	12.7	12.0	106	40 - 160	P	03/24/2025	10:54	LB135158
	Nickel	38.1	40.0	95	40 - 160	P	03/24/2025	10:54	LB135158
	Zinc	41.8	40.0	104	40 - 160	P	03/24/2025	10:54	LB135158
CRA	Mercury	0.049	0.05	98	40 - 160	CV	03/25/2025	11:45	LB135174