

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP031225

Review By	anahy	Review On	3/13/2025 11:41:15 AM
Supervise By	Jagrut	Supervise On	3/17/2025 4:44:19 PM
SubDirectory	BP031225	HP Acquire Method	BNA_P
		HP Processing Method	bp031225

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12653,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024159.D	12 Mar 2025 15:36		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP024160.D	12 Mar 2025 16:17		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP024161.D	12 Mar 2025 16:57	Compound #32,54,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP024162.D	12 Mar 2025 17:38		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BP024163.D	12 Mar 2025 18:19		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BP024164.D	12 Mar 2025 19:00	This calibration is good for 8270E and 8270E DOD methods	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BP024165.D	12 Mar 2025 19:41		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP024166.D	12 Mar 2025 20:21	Compound #69 Removed from 60PPM	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP024167.D	12 Mar 2025 21:02	Compound #32,54,69 Removed from 80PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBP031225	BP024168.D	12 Mar 2025 21:43		RC/JU	Ok,M
11	PB167078BL	PB167078BL	BP024169.D	12 Mar 2025 22:24		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP031825

Review By	anahy	Review On	3/19/2025 9:19:50 AM
Supervise By		Supervise On	
SubDirectory	BP031825	HP Acquire Method	BNA_P
		HP Processing Method	bp031225

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12655,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024178.D	18 Mar 2025 09:05		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024179.D	18 Mar 2025 09:46		RC/JU	Ok
3	PB167157BL	PB167157BL	BP024180.D	18 Mar 2025 10:26		RC/JU	Ok
4	PB167157BS	PB167157BS	BP024181.D	18 Mar 2025 11:07		RC/JU	Ok
5	PB167171BS	PB167171BS	BP024182.D	18 Mar 2025 11:48		RC/JU	Ok,NS
6	PB167171BSD	PB167171BSD	BP024183.D	18 Mar 2025 12:29		RC/JU	Ok,NS
7	PB167171BL	PB167171BL	BP024184.D	18 Mar 2025 13:10		RC/JU	Ok
8	Q1583-02	EFF-WASTE WATER	BP024185.D	18 Mar 2025 13:57	Surrogate Fail	RC/JU	Ok,NS

M : Manual Integration