

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX022125

Review By	John Carlone	Review On	2/24/2025 9:46:31 AM
Supervise By	Maresh Dadoda	Supervise On	2/24/2025 1:33:06 PM
SubDirectory	VX022125	HP Acquire Method	HP Processing Method 624X022125W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133095		
Initial Calibration Stds	VP133103,VP133104,VP133105,VP133106,VP133107		
CCC	VP133096,VP133109,VP133110		
Internal Standard/PEM			
ICV/I.BLK	VP133108		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045007.D	21 Feb 2025 09:31		JC/MD	Ok
2	VSTDIC005	VSTDIC005	VX045008.D	21 Feb 2025 09:58		JC/MD	Ok,M
3	VSTDICCC020	VSTDICCC020	VX045009.D	21 Feb 2025 10:21		JC/MD	Ok,M
4	VSTDIC050	VSTDIC050	VX045010.D	21 Feb 2025 10:44		JC/MD	Ok,M
5	VSTDIC100	VSTDIC100	VX045011.D	21 Feb 2025 11:07		JC/MD	Ok,M
6	VSTDIC150	VSTDIC150	VX045012.D	21 Feb 2025 11:29		JC/MD	Ok,M
7	IBLK	IBLK	VX045013.D	21 Feb 2025 11:52		JC/MD	Ok
8	VSTDICV020	ICVVX022125	VX045014.D	21 Feb 2025 13:09		JC/MD	Ok,M
9	VX0221WBS01	VX0221WBS01	VX045015.D	21 Feb 2025 13:41		JC/MD	Ok,M
10	VX0221WBSD01	VX0221WBSD01	VX045016.D	21 Feb 2025 14:03		JC/MD	Ok,M
11	VX0221WBL01	VX0221WBL01	VX045017.D	21 Feb 2025 14:39		JC/MD	Ok
12	Q1168-07 2.5PPB	LOD-MDL-WATER-01-0	VX045018.D	21 Feb 2025 15:12		JC/MD	Ok,M
13	Q1168-08 5.0PPB	LOQ-WATER-02-QT1-2	VX045019.D	21 Feb 2025 15:38		JC/MD	Ok,M
14	VSTDCCC020	VSTDCCC020EC	VX045020.D	21 Feb 2025 16:05		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX031425

Review By	Review On
Supervise By	Supervise On
SubDirectory	VX031425
HP Acquire Method	HP Processing Method
	624X022125W.M
STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133291
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133292,VP133293

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045284.D	14 Mar 2025 09:28		JC/MD	Ok
2	VSTDCCC020	VSTDCCC020	VX045285.D	14 Mar 2025 10:17		JC/MD	Ok,NR
3	VX0314WBS01	VX0314WBS01	VX045286.D	14 Mar 2025 10:50		JC/MD	Ok,NR
4	VX0314WBS02	VX0314WBS02	VX045287.D	14 Mar 2025 11:17	Surrogate fail	JC/MD	Not Ok
5	VX0314WBL01	VX0314WBL01	VX045288.D	14 Mar 2025 11:40		JC/MD	Ok
6	Q1502-02	PT-VOA-WP	VX045289.D	14 Mar 2025 12:09		JC/MD	Ok
7	IBLK	IBLK	VX045290.D	14 Mar 2025 12:32		JC/MD	Ok
8	Q1495-01	001-WILLETS-PT-BLVD	VX045291.D	14 Mar 2025 12:59		JC/MD	Ok,NR
9	Q1495-02	002-35TH-AVE(MAR)	VX045292.D	14 Mar 2025 13:22		JC/MD	Ok,NR
10	VX0314WBSD01	VX0314WBSD01	VX045293.D	14 Mar 2025 15:22		JC/MD	Ok,NR
11	Q1583-01	EFF-WASTE WATER	VX045294.D	14 Mar 2025 16:18		JC/MD	Ok
12	VSTDCCC020	VSTDCCC020EC	VX045295.D	14 Mar 2025 16:41		JC/MD	Ok,NR

M : Manual Integration