

Surrogate Summary

SDG No.: Q1583

Client: Ardmore Chemical

Analytical Method: SW624.1

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1583-01	EFF-WASTE WATER	1,2-Dichloroethane-d4	30	32.0	107	91	110
		Toluene-d8	30	28.9	96	91	112
		4-Bromofluorobenzene	30	31.8	106	63	112
VX0314WBL01	VX0314WBL01	1,2-Dichloroethane-d4	30	32.6	109	91	110
		Toluene-d8	30	30.8	102	91	112
		4-Bromofluorobenzene	30	27.3	91	63	112
VX0314WBS01	VX0314WBS01	1,2-Dichloroethane-d4	30	33.0	110	91	110
		Toluene-d8	30	30.9	103	91	112
		4-Bromofluorobenzene	30	30.8	103	63	112
VX0314WBSD0	VX0314WBSD01	1,2-Dichloroethane-d4	30	31.5	105	91	110
		Toluene-d8	30	29.8	99	91	112
		4-Bromofluorobenzene	30	30.7	102	63	112