

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1583

Client: Ardmore Chemical

Analytical Method: SW624.1

Datafile : VX045286.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0314WBS01	Chloromethane	20	17.9	ug/L	90			1	205	
	Vinyl Chloride	20	16.6	ug/L	83			5	195	
	Bromomethane	20	20.5	ug/L	103			15	185	
	Chloroethane	20	17.0	ug/L	85			40	160	
	Trichlorofluoromethane	20	17.9	ug/L	90			50	150	
	1,1-Dichloroethene	20	18.7	ug/L	94			50	150	
	Acrolein	100	84.9	ug/L	85			60	140	
	Acrylonitrile	100	100	ug/L	100			60	140	
	Methylene Chloride	20	19.0	ug/L	95			60	140	
	trans-1,2-Dichloroethene	20	18.5	ug/L	93			70	130	
	1,1-Dichloroethane	20	19.0	ug/L	95			70	130	
	Carbon Tetrachloride	20	19.3	ug/L	97			70	130	
	Chloroform	20	19.6	ug/L	98			70	135	
	1,1,1-Trichloroethane	20	18.8	ug/L	94			70	130	
	Benzene	20	18.6	ug/L	93			65	135	
	1,2-Dichloroethane	20	20.2	ug/L	101			70	130	
	Trichloroethene	20	18.9	ug/L	95			65	135	
	1,2-Dichloropropane	20	19.1	ug/L	96			35	165	
	Bromodichloromethane	20	19.8	ug/L	99			65	135	
	Toluene	20	19.1	ug/L	96			70	130	
	t-1,3-Dichloropropene	20	17.1	ug/L	86			50	150	
	cis-1,3-Dichloropropene	20	17.7	ug/L	89			25	175	
	1,1,2-Trichloroethane	20	19.3	ug/L	97			70	130	
	2-Chloroethyl vinyl ether	100	99.4	ug/L	99			1	225	
	Dibromochloromethane	20	19.8	ug/L	99			70	135	
	Tetrachloroethene	20	20.0	ug/L	100			70	130	
	Chlorobenzene	20	19.0	ug/L	95			65	135	
	Ethyl Benzene	20	19.1	ug/L	96			60	140	
	m/p-Xylenes	40	38.5	ug/L	96			87	111	
	o-Xylene	20	19.3	ug/L	97			87	111	
	Bromoform	20	19.2	ug/L	96			70	130	
	1,1,2,2-Tetrachloroethane	20	20.9	ug/L	104			60	140	
	1,3-Dichlorobenzene	20	19.3	ug/L	97			70	130	
	1,4-Dichlorobenzene	20	19.3	ug/L	97			65	135	
	1,2-Dichlorobenzene	20	20.0	ug/L	100			65	135	