

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1583

Client: Ardmore Chemical

Analytical Method: SW624.1

Datafile : VX045293.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0314WBSD01	Chloromethane	20	16.6	ug/L	83			1	205	20
	Vinyl Chloride	20	15.6	ug/L	78			5	195	20
	Bromomethane	20	20.4	ug/L	102			15	185	20
	Chloroethane	20	15.4	ug/L	77			40	160	20
	Trichlorofluoromethane	20	17.6	ug/L	88			50	150	20
	1,1-Dichloroethene	20	18.3	ug/L	92			50	150	20
	Acrolein	100	100	ug/L	100			60	140	20
	Acrylonitrile	100	100	ug/L	100			60	140	20
	Methylene Chloride	20	18.5	ug/L	93			60	140	20
	trans-1,2-Dichloroethene	20	17.4	ug/L	87			70	130	20
	1,1-Dichloroethane	20	18.7	ug/L	94			70	130	20
	Carbon Tetrachloride	20	19.5	ug/L	98			70	130	20
	Chloroform	20	19.3	ug/L	97			70	135	20
	1,1,1-Trichloroethane	20	19.3	ug/L	97			70	130	20
	Benzene	20	19.4	ug/L	97			65	135	20
	1,2-Dichloroethane	20	21.4	ug/L	107			70	130	20
	Trichloroethene	20	18.9	ug/L	95			65	135	20
	1,2-Dichloropropane	20	19.6	ug/L	98			35	165	20
	Bromodichloromethane	20	20.0	ug/L	100			65	135	20
	Toluene	20	18.7	ug/L	94			70	130	20
	t-1,3-Dichloropropene	20	17.3	ug/L	86			50	150	20
	cis-1,3-Dichloropropene	20	18.3	ug/L	92			25	175	20
	1,1,2-Trichloroethane	20	19.6	ug/L	98			70	130	20
	2-Chloroethyl vinyl ether	100	100	ug/L	100			1	225	20
	Dibromochloromethane	20	19.3	ug/L	97			70	135	20
	Tetrachloroethene	20	18.6	ug/L	93			70	130	20
	Chlorobenzene	20	18.5	ug/L	93			65	135	20
	Ethyl Benzene	20	18.6	ug/L	93			60	140	20
	m/p-Xylenes	40	37.9	ug/L	95			87	111	20
	o-Xylene	20	18.4	ug/L	92			87	111	20
	Bromoform	20	19.2	ug/L	96			70	130	20
	1,1,2,2-Tetrachloroethane	20	20.7	ug/L	104			60	140	20
	1,3-Dichlorobenzene	20	18.5	ug/L	93			70	130	20
	1,4-Dichlorobenzene	20	18.9	ug/L	95			65	135	20
	1,2-Dichlorobenzene	20	19.3	ug/L	97			65	135	20