

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1584 SAS No.: Q1584 SDG NO.: Q1584

Continuing Calib Date: 03/18/2025 Initial Calibration Date(s): 03/17/2025 03/17/2025

Continuing Calib Time: 12:23 Initial Calibration Time(s): 12:56 22:25

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND       |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------|-----|------------|-----------|-----------|------|------------|
|                |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1 | (1) | 4.49       | 4.49      | 4.42      | 4.56 | 0.00       |
| Aroclor-1016-2 | (2) | 4.51       | 4.50      | 4.43      | 4.57 | -0.01      |
| Aroclor-1016-3 | (3) | 4.56       | 4.56      | 4.49      | 4.63 | 0.00       |
| Aroclor-1016-4 | (4) | 4.65       | 4.65      | 4.58      | 4.72 | 0.00       |
| Aroclor-1016-5 | (5) | 4.94       | 4.94      | 4.87      | 5.01 | 0.00       |
| Aroclor-1260-1 | (1) | 6.03       | 6.03      | 5.96      | 6.10 | 0.00       |
| Aroclor-1260-2 | (2) | 6.28       | 6.28      | 6.21      | 6.35 | 0.00       |
| Aroclor-1260-3 | (3) | 6.64       | 6.63      | 6.56      | 6.70 | -0.01      |
| Aroclor-1260-4 | (4) | 6.85       | 6.85      | 6.78      | 6.92 | 0.00       |
| Aroclor-1260-5 | (5) | 7.15       | 7.15      | 7.08      | 7.22 | 0.00       |
| TCX            |     | 3.41       | 3.41      | 3.36      | 3.46 | 0.00       |
| DCB            |     | 8.51       | 8.50      | 8.40      | 8.60 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1584 SAS No.: Q1584 SDG NO.: Q1584

Continuing Calib Date: 03/18/2025 Initial Calibration Date(s): 03/17/2025 03/17/2025

Continuing Calib Time: 12:23 Initial Calibration Time(s): 12:56 22:25

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND       |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1 | (1) | 3.78       | 3.79      | 3.72                   | 3.86 | 0.01       |
| Aroclor-1016-2 | (2) | 3.80       | 3.80      | 3.73                   | 3.87 | 0.00       |
| Aroclor-1016-3 | (3) | 3.96       | 3.96      | 3.89                   | 4.03 | 0.00       |
| Aroclor-1016-4 | (4) | 4.01       | 4.01      | 3.94                   | 4.08 | 0.00       |
| Aroclor-1016-5 | (5) | 4.20       | 4.21      | 4.14                   | 4.28 | 0.01       |
| Aroclor-1260-1 | (1) | 5.20       | 5.20      | 5.13                   | 5.27 | 0.00       |
| Aroclor-1260-2 | (2) | 5.39       | 5.39      | 5.32                   | 5.46 | 0.00       |
| Aroclor-1260-3 | (3) | 5.53       | 5.53      | 5.46                   | 5.60 | 0.00       |
| Aroclor-1260-4 | (4) | 5.99       | 5.99      | 5.92                   | 6.06 | 0.00       |
| Aroclor-1260-5 | (5) | 6.24       | 6.24      | 6.17                   | 6.31 | 0.00       |
| TCX            |     | 2.78       | 2.78      | 2.73                   | 2.83 | 0.00       |
| DCB            |     | 7.55       | 7.55      | 7.45                   | 7.65 | 0.00       |