

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1584 SAS No.: Q1584 SDG NO.: Q1584

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL01 Date Analyzed: 03/18/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070091.D Time Analyzed: 10:24

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.485	4.386	4.586	230024400.000	251597000.000	9.4
Aroclor-1016-2	4.504	4.404	4.604	387085800.000	402658000.000	4.0
Aroclor-1016-3	4.562	4.463	4.663	252199800.000	260626000.000	3.3
Aroclor-1016-4	4.651	4.551	4.751	199753500.000	210909000.000	5.6
Aroclor-1016-5	4.936	4.836	5.036	196065600.000	215209000.000	9.8
Aroclor-1260-1	6.025	5.926	6.126	342744100.000	343826000.000	0.3
Aroclor-1260-2	6.280	6.181	6.381	375376700.000	394377000.000	5.1
Aroclor-1260-3	6.632	6.533	6.733	301270200.000	309137000.000	2.6
Aroclor-1260-4	6.846	6.746	6.946	326357800.000	337745000.000	3.5
Aroclor-1260-5	7.151	7.052	7.252	583624600.000	621299000.000	6.5
DCB	8.503	8.403	8.603	3797495000.000	3854630000.000	1.5
TCX	3.405	3.305	3.505	8939766000.000	9157800000.000	2.5

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1584 SAS No.: Q1584 SDG NO.: Q1584

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL01 Date Analyzed: 03/18/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070091.D Time Analyzed: 10:24

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.786	3.687	3.887	223585300.000	238378000.000	6.6
Aroclor-1016-2	3.801	3.702	3.902	365003900.000	365919000.000	0.3
Aroclor-1016-3	3.961	3.863	4.063	206649600.000	226273000.000	9.5
Aroclor-1016-4	4.010	3.911	4.111	189703100.000	196490000.000	3.6
Aroclor-1016-5	4.205	4.106	4.306	222979400.000	244713000.000	9.7
Aroclor-1260-1	5.199	5.100	5.300	400599300.000	385519000.000	-3.8
Aroclor-1260-2	5.388	5.289	5.489	466852700.000	459071000.000	-1.7
Aroclor-1260-3	5.530	5.431	5.631	444827800.000	421757000.000	-5.2
Aroclor-1260-4	5.990	5.892	6.092	392153300.000	376549000.000	-4.0
Aroclor-1260-5	6.235	6.136	6.336	828820000.000	830116000.000	0.2
DCB	7.547	7.449	7.649	5861421000.000	5763900000.000	-1.7
TCX	2.783	2.683	2.883	8065420000.000	7457040000.000	-7.6