

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1584 SAS No.: Q1584 SDG NO.: Q1584

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL02 Date Analyzed: 03/18/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070095.D Time Analyzed: 12:23

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.487	4.386	4.586	230024400.000	259126000.000	12.7
Aroclor-1016-2	4.506	4.404	4.604	387085800.000	416504000.000	7.6
Aroclor-1016-3	4.564	4.463	4.663	252199800.000	265111000.000	5.1
Aroclor-1016-4	4.653	4.551	4.751	199753500.000	216949000.000	8.6
Aroclor-1016-5	4.938	4.836	5.036	196065600.000	223284000.000	13.9
Aroclor-1260-1	6.027	5.926	6.126	342744100.000	360713000.000	5.2
Aroclor-1260-2	6.283	6.181	6.381	375376700.000	411468000.000	9.6
Aroclor-1260-3	6.635	6.533	6.733	301270200.000	338559000.000	12.4
Aroclor-1260-4	6.848	6.746	6.946	326357800.000	350002000.000	7.2
Aroclor-1260-5	7.154	7.052	7.252	583624600.000	651822000.000	11.7
DCB	8.507	8.403	8.603	3797495000.000	4041010000.000	6.4
TCX	3.406	3.305	3.505	8939766000.000	9502110000.000	6.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1584 SAS No.: Q1584 SDG NO.: Q1584

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL02 Date Analyzed: 03/18/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070095.D Time Analyzed: 12:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.784	3.687	3.887	223585300.000	273816000.000	22.5
Aroclor-1016-2	3.801	3.702	3.902	365003900.000	338370000.000	-7.3
Aroclor-1016-3	3.961	3.863	4.063	206649600.000	223522000.000	8.2
Aroclor-1016-4	4.009	3.911	4.111	189703100.000	192567000.000	1.5
Aroclor-1016-5	4.204	4.106	4.306	222979400.000	224896000.000	0.9
Aroclor-1260-1	5.199	5.100	5.300	400599300.000	413821000.000	3.3
Aroclor-1260-2	5.388	5.289	5.489	466852700.000	482104000.000	3.3
Aroclor-1260-3	5.530	5.431	5.631	444827800.000	438114000.000	-1.5
Aroclor-1260-4	5.991	5.892	6.092	392153300.000	392055000.000	0.0
Aroclor-1260-5	6.235	6.136	6.336	828820000.000	841244000.000	1.5
DCB	7.548	7.449	7.649	5861421000.000	6119090000.000	4.4
TCX	2.783	2.683	2.883	8065420000.000	7693040000.000	-4.6