

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1584 SAS No.: Q1584 SDG NO.: Q1584

EPA Sample No.: SSTD020090 Date Analyzed: 03/20/2025

Lab File ID: BM049726.D Time Analyzed: 09:09

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	371691	7.798	1408580	10.60	937063	14.44
UPPER LIMIT	743382	8.298	2817160	11.098	1874130	14.939
LOWER LIMIT	185846	7.298	704290	10.098	468532	13.939
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	315159	7.80	1122590	10.60	832020	14.44

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.