

DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

Laboratory Name : CHEMTECH

Client : RTP Environmental

Project Location : _____

Project Number : 123 - Yellow Cab - TO-15

Laboratory Sample ID(s) : Q1586

Sampling Date(s) : 3/14/2025

List DKQP Methods Used (e.g., 8260,8270, et Cetra) **TO-15**

1	For each analytical method referenced in this laboratory report package, were all specified QA/QC performance criteria followed, including the requirement to explain any criteria falling outside of acceptable guidelines, as specified in the NJDEP Data of Known Quality performance standards?	☒ Yes ☐ No
1A	Were the method specified handling, preservation, and holding time requirements met?	☒ Yes ☐ No
1B	EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)	☐ Yes ☐ No ☒ N/A
2	Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?	☒ Yes ☐ No
3	Were samples received at an appropriate temperature (4±2° C)?	☐ Yes ☐ No ☒ N/A
4	Were all QA/QC performance criteria specified in the NJDEP DKQP standards achieved?	☒ Yes ☐ No
5	a)Were reporting limits specified or referenced on the chain-of-custody or communicated to the laboratory prior to sample receipt? b)Were these reporting limits met?	☒ Yes ☐ No ☒ Yes ☐ No ☐ N/A
6	For each analytical method referenced in this laboratory report package, were results reported for all constituents identified in the method-specific analyte lists presented in the DKQP documents and/or site-specific QAPP?	☒ Yes ☐ No
7	Are project-specific matrix spikes and/or laboratory duplicates included in this data set?	☒ Yes ☐ No

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."

Cover Page

Order ID : Q1586

Project ID : Yellow Cab - TO-15

Client : RTP Environmental

Lab Sample Number

Q1586-01
Q1586-02
Q1586-03

Client Sample Number

565-SG-1
557-SG-1
557-SG-2

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 3/20/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

RTP Environmental

Project Name: Yellow Cab - TO-15

Project # N/A

Chemtech Project # Q1586

Test Name: TO-15

A. Number of Samples and Date of Receipt:

3 Air samples were received on 03/14/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: TO-15. This data package contains results for TO-15.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_L were done using GC column RTX-1, which is 60 meters, 0.32 mm id, 1.0 um df, Restek Cat. #10157. The Trap was supplied by Entech, glass bead and Tenax , Entech 7100A Preconcentrator. The analysis of TO-15 was based on method TO-15.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

Due to potential high concentration of target analytes, Samples 565-S6-1 was initially diluted.

Samples 565-S6-1, 565-S6-1DL and 557-S6-2 were diluted due to high concentrations.

E. Additional Comments:

The Sample #565-SG-1, 557-SG-2 have the concentration of target compound below Method detection limits, therefore it is not reported as Hit in Form1.

The not QT review data is reported in the Miscellaneous.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The Manual Integrations are performed for the followings.

Manual Integration Report			
Sequence	VL022525	Instrument	MSVOA_I

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICCC010	VL042041.D	m/p-Xylene	MMDadoda	3/6/2025 1:12:13 AM	SAM	3/6/2025 1:24:10 AM	Peak Integrated by Software incorrectly
VSTDICCC002	VL042042.D	m/p-Xylene	SAM	2/26/2025 7:41:55 AM	MMDadoda	2/26/2025 1:45:07 PM	Peak Integrated by Software incorrectly
VSTDICCC001	VL042043.D	1,1,2-Trichloroethane	SAM	2/26/2025 7:42:26 AM	MMDadoda	2/26/2025 1:45:09 PM	Peak Integrated by Software incorrectly
VSTDICCC001	VL042043.D	1,4-Dioxane	SAM	2/26/2025 7:42:26 AM	MMDadoda	2/26/2025 1:45:09 PM	Peak Integrated by Software incorrectly
VSTDICCC001	VL042043.D	4-Methyl-2-Pentanone	SAM	2/26/2025 7:42:26 AM	MMDadoda	2/26/2025 1:45:09 PM	Peak Integrated by Software incorrectly
VSTDICCC001	VL042043.D	Benzyl Chloride	SAM	2/26/2025 7:42:26 AM	MMDadoda	2/26/2025 1:45:09 PM	Peak Integrated by Software incorrectly
VSTDICCC001	VL042043.D	cis-1,3-Dichloropropene	SAM	2/26/2025 7:42:26 AM	MMDadoda	2/26/2025 1:45:09 PM	Peak Integrated by Software incorrectly
VSTDICCC001	VL042043.D	Ethanol	SAM	2/26/2025 7:42:26 AM	MMDadoda	2/26/2025 1:45:09 PM	Peak Integrated by

				AM			Software incorrect ly
VSTDIC001	VL042043.D	Heptane	SAM	2/26/2025 7:42:26 AM	MMDadoda	2/26/2025 1:45:09 PM	Peak Integrat ed by Software incorrect ly
VSTDIC001	VL042043.D	m/p-Xylene	SAM	2/26/2025 7:42:26 AM	MMDadoda	2/26/2025 1:45:09 PM	Peak Integrat ed by Software incorrect ly
VSTDIC001	VL042043.D	t-1,3-Dichloropropene	SAM	2/26/2025 7:42:26 AM	MMDadoda	2/26/2025 1:45:09 PM	Peak Integrat ed by Software incorrect ly
VSTDIC001	VL042043.D	trans-1,2-Dichloroethene	SAM	2/26/2025 7:42:26 AM	MMDadoda	2/26/2025 1:45:09 PM	Peak Integrat ed by Software incorrect ly
VSTDIC0.5	VL042044.D	1,1,2-Trichloroethane	SAM	2/26/2025 7:42:02 AM	MMDadoda	2/26/2025 1:45:11 PM	Peak Integrat ed by Software incorrect ly
VSTDIC0.5	VL042044.D	1,2-Dibromoethane	SAM	2/26/2025 7:42:02 AM	MMDadoda	2/26/2025 1:45:11 PM	Peak Integrat ed by Software incorrect ly
VSTDIC0.5	VL042044.D	1,2-Dichloropropane	SAM	2/26/2025 7:42:02 AM	MMDadoda	2/26/2025 1:45:11 PM	Peak Integrat ed by Software incorrect ly
VSTDIC0.5	VL042044.D	1,4-Dioxane	SAM	2/26/2025 7:42:02 AM	MMDadoda	2/26/2025 1:45:11 PM	Peak Integrat ed by Software incorrect ly
VSTDIC0.5	VL042044.D	2-Hexanone	SAM	2/26/2025 7:42:02 AM	MMDadoda	2/26/2025 1:45:11 PM	Peak Integrat ed by Software incorrect

							ly
VSTDIC0.5	VL042044.D	4-Methyl-2-Pentanone	SAM	2/26/2025 7:42:02 AM	MMDadoda	2/26/2025 1:45:11 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042044.D	cis-1,3-Dichloropropene	SAM	2/26/2025 7:42:02 AM	MMDadoda	2/26/2025 1:45:11 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042044.D	Dibromochloromethane	SAM	2/26/2025 7:42:02 AM	MMDadoda	2/26/2025 1:45:11 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042044.D	m/p-Xylene	SAM	2/26/2025 7:42:02 AM	MMDadoda	2/26/2025 1:45:11 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042044.D	t-1,3-Dichloropropene	SAM	2/26/2025 7:42:02 AM	MMDadoda	2/26/2025 1:45:11 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042044.D	trans-1,2-Dichloroethene	SAM	2/26/2025 7:42:02 AM	MMDadoda	2/26/2025 1:45:11 PM	Peak Integrated by Software incorrectly
VSTDIC0.1	VL042045.D	1,1,2,2-Tetrachloroethane	SAM	2/26/2025 7:42:22 AM	MMDadoda	2/26/2025 1:45:12 PM	Peak Integrated by Software incorrectly
VSTDIC0.1	VL042045.D	1,2-Dibromoethane	SAM	2/26/2025 7:42:22 AM	MMDadoda	2/26/2025 1:45:12 PM	Peak Integrated by Software incorrectly
VSTDIC0.1	VL042045.D	Carbon Tetrachloride	SAM	2/26/2025 7:42:22 AM	MMDadoda	2/26/2025 1:45:12 PM	Peak Integrated by Software incorrectly
VSTDIC0.1	VL042045	Tetrachloroethene	SAM	2/26/20	MMDadoda	2/26/2025	Peak

	.D			25 7:42:22 AM		1:45:12 PM	Integrat ed by Software incorrect ly
VSTDICC0.1	VL042045 .D	Trichloroethene	SAM	2/26/20 25 7:42:22 AM	MMDadoda	2/26/2025 1:45:12 PM	Peak Integrat ed by Software incorrect ly
VSTDICC0.0 3	VL042046 .D	1,1,1- Trichloroethane	SAM	2/26/20 25 7:43:48 AM	MMDadoda	2/26/2025 1:45:14 PM	Peak Integrat ed by Software incorrect ly
VSTDICC0.0 3	VL042046 .D	1,1,2,2- Tetrachloroethane	SAM	2/26/20 25 7:43:48 AM	MMDadoda	2/26/2025 1:45:14 PM	Peak Integrat ed by Software incorrect ly
VSTDICC0.0 3	VL042046 .D	Carbon Tetrachloride	SAM	2/26/20 25 7:43:48 AM	MMDadoda	2/26/2025 1:45:14 PM	Peak Integrat ed by Software incorrect ly
VSTDICC0.0 3	VL042046 .D	Tetrachloroethene	SAM	2/26/20 25 7:43:48 AM	MMDadoda	2/26/2025 1:45:14 PM	Peak Integrat ed by Software incorrect ly
VSTDICC0.0 3	VL042046 .D	Trichloroethene	SAM	2/26/20 25 7:43:48 AM	MMDadoda	2/26/2025 1:45:14 PM	Peak Integrat ed by Software incorrect ly
VSTDICC01 5	VL042047 .D	m/p-Xylene	SAM	2/26/20 25 7:42:16 AM	MMDadoda	2/26/2025 1:45:16 PM	Peak Integrat ed by Software incorrect ly
VSTDICV01 0	VL042048 .D	m/p-Xylene	SAM	2/26/20 25 7:42:09 AM	MMDadoda	2/26/2025 1:45:17 PM	Peak Integrat ed by Software incorrect ly

Manual Integration Report

Sequence	VL031725	Instrument	MSVOA_I
----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC010	VL042150.D	m/p-Xylene	SAM	3/18/2025 8:21:13 AM	MMDadoda	3/18/2025 1:56:42 PM	Peak Integrated by Software incorrectly
VL0317ABS01	VL042152.D	m/p-Xylene	SAM	3/18/2025 8:21:23 AM	MMDadoda	3/18/2025 1:56:43 PM	Peak Integrated by Software incorrectly
Q1586-01	VL042154.D	Propene	SAM	3/18/2025 8:23:13 AM	MMDadoda	3/18/2025 1:56:47 PM	Peak Integrated by Software incorrectly
Q1586-01	VL042154.D	Trichloroethene	SAM	3/18/2025 8:23:13 AM	MMDadoda	3/18/2025 1:56:47 PM	Peak Integrated by Software incorrectly
Q1586-01DL	VL042155.D	2-Butanone	SAM	3/18/2025 8:21:32 AM	MMDadoda	3/18/2025 1:56:49 PM	Peak Integrated by Software incorrectly
Q1586-01DL	VL042155.D	Ethanol	SAM	3/18/2025 8:21:32 AM	MMDadoda	3/18/2025 1:56:49 PM	Peak Integrated by Software incorrectly
Q1586-01DL	VL042155.D	o-Xylene	SAM	3/18/2025 8:21:32 AM	MMDadoda	3/18/2025 1:56:49 PM	Peak Integrated by Software incorrectly
Q1586-01DL	VL042155.D	Propene	SAM	3/18/2025 8:21:32 AM	MMDadoda	3/18/2025 1:56:49 PM	Peak Integrated by Software incorrectly

Q1586-03	VL042158.D	Bromodichloromethane	SAM	3/18/2025 8:23:22 AM	MMDadoda	3/18/2025 1:56:52 PM	Peak Integrated by Software incorrectly
Q1586-03	VL042158.D	Chlorodifluoromethane	SAM	3/18/2025 8:23:22 AM	MMDadoda	3/18/2025 1:56:52 PM	Peak Integrated by Software incorrectly
Q1586-03	VL042158.D	Cyclohexane	SAM	3/18/2025 8:23:22 AM	MMDadoda	3/18/2025 1:56:52 PM	Peak Integrated by Software incorrectly
Q1586-03	VL042158.D	Heptane	SAM	3/18/2025 8:23:22 AM	MMDadoda	3/18/2025 1:56:52 PM	Peak Integrated by Software incorrectly
Q1586-03	VL042158.D	m/p-Xylene	SAM	3/18/2025 8:23:22 AM	MMDadoda	3/18/2025 1:56:52 PM	Peak Integrated by Software incorrectly
Q1586-03	VL042158.D	Tetrahydrofuran	SAM	3/18/2025 8:23:22 AM	MMDadoda	3/18/2025 1:56:52 PM	Peak Integrated by Software incorrectly
Q1586-03	VL042158.D	Toluene	SAM	3/18/2025 8:23:22 AM	MMDadoda	3/18/2025 1:56:52 PM	Peak Integrated by Software incorrectly
Q1586-03	VL042158.D	Trichloroethene	SAM	3/18/2025 8:23:22 AM	MMDadoda	3/18/2025 1:56:52 PM	Peak Integrated by Software incorrectly
Q1586-02	VL042159.D	Carbon Tetrachloride	SAM	3/18/2025 8:23:27 AM	MMDadoda	3/18/2025 1:56:55 PM	Peak Integrated by Software incorrectly
Q1586-02	VL042159.D	Propene	SAM	3/18/2025	MMDadoda	3/18/2025 1:56:55	Peak Integrated



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

				8:23:27 AM		PM	d by Software incorrectl y
--	--	--	--	---------------	--	----	-------------------------------------

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1586

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: MAHESH PATEL

Date: 03/20/2025

LAB CHRONICLE

OrderID:	Q1586	OrderDate:	3/17/2025 9:31:17 AM
Client:	RTP Environmental	Project:	Yellow Cab - TO-15
Contact:	David Gabel	Location:	F11

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1586-01	565-SG-1	Air	TO-15	TO-15	03/14/25		03/17/25	03/14/25
Q1586-01DL	565-SG-1DL	Air	TO-15	TO-15	03/14/25		03/17/25	03/14/25
Q1586-01DL 2	565-SG-1DL2	Air	TO-15	TO-15	03/14/25		03/17/25	03/14/25
Q1586-02	557-SG-1	Air	TO-15	TO-15	03/14/25		03/17/25	03/14/25
Q1586-03	557-SG-2	Air	TO-15	TO-15	03/14/25		03/17/25	03/14/25
Q1586-03DL	557-SG-2DL	Air	TO-15	TO-15	03/14/25		03/17/25	03/14/25

Hit Summary Sheet SW-846

SDG No.: Q1586
Client: RTP Environmental

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: 565-SG-1								
Q1586-01	565-SG-1	Air	Heptane	2340	E	1.80	8.20	ug/m3
Q1586-01	565-SG-1	Air	Acetone	285	E	2.28	4.75	ug/m3
Q1586-01	565-SG-1	Air	Cyclohexane	2480	E	3.03	6.88	ug/m3
Q1586-01	565-SG-1	Air	2-Butanone	98.8		1.42	5.90	ug/m3
Q1586-01	565-SG-1	Air	Benzene	926	E	1.12	6.39	ug/m3
Q1586-01	565-SG-1	Air	Trichloroethene	2.10		0.38	0.64	ug/m3
Q1586-01	565-SG-1	Air	Toluene	74.6		1.66	7.54	ug/m3
Q1586-01	565-SG-1	Air	Tetrachloroethene	121		0.41	0.81	ug/m3
Q1586-01	565-SG-1	Air	Ethyl Benzene	46.5		2.08	8.69	ug/m3
Q1586-01	565-SG-1	Air	m/p-Xylene	169		3.65	17.4	ug/m3
Q1586-01	565-SG-1	Air	o-Xylene	50.4		2.08	8.69	ug/m3
Q1586-01	565-SG-1	Air	1,3,5-Trimethylbenzene	10.3		2.16	9.83	ug/m3
Q1586-01	565-SG-1	Air	1,2,4-Trimethylbenzene	17.7		1.52	9.83	ug/m3
Q1586-01	565-SG-1	Air	4-Ethyltoluene	5.90	J	2.36	9.83	ug/m3
Q1586-01	565-SG-1	Air	Hexane	3420	E	1.55	7.05	ug/m3
Total Voc :				10000				
Total Concentration:				10000				
Client ID: 565-SG-1DL								
Q1586-01DL	565-SG-1DL	Air	Heptane	3400	ED	18.0	82.0	ug/m3
Q1586-01DL	565-SG-1DL	Air	Cyclohexane	3790	ED	30.3	68.8	ug/m3
Q1586-01DL	565-SG-1DL	Air	2-Butanone	117	D	14.2	59.0	ug/m3
Q1586-01DL	565-SG-1DL	Air	Benzene	1410	D	11.2	63.9	ug/m3
Q1586-01DL	565-SG-1DL	Air	Toluene	74.6	JD	16.6	75.4	ug/m3
Q1586-01DL	565-SG-1DL	Air	Tetrachloroethene	142	D	4.07	8.14	ug/m3
Q1586-01DL	565-SG-1DL	Air	Ethyl Benzene	38.2	JD	20.9	86.9	ug/m3
Q1586-01DL	565-SG-1DL	Air	m/p-Xylene	148	JD	36.5	174	ug/m3
Q1586-01DL	565-SG-1DL	Air	o-Xylene	44.7	JD	20.9	86.9	ug/m3
Q1586-01DL	565-SG-1DL	Air	Hexane	16200	ED	15.5	70.5	ug/m3
Total Voc :				25400				
Total Concentration:				25400				
Client ID: 565-SG-1DL2								
Q1586-01DL2	565-SG-1DL2	Air	Heptane	2420	D	270	1230	ug/m3
Q1586-01DL2	565-SG-1DL2	Air	Cyclohexane	2680	D	447	1030	ug/m3
Q1586-01DL2	565-SG-1DL2	Air	Benzene	1340	D	167	958	ug/m3
Q1586-01DL2	565-SG-1DL2	Air	Hexane	21900	D	233	1060	ug/m3
Total Voc :				28300				
Total Concentration:				28300				

Hit Summary Sheet SW-846

SDG No.: Q1586

Client: RTP Environmental

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: 557-SG-1								
Q1586-02	557-SG-1	Air	Dichlorodifluoromethane	2.67		1.04	2.47	ug/m3
Q1586-02	557-SG-1	Air	Trichlorofluoromethane	1.74	J	0.90	2.81	ug/m3
Q1586-02	557-SG-1	Air	tert-Butyl alcohol	0.73	J	0.58	1.52	ug/m3
Q1586-02	557-SG-1	Air	Acetone	20.2		0.57	1.19	ug/m3
Q1586-02	557-SG-1	Air	Carbon Disulfide	4.67		0.50	1.56	ug/m3
Q1586-02	557-SG-1	Air	Methylene Chloride	1.74		0.66	1.74	ug/m3
Q1586-02	557-SG-1	Air	2-Butanone	2.01		0.35	1.47	ug/m3
Q1586-02	557-SG-1	Air	Carbon Tetrachloride	0.50		0.060	0.19	ug/m3
Q1586-02	557-SG-1	Air	Chloroform	3.81		0.39	2.44	ug/m3
Q1586-02	557-SG-1	Air	Benzene	0.35	J	0.29	1.60	ug/m3
Q1586-02	557-SG-1	Air	Tetrachloroethene	21.0		0.14	0.20	ug/m3
Q1586-02	557-SG-1	Air	Hexane	1.02	J	0.39	1.76	ug/m3
Total Voc :				60.5				
Total Concentration:				60.5				
Client ID: 557-SG-2								
Q1586-03	557-SG-2	Air	Dichlorodifluoromethane	2.77		1.04	2.47	ug/m3
Q1586-03	557-SG-2	Air	Chloromethane	0.60	J	0.21	1.03	ug/m3
Q1586-03	557-SG-2	Air	Trichlorofluoromethane	1.69	J	0.90	2.81	ug/m3
Q1586-03	557-SG-2	Air	tert-Butyl alcohol	1.73		0.58	1.52	ug/m3
Q1586-03	557-SG-2	Air	Heptane	1.15	J	0.45	2.05	ug/m3
Q1586-03	557-SG-2	Air	Acetone	36.1	E	0.57	1.19	ug/m3
Q1586-03	557-SG-2	Air	Carbon Disulfide	7.47		0.50	1.56	ug/m3
Q1586-03	557-SG-2	Air	Methylene Chloride	2.08		0.66	1.74	ug/m3
Q1586-03	557-SG-2	Air	2-Butanone	3.83		0.35	1.47	ug/m3
Q1586-03	557-SG-2	Air	Carbon Tetrachloride	0.57		0.060	0.19	ug/m3
Q1586-03	557-SG-2	Air	cis-1,2-Dichloroethene	0.56	J	0.36	1.98	ug/m3
Q1586-03	557-SG-2	Air	Chloroform	41.5		0.39	2.44	ug/m3
Q1586-03	557-SG-2	Air	Benzene	11.2		0.29	1.60	ug/m3
Q1586-03	557-SG-2	Air	Trichloroethene	2.53		0.11	0.16	ug/m3
Q1586-03	557-SG-2	Air	Bromodichloromethane	2.14	J	0.54	3.35	ug/m3
Q1586-03	557-SG-2	Air	Toluene	0.75	J	0.41	1.88	ug/m3
Q1586-03	557-SG-2	Air	Tetrachloroethene	88.8		0.14	0.20	ug/m3
Q1586-03	557-SG-2	Air	1,2,4-Trimethylbenzene	1.67	J	0.39	2.46	ug/m3
Q1586-03	557-SG-2	Air	Naphthalene	4.04		0.42	0.52	ug/m3
Q1586-03	557-SG-2	Air	Hexane	2.47		0.39	1.76	ug/m3
Total Voc :				214				

Hit Summary Sheet
SW-846

SDG No.: Q1586

Client: RTP Environmental

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Total Concentration:				214				
Client ID:	557-SG-2DL							
Q1586-03DL	557-SG-2DL	Air	Acetone	38.0	D	5.70	11.9	ug/m3
Q1586-03DL	557-SG-2DL	Air	2-Butanone	4.72	JD	3.54	14.8	ug/m3
Q1586-03DL	557-SG-2DL	Air	Chloroform	41.5	D	4.05	24.4	ug/m3
Q1586-03DL	557-SG-2DL	Air	Benzene	9.90	JD	2.78	16.0	ug/m3
Q1586-03DL	557-SG-2DL	Air	Tetrachloroethene	79.3	D	1.02	2.03	ug/m3
Total Voc :				173				
Total Concentration:				173				

Project : Yellow Cab - TO-15

Sampling Date : 03/14/25

Field Id Number : 565-SG-1

Analysis Date : 03/17/25

Laboratory Id Number : Q1586-01

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.84	U	4.15		
Chloromethane	74-87-3	50.49	0.39	U	0.81		
Vinyl Chloride	75-01-4	62.5	0.06	U	0.15		
Bromomethane	74-83-9	94.94	0.56	U	2.17		
Chloroethane	75-00-3	64.52	0.68	U	1.79		
Tetrahydrofuran	109-99-9	72.11	0.52	U	1.53		
Trichlorofluoromethane	75-69-4	137.4	0.64	U	3.6		
Dichlorotetrafluoroethane	76-14-2	170.9	0.35	U	2.45		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.64	U	4.91		
tert-Butyl alcohol	75-65-0	74.12	0.76	U	2.3		
Heptane	142-82-5	100.2	570	E	2335		
1,1-Dichloroethene	75-35-4	96.94	0.56	U	2.22		
Acetone	67-64-1	58.08	120	E	285		
Carbon Disulfide	75-15-0	76.14	0.64	U	1.99		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.48	U	1.73		
Methylene Chloride	75-09-2	84.94	0.76	U	2.64		
trans-1,2-Dichloroethene	156-60-5	96.94	0.76	U	3.01		
1,1-Dichloroethane	75-34-3	98.96	0.52	U	2.1		
Cyclohexane	110-82-7	84.16	720	E	2478		
2-Butanone	78-93-3	72.11	33.5		98.8		
Carbon Tetrachloride	56-23-5	153.8	0.04	U	0.25		
cis-1,2-Dichloroethene	156-59-2	96.94	0.36	U	1.43		
Chloroform	67-66-3	119.4	0.33	U	1.61		
1,1,1-Trichloroethane	71-55-6	133.4	0.04	U	0.22		
2,2,4-Trimethylpentane	540-84-1	114.2	0.4	U	1.87		
Benzene	71-43-2	78.11	290	E	926		
1,2-Dichloroethane	107-06-2	98.96	0.36	U	1.46		
Trichloroethene	79-01-6	131.4	0.39		2.1		
1,2-Dichloropropane	78-87-5	113	0.44	U	2.03		
Bromodichloromethane	75-27-4	163.8	0.31	U	2.08		
4-Methyl-2-Pentanone	108-10-1	100.2	0.37	U	1.52		
Toluene	108-88-3	92.14	19.8		74.6		
t-1,3-Dichloropropene	10061-02-6	111	0.24	U	1.09		
cis-1,3-Dichloropropene	10061-01-5	111	0.26	U	1.18		
1,1,2-Trichloroethane	79-00-5	133.4	0.28	U	1.53		

Project : Yellow Cab - TO-15

Sampling Date : 03/14/25

Field Id Number : 565-SG-1

Analysis Date : 03/17/25

Laboratory Id Number : Q1586-01

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.26	U	2.22		
1,2-Dibromoethane	106-93-4	187.9	0.29	U	2.23		
Tetrachloroethene	127-18-4	165.8	17.9		121		
Chlorobenzene	108-90-7	112.6	0.3	U	1.38		
Ethyl Benzene	100-41-4	106.2	10.7		46.5		
m/p-Xylene	179601-23-1	106.2	39		169		
o-Xylene	95-47-6	106.2	11.6		50.4		
Styrene	100-42-5	104.1	0.44	U	1.87		
Bromoform	75-25-2	252.8	0.23	U	2.38		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.04	U	0.27		
2-Chlorotoluene	95-49-8	126.6	0.44	U	2.28		
1,3,5-Trimethylbenzene	108-67-8	120.2	2.1		10.3		
1,2,4-Trimethylbenzene	95-63-6	120.2	3.6		17.7		
1,3-Dichlorobenzene	541-73-1	147	0.32	U	1.92		
1,4-Dichlorobenzene	106-46-7	147	0.21	U	1.26		
1,2-Dichlorobenzene	95-50-1	147	0.3	U	1.8		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.34	U	2.52		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.34	U	3.63		
Naphthalene	91-20-3	128.17	0.3	U	1.57		
1,3-Butadiene	106-99-0	54.09	0.52	U	1.15		
4-Ethyltoluene	622-96-8	120.2	1.2	J	5.9		
Hexane	110-54-3	86.17	970	E	3418		
Allyl Chloride	107-05-1	76.53	0.6	U	1.88		
1,4-Dioxane	123-91-1	88.12	0.84	U	3.03		
Methyl Methacrylate	80-62-6	100.117	0.34	U	1.39		

Project : Yellow Cab - TO-15

Sampling Date : 03/14/25

Field Id Number : 565-SG-1DL

Analysis Date : 03/17/25

Laboratory Id Number : Q1586-01DL

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	8.4	UD	41.5		
Chloromethane	74-87-3	50.49	3.9	UD	8.05		
Vinyl Chloride	75-01-4	62.5	0.56	UD	1.43		
Bromomethane	74-83-9	94.94	5.6	UD	21.7		
Chloroethane	75-00-3	64.52	6.8	UD	17.9		
Tetrahydrofuran	109-99-9	72.11	5.2	UD	15.3		
Trichlorofluoromethane	75-69-4	137.4	6.4	UD	36.0		
Dichlorotetrafluoroethane	76-14-2	170.9	3.5	UD	24.5		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	6.4	UD	49.0		
tert-Butyl alcohol	75-65-0	74.12	7.6	UD	23.0		
Heptane	142-82-5	100.2	830	ED	3401		
1,1-Dichloroethene	75-35-4	96.94	5.6	UD	22.2		
Acetone	67-64-1	58.08	9.6	UD	22.8		
Carbon Disulfide	75-15-0	76.14	6.4	UD	19.9		
Methyl tert-Butyl Ether	1634-04-4	88.15	4.8	UD	17.3		
Methylene Chloride	75-09-2	84.94	7.6	UD	26.4		
trans-1,2-Dichloroethene	156-60-5	96.94	7.6	UD	30.1		
1,1-Dichloroethane	75-34-3	98.96	5.2	UD	21.0		
Cyclohexane	110-82-7	84.16	1100	ED	3786		
2-Butanone	78-93-3	72.11	39.7	D	117		
Carbon Tetrachloride	56-23-5	153.8	0.44	UD	2.77		
cis-1,2-Dichloroethene	156-59-2	96.94	3.6	UD	14.3		
Chloroform	67-66-3	119.4	3.3	UD	16.1		
1,1,1-Trichloroethane	71-55-6	133.4	0.38	UD	2.07		
2,2,4-Trimethylpentane	540-84-1	114.2	4	UD	18.7		
Benzene	71-43-2	78.11	440	D	1405		
1,2-Dichloroethane	107-06-2	98.96	3.6	UD	14.6		
Trichloroethene	79-01-6	131.4	0.68	UD	3.65		
1,2-Dichloropropane	78-87-5	113	4.4	UD	20.3		
Bromodichloromethane	75-27-4	163.8	3.1	UD	20.8		
4-Methyl-2-Pentanone	108-10-1	100.2	3.7	UD	15.2		
Toluene	108-88-3	92.14	19.8	JD	74.6		
t-1,3-Dichloropropene	10061-02-6	111	2.4	UD	10.9		
cis-1,3-Dichloropropene	10061-01-5	111	2.6	UD	11.8		
1,1,2-Trichloroethane	79-00-5	133.4	2.8	UD	15.3		

Project : Yellow Cab - TO-15

Sampling Date : 03/14/25

Field Id Number : 565-SG-1DL

Analysis Date : 03/17/25

Laboratory Id Number : Q1586-01DL

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	2.6	UD	22.2		
1,2-Dibromoethane	106-93-4	187.9	2.9	UD	22.3		
Tetrachloroethene	127-18-4	165.8	21	D	142		
Chlorobenzene	108-90-7	112.6	3	UD	13.8		
Ethyl Benzene	100-41-4	106.2	8.8	JD	38.2		
m/p-Xylene	179601-23-1	106.2	34	JD	147		
o-Xylene	95-47-6	106.2	10.3	JD	44.7		
Styrene	100-42-5	104.1	4.4	UD	18.7		
Bromoform	75-25-2	252.8	2.3	UD	23.8		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.44	UD	3.02		
2-Chlorotoluene	95-49-8	126.6	4.4	UD	22.8		
1,3,5-Trimethylbenzene	108-67-8	120.2	4.4	UD	21.6		
1,2,4-Trimethylbenzene	95-63-6	120.2	3.1	UD	15.2		
1,3-Dichlorobenzene	541-73-1	147	3.2	UD	19.2		
1,4-Dichlorobenzene	106-46-7	147	2.1	UD	12.6		
1,2-Dichlorobenzene	95-50-1	147	3	UD	18.0		
1,2,4-Trichlorobenzene	120-82-1	181.5	3.4	UD	25.2		
Hexachloro-1,3-Butadiene	87-68-3	260.8	3.4	UD	36.3		
Naphthalene	91-20-3	128.17	3	UD	15.7		
1,3-Butadiene	106-99-0	54.09	5.2	UD	11.5		
4-Ethyltoluene	622-96-8	120.2	4.8	UD	23.6		
Hexane	110-54-3	86.17	4600	ED	16211		
Allyl Chloride	107-05-1	76.53	6	UD	18.8		
1,4-Dioxane	123-91-1	88.12	8.4	UD	30.3		
Methyl Methacrylate	80-62-6	100.117	3.4	UD	13.9		

Project : Yellow Cab - TO-15

Sampling Date : 03/14/25

Field Id Number : 565-SG-1DL2

Analysis Date : 03/17/25

Laboratory Id Number : Q1586-01DL2

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	130	UD	642		
Chloromethane	74-87-3	50.49	58.8	UD	121		
Vinyl Chloride	75-01-4	62.5	8.4	UD	21.5		
Bromomethane	74-83-9	94.94	84	UD	326		
Chloroethane	75-00-3	64.52	100	UD	263		
Tetrahydrofuran	109-99-9	72.11	78	UD	230		
Trichlorofluoromethane	75-69-4	137.4	96	UD	539		
Dichlorotetrafluoroethane	76-14-2	170.9	52.8	UD	369		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	96	UD	735		
tert-Butyl alcohol	75-65-0	74.12	110	UD	333		
Heptane	142-82-5	100.2	590	D	2417		
1,1-Dichloroethene	75-35-4	96.94	84	UD	333		
Acetone	67-64-1	58.08	140	UD	332		
Carbon Disulfide	75-15-0	76.14	96	UD	298		
Methyl tert-Butyl Ether	1634-04-4	88.15	72	UD	259		
Methylene Chloride	75-09-2	84.94	110	UD	382		
trans-1,2-Dichloroethene	156-60-5	96.94	110	UD	436		
1,1-Dichloroethane	75-34-3	98.96	78	UD	315		
Cyclohexane	110-82-7	84.16	780	D	2684		
2-Butanone	78-93-3	72.11	72	UD	212		
Carbon Tetrachloride	56-23-5	153.8	6.6	UD	41.5		
cis-1,2-Dichloroethene	156-59-2	96.94	53.4	UD	211		
Chloroform	67-66-3	119.4	49.8	UD	243		
1,1,1-Trichloroethane	71-55-6	133.4	5.7	UD	31.1		
2,2,4-Trimethylpentane	540-84-1	114.2	60	UD	280		
Benzene	71-43-2	78.11	420	D	1341		
1,2-Dichloroethane	107-06-2	98.96	54.6	UD	220		
Trichloroethene	79-01-6	131.4	10.2	UD	54.8		
1,2-Dichloropropane	78-87-5	113	66	UD	305		
Bromodichloromethane	75-27-4	163.8	46.2	UD	309		
4-Methyl-2-Pentanone	108-10-1	100.2	55.8	UD	228		
Toluene	108-88-3	92.14	66	UD	248		
t-1,3-Dichloropropene	10061-02-6	111	36	UD	163		
cis-1,3-Dichloropropene	10061-01-5	111	38.4	UD	174		
1,1,2-Trichloroethane	79-00-5	133.4	42.6	UD	232		

Project : Yellow Cab - TO-15

Sampling Date : 03/14/25

Field Id Number : 565-SG-1DL2

Analysis Date : 03/17/25

Laboratory Id Number : Q1586-01DL2

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	39.6	UD	337		
1,2-Dibromoethane	106-93-4	187.9	43.2	UD	332		
Tetrachloroethene	127-18-4	165.8	9	UD	61.0		
Chlorobenzene	108-90-7	112.6	45.6	UD	210		
Ethyl Benzene	100-41-4	106.2	72	UD	312		
m/p-Xylene	179601-23-1	106.2	130	UD	564		
o-Xylene	95-47-6	106.2	72	UD	312		
Styrene	100-42-5	104.1	66	UD	281		
Bromoform	75-25-2	252.8	34.2	UD	353		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	6.6	UD	45.3		
2-Chlorotoluene	95-49-8	126.6	66	UD	341		
1,3,5-Trimethylbenzene	108-67-8	120.2	66	UD	324		
1,2,4-Trimethylbenzene	95-63-6	120.2	46.8	UD	230		
1,3-Dichlorobenzene	541-73-1	147	48	UD	288		
1,4-Dichlorobenzene	106-46-7	147	31.8	UD	191		
1,2-Dichlorobenzene	95-50-1	147	45	UD	270		
1,2,4-Trichlorobenzene	120-82-1	181.5	51	UD	378		
Hexachloro-1,3-Butadiene	87-68-3	260.8	51	UD	544		
Naphthalene	91-20-3	128.17	45	UD	235		
1,3-Butadiene	106-99-0	54.09	78	UD	172		
4-Ethyltoluene	622-96-8	120.2	72	UD	353		
Hexane	110-54-3	86.17	6200	D	21850		
Allyl Chloride	107-05-1	76.53	90	UD	281		
1,4-Dioxane	123-91-1	88.12	130	UD	468		
Methyl Methacrylate	80-62-6	100.117	51.6	UD	211		

Project : Yellow Cab - TO-15

Field Id Number : 557-SG-1

Laboratory Id Number : Q1586-02

Sampling Date : 03/14/25

Analysis Date : 03/17/25

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.54		2.67		
Chloromethane	74-87-3	50.49	0.1	U	0.21		
Vinyl Chloride	75-01-4	62.5	0.01	U	0.03		
Bromomethane	74-83-9	94.94	0.14	U	0.54		
Chloroethane	75-00-3	64.52	0.17	U	0.45		
Tetrahydrofuran	109-99-9	72.11	0.13	U	0.38		
Trichlorofluoromethane	75-69-4	137.4	0.31	J	1.74		
Dichlorotetrafluoroethane	76-14-2	170.9	0.09	U	0.63		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.16	U	1.23		
tert-Butyl alcohol	75-65-0	74.12	0.24	J	0.73		
Heptane	142-82-5	100.2	0.11	U	0.45		
1,1-Dichloroethene	75-35-4	96.94	0.14	U	0.56		
Acetone	67-64-1	58.08	8.5		20.2		
Carbon Disulfide	75-15-0	76.14	1.5		4.67		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.12	U	0.43		
Methylene Chloride	75-09-2	84.94	0.5		1.74		
trans-1,2-Dichloroethene	156-60-5	96.94	0.19	U	0.75		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	0.68		2.01		
Carbon Tetrachloride	56-23-5	153.8	0.08		0.5		
cis-1,2-Dichloroethene	156-59-2	96.94	0.09	U	0.36		
Chloroform	67-66-3	119.4	0.78		3.81		
1,1,1-Trichloroethane	71-55-6	133.4	0.01	U	0.05		
2,2,4-Trimethylpentane	540-84-1	114.2	0.1	U	0.47		
Benzene	71-43-2	78.11	0.11	J	0.35		
1,2-Dichloroethane	107-06-2	98.96	0.09	U	0.36		
Trichloroethene	79-01-6	131.4	0.02	U	0.11		
1,2-Dichloropropane	78-87-5	113	0.11	U	0.51		
Bromodichloromethane	75-27-4	163.8	0.08	U	0.54		
4-Methyl-2-Pentanone	108-10-1	100.2	0.09	U	0.37		
Toluene	108-88-3	92.14	0.11	U	0.41		
t-1,3-Dichloropropene	10061-02-6	111	0.06	U	0.27		
cis-1,3-Dichloropropene	10061-01-5	111	0.06	U	0.27		
1,1,2-Trichloroethane	79-00-5	133.4	0.07	U	0.38		

Project : Yellow Cab - TO-15

Sampling Date : 03/14/25

Field Id Number : 557-SG-1

Analysis Date : 03/17/25

Laboratory Id Number : Q1586-02

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.07	U	0.6		
1,2-Dibromoethane	106-93-4	187.9	0.07	U	0.54		
Tetrachloroethene	127-18-4	165.8	3.1		21.0		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.12	U	0.52		
m/p-Xylene	179601-23-1	106.2	0.21	U	0.91		
o-Xylene	95-47-6	106.2	0.12	U	0.52		
Styrene	100-42-5	104.1	0.11	U	0.47		
Bromoform	75-25-2	252.8	0.06	U	0.62		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.01	U	0.07		
2-Chlorotoluene	95-49-8	126.6	0.11	U	0.57		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.11	U	0.54		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.08	U	0.39		
1,3-Dichlorobenzene	541-73-1	147	0.08	U	0.48		
1,4-Dichlorobenzene	106-46-7	147	0.05	U	0.3		
1,2-Dichlorobenzene	95-50-1	147	0.08	U	0.48		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.09	U	0.67		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.09	U	0.96		
Naphthalene	91-20-3	128.17	0.08	U	0.42		
1,3-Butadiene	106-99-0	54.09	0.13	U	0.29		
4-Ethyltoluene	622-96-8	120.2	0.12	U	0.59		
Hexane	110-54-3	86.17	0.29	J	1.02		
Allyl Chloride	107-05-1	76.53	0.15	U	0.47		
1,4-Dioxane	123-91-1	88.12	0.21	U	0.76		
Methyl Methacrylate	80-62-6	100.117	0.09	U	0.37		

Project : Yellow Cab - TO-15

Sampling Date : 03/14/25

Field Id Number : 557-SG-2

Analysis Date : 03/17/25

Laboratory Id Number : Q1586-03

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.56		2.77		
Chloromethane	74-87-3	50.49	0.29	J	0.6		
Vinyl Chloride	75-01-4	62.5	0.01	U	0.03		
Bromomethane	74-83-9	94.94	0.14	U	0.54		
Chloroethane	75-00-3	64.52	0.17	U	0.45		
Tetrahydrofuran	109-99-9	72.11	0.13	U	0.38		
Trichlorofluoromethane	75-69-4	137.4	0.3	J	1.69		
Dichlorotetrafluoroethane	76-14-2	170.9	0.09	U	0.63		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.16	U	1.23		
tert-Butyl alcohol	75-65-0	74.12	0.57		1.73		
Heptane	142-82-5	100.2	0.28	J	1.15		
1,1-Dichloroethene	75-35-4	96.94	0.14	U	0.56		
Acetone	67-64-1	58.08	15.2	E	36.1		
Carbon Disulfide	75-15-0	76.14	2.4		7.47		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.12	U	0.43		
Methylene Chloride	75-09-2	84.94	0.6		2.08		
trans-1,2-Dichloroethene	156-60-5	96.94	0.19	U	0.75		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	1.3		3.83		
Carbon Tetrachloride	56-23-5	153.8	0.09		0.57		
cis-1,2-Dichloroethene	156-59-2	96.94	0.14	J	0.56		
Chloroform	67-66-3	119.4	8.5		41.5		
1,1,1-Trichloroethane	71-55-6	133.4	0.01	U	0.05		
2,2,4-Trimethylpentane	540-84-1	114.2	0.1	U	0.47		
Benzene	71-43-2	78.11	3.5		11.2		
1,2-Dichloroethane	107-06-2	98.96	0.09	U	0.36		
Trichloroethene	79-01-6	131.4	0.47		2.53		
1,2-Dichloropropane	78-87-5	113	0.11	U	0.51		
Bromodichloromethane	75-27-4	163.8	0.32	J	2.14		
4-Methyl-2-Pentanone	108-10-1	100.2	0.09	U	0.37		
Toluene	108-88-3	92.14	0.2	J	0.75		
t-1,3-Dichloropropene	10061-02-6	111	0.06	U	0.27		
cis-1,3-Dichloropropene	10061-01-5	111	0.06	U	0.27		
1,1,2-Trichloroethane	79-00-5	133.4	0.07	U	0.38		

Project : Yellow Cab - TO-15

Sampling Date : 03/14/25

Field Id Number : 557-SG-2

Analysis Date : 03/17/25

Laboratory Id Number : Q1586-03

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.07	U	0.6		
1,2-Dibromoethane	106-93-4	187.9	0.07	U	0.54		
Tetrachloroethene	127-18-4	165.8	13.1		88.8		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.12	U	0.52		
m/p-Xylene	179601-23-1	106.2	0.21	U	0.91		
o-Xylene	95-47-6	106.2	0.12	U	0.52		
Styrene	100-42-5	104.1	0.11	U	0.47		
Bromoform	75-25-2	252.8	0.06	U	0.62		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.01	U	0.07		
2-Chlorotoluene	95-49-8	126.6	0.11	U	0.57		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.11	U	0.54		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.34	J	1.67		
1,3-Dichlorobenzene	541-73-1	147	0.08	U	0.48		
1,4-Dichlorobenzene	106-46-7	147	0.05	U	0.3		
1,2-Dichlorobenzene	95-50-1	147	0.08	U	0.48		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.09	U	0.67		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.09	U	0.96		
Naphthalene	91-20-3	128.17	0.77		4.04		
1,3-Butadiene	106-99-0	54.09	0.13	U	0.29		
4-Ethyltoluene	622-96-8	120.2	0.12	U	0.59		
Hexane	110-54-3	86.17	0.7		2.47		
Allyl Chloride	107-05-1	76.53	0.15	U	0.47		
1,4-Dioxane	123-91-1	88.12	0.21	U	0.76		
Methyl Methacrylate	80-62-6	100.117	0.09	U	0.37		

Project : Yellow Cab - TO-15

Sampling Date : 03/14/25

Field Id Number : 557-SG-2DL

Analysis Date : 03/17/25

Laboratory Id Number : Q1586-03DL

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	2.1	UD	10.4		
Chloromethane	74-87-3	50.49	0.98	UD	2.02		
Vinyl Chloride	75-01-4	62.5	0.14	UD	0.36		
Bromomethane	74-83-9	94.94	1.4	UD	5.44		
Chloroethane	75-00-3	64.52	1.7	UD	4.49		
Tetrahydrofuran	109-99-9	72.11	1.3	UD	3.83		
Trichlorofluoromethane	75-69-4	137.4	1.6	UD	8.99		
Dichlorotetrafluoroethane	76-14-2	170.9	0.88	UD	6.15		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	1.6	UD	12.3		
tert-Butyl alcohol	75-65-0	74.12	1.9	UD	5.76		
Heptane	142-82-5	100.2	1.1	UD	4.51		
1,1-Dichloroethene	75-35-4	96.94	1.4	UD	5.55		
Acetone	67-64-1	58.08	16	D	38.0		
Carbon Disulfide	75-15-0	76.14	1.6	UD	4.98		
Methyl tert-Butyl Ether	1634-04-4	88.15	1.2	UD	4.33		
Methylene Chloride	75-09-2	84.94	1.9	UD	6.6		
trans-1,2-Dichloroethene	156-60-5	96.94	1.9	UD	7.53		
1,1-Dichloroethane	75-34-3	98.96	1.3	UD	5.26		
Cyclohexane	110-82-7	84.16	2.2	UD	7.57		
2-Butanone	78-93-3	72.11	1.6	JD	4.72		
Carbon Tetrachloride	56-23-5	153.8	0.11	UD	0.69		
cis-1,2-Dichloroethene	156-59-2	96.94	0.89	UD	3.53		
Chloroform	67-66-3	119.4	8.5	D	41.5		
1,1,1-Trichloroethane	71-55-6	133.4	0.1	UD	0.55		
2,2,4-Trimethylpentane	540-84-1	114.2	1	UD	4.67		
Benzene	71-43-2	78.11	3.1	JD	9.9		
1,2-Dichloroethane	107-06-2	98.96	0.91	UD	3.68		
Trichloroethene	79-01-6	131.4	0.17	UD	0.91		
1,2-Dichloropropane	78-87-5	113	1.1	UD	5.08		
Bromodichloromethane	75-27-4	163.8	0.77	UD	5.16		
4-Methyl-2-Pentanone	108-10-1	100.2	0.93	UD	3.81		
Toluene	108-88-3	92.14	1.1	UD	4.15		
t-1,3-Dichloropropene	10061-02-6	111	0.6	UD	2.72		
cis-1,3-Dichloropropene	10061-01-5	111	0.64	UD	2.91		
1,1,2-Trichloroethane	79-00-5	133.4	0.71	UD	3.87		

Project : Yellow Cab - TO-15

Sampling Date : 03/14/25

Field Id Number : 557-SG-2DL

Analysis Date : 03/17/25

Laboratory Id Number : Q1586-03DL

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.66	UD	5.62		
1,2-Dibromoethane	106-93-4	187.9	0.72	UD	5.53		
Tetrachloroethene	127-18-4	165.8	11.7	D	79.3		
Chlorobenzene	108-90-7	112.6	0.76	UD	3.5		
Ethyl Benzene	100-41-4	106.2	1.2	UD	5.21		
m/p-Xylene	179601-23-1	106.2	2.1	UD	9.12		
o-Xylene	95-47-6	106.2	1.2	UD	5.21		
Styrene	100-42-5	104.1	1.1	UD	4.68		
Bromoform	75-25-2	252.8	0.57	UD	5.89		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.11	UD	0.76		
2-Chlorotoluene	95-49-8	126.6	1.1	UD	5.7		
1,3,5-Trimethylbenzene	108-67-8	120.2	1.1	UD	5.41		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.78	UD	3.83		
1,3-Dichlorobenzene	541-73-1	147	0.8	UD	4.81		
1,4-Dichlorobenzene	106-46-7	147	0.53	UD	3.19		
1,2-Dichlorobenzene	95-50-1	147	0.75	UD	4.51		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.85	UD	6.31		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.85	UD	9.07		
Naphthalene	91-20-3	128.17	0.75	UD	3.93		
1,3-Butadiene	106-99-0	54.09	1.3	UD	2.88		
4-Ethyltoluene	622-96-8	120.2	1.2	UD	5.9		
Hexane	110-54-3	86.17	1.1	UD	3.88		
Allyl Chloride	107-05-1	76.53	1.5	UD	4.7		
1,4-Dioxane	123-91-1	88.12	2.1	UD	7.57		
Methyl Methacrylate	80-62-6	100.117	0.86	UD	3.52		

Method	Matrix	Parameter	Spike_Amt_ Added	Calculated MDL	MB Data	MB Datafile	New MDL TO Be Used	MDL Unit
TO-15	Air	1,1,1,2-Tetrachloroethane	0.4	0.090683682	0		0.091	ppbv
TO-15	Air	1,1,1-Trichloroethane	0.03	0.00947589	0		0.0095	ppbv
TO-15	Air	1,1,2,2-Tetrachloroethane	0.03	0.010183321	0		0.011	ppbv
TO-15	Air	1,1,2-Trichloroethane	0.4	0.070112514	0		0.071	ppbv
TO-15	Air	1,1,2-Trichlorotrifluoroethane	0.4	0.155560652	0		0.16	ppbv
TO-15	Air	1,1-Dichloroethane	0.4	0.122378743	0		0.13	ppbv
TO-15	Air	1,1-Dichloroethene	0.4	0.139735617	0		0.14	ppbv
TO-15	Air	1,2,4-Trichlorobenzene	0.4	0.084677012	0		0.085	ppbv
TO-15	Air	1,2,4-Trimethylbenzene	0.4	0.077496367	0		0.078	ppbv
TO-15	Air	1,2-Dibromoethane	0.4	0.071318853	0		0.072	ppbv
TO-15	Air	1,2-Dichlorobenzene	0.4	0.074250332	0		0.075	ppbv
TO-15	Air	1,2-Dichloroethane	0.4	0.090890005	0		0.091	ppbv
TO-15	Air	1,2-Dichloropropane	0.4	0.073346868	0.109	VL038919.D	0.11	ppbv
TO-15	Air	1,3,5-Trimethylbenzene	0.4	0.101198372	0		0.11	ppbv
TO-15	Air	1,3-Butadiene	0.4	0.124009143	0		0.13	ppbv
TO-15	Air	1,3-Dichlorobenzene	0.4	0.079013305	0		0.08	ppbv
TO-15	Air	1,4-Dichlorobenzene	0.4	0.052746151	0		0.053	ppbv
TO-15	Air	1,4-Dioxane	0.4	0.202123585	0		0.21	ppbv
TO-15	Air	2,2,4-Trimethylpentane	0.4	0.09988427	0		0.10	ppbv
TO-15	Air	2-Butanone	0.4	0.118687999	0		0.12	ppbv
TO-15	Air	2-Chlorotoluene	0.4	0.104141644	0		0.11	ppbv
TO-15	Air	2-Hexanone	0.4	0.080656229	0		0.081	ppbv
TO-15	Air	4-Ethyltoluene	0.4	0.112037051	0		0.12	ppbv
TO-15	Air	4-Methyl-2-Pentanone	0.4	0.092836933	0		0.093	ppbv
TO-15	Air	Acetone	0.4	0.234416891	0		0.24	ppbv
TO-15	Air	Allyl Chloride	0.4	0.145618654	0		0.15	ppbv
TO-15	Air	Benzene	0.4	0.086407689	0		0.087	ppbv
TO-15	Air	Benzyl Chloride	0.4	0.173895169	0		0.18	ppbv
TO-15	Air	Bromodichloromethane	0.4	0.076780313	0		0.077	ppbv
TO-15	Air	Bromoethene	0.4	0.115560124	0		0.12	ppbv
TO-15	Air	Bromoform	0.4	0.056947551	0		0.057	ppbv
TO-15	Air	Bromomethane	0.4	0.137678239	0		0.14	ppbv
TO-15	Air	Carbon Disulfide	0.4	0.157343925	0		0.16	ppbv
TO-15	Air	Carbon Tetrachloride	0.03	0.010617308	0		0.011	ppbv
TO-15	Air	Chlorobenzene	0.4	0.075202602	0		0.076	ppbv
TO-15	Air	Chlorodifluoromethane	0.4	0.087466815	0		0.088	ppbv
TO-15	Air	Chloroethane	0.4	0.163530318	0		0.17	ppbv
TO-15	Air	Chloroform	0.4	0.082232394	0		0.083	ppbv
TO-15	Air	Chloromethane	0.4	0.097292998	0		0.098	ppbv
TO-15	Air	cis-1,2-Dichloroethene	0.4	0.088332535	0		0.089	ppbv
TO-15	Air	cis-1,3-Dichloropropene	0.4	0.063248763	0		0.064	ppbv
TO-15	Air	Cyclohexane	0.4	0.217999816	0		0.22	ppbv
TO-15	Air	Dibromochloromethane	0.4	0.065703935	0		0.066	ppbv
TO-15	Air	Dichlorodifluoromethane	0.4	0.20957733	0		0.21	ppbv
TO-15	Air	Dichlorotetrafluoroethane	0.4	0.087348046	0		0.088	ppbv
TO-15	Air	Ethyl Acetate	0.4	0.078161123	0		0.079	ppbv
TO-15	Air	Ethyl Benzene	0.4	0.115413562	0		0.12	ppbv
TO-15	Air	Heptane	0.4	0.100800637	0		0.11	ppbv
TO-15	Air	Hexachloro-1,3-Butadiene	0.4	0.084169773	0		0.085	ppbv
TO-15	Air	Hexane	0.4	0.106485082	0		0.11	ppbv
TO-15	Air	Isopropyl Alcohol	0.4	0.068329134	0		0.069	ppbv
TO-15	Air	Isopropylbenzene	0.4	0.086744608	0		0.087	ppbv
TO-15	Air	m/p-Xylene	0.4	0.205831125	0		0.21	ppbv
TO-15	Air	Methyl Methacrylate	0.4	0.085528753	0		0.086	ppbv
TO-15	Air	Methyl tert-Butyl Ether	0.4	0.119546728	0		0.12	ppbv
TO-15	Air	Methylene Chloride	0.4	0.18405265	0.11	VL039022.D	0.19	ppbv
TO-15	Air	Naphthalene	0.4	0.074961199	0		0.075	ppbv
TO-15	Air	Naphthalene,2-methyl-	0.4	0.148496389	0		0.15	ppbv
TO-15	Air	n-Butylbenzene	0.4	0.08201719	0		0.083	ppbv
TO-15	Air	n-propylbenzene	0.4	0.0778963	0		0.078	ppbv
TO-15	Air	o-Xylene	0.4	0.117630235	0		0.12	ppbv
TO-15	Air	p-Isopropyltoluene	0.4	0.085704866	0		0.086	ppbv
TO-15	Air	Propene	0.4	0.143790814	0.14	VL039337.D	0.15	ppbv
TO-15	Air	sec-Butylbenzene	0.4	0.093164076	0		0.094	ppbv
TO-15	Air	Styrene	0.4	0.109444442	0		0.11	ppbv
TO-15	Air	t-1,3-Dichloropropene	0.4	0.059734415	0		0.06	ppbv
TO-15	Air	tert-Butyl alcohol	0.4	0.187711954	0		0.19	ppbv
TO-15	Air	tert-Butylbenzene	0.4	0.092760327	0		0.093	ppbv
TO-15	Air	Tetrachloroethene	0.03	0.014462962	0		0.015	ppbv
TO-15	Air	Tetrahydrofuran	0.4	0.122655048	0		0.13	ppbv
TO-15	Air	Toluene	0.4	0.109284	0		0.11	ppbv
TO-15	Air	trans-1,2-Dichloroethene	0.4	0.185264638	0		0.19	ppbv
TO-15	Air	Trichloroethene	0.03	0.016558175	0		0.017	ppbv
TO-15	Air	Trichlorofluoromethane	0.4	0.157465794	0		0.16	ppbv
TO-15	Air	Vinyl Acetate	0.4	0.152371485	0		0.16	ppbv
TO-15	Air	Vinyl Chloride	0.03	0.013585954	0		0.014	ppbv
TO-15	Air	Ethanol	0.4	0.317031286	0		0.32	ug/L

StddevVal	StudentsT Value	Instrument ID	Datafile1	Concentration1	Adate1	Datafile2	Concentration2
0.03485153	2.602	MSVOA_L	VL038602.D	0.51	03142022T18:52:00	VL038645.D	0.494
0.003641772	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.029
0.003913651	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.035
0.026945624	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.439
0.059785032	2.602	MSVOA_L	VL038602.D	0.47	03142022T18:52:00	VL038645.D	0.466
0.047032569	2.602	MSVOA_L	VL038602.D	0.45	03142022T18:52:00	VL038645.D	0.456
0.053703158	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.411
0.032543048	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.336
0.029783385	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.396
0.027409244	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.352
0.028535869	2.602	MSVOA_L	VL038602.D	0.45	03142022T18:52:00	VL038645.D	0.407
0.034930824	2.602	MSVOA_L	VL038602.D	0.41	03142022T18:52:00	VL038645.D	0.445
0.02818865	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.406
0.038892534	2.602	MSVOA_L	VL038602.D	0.33	03142022T18:52:00	VL038645.D	0.374
0.047659163	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.419
0.030366374	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.428
0.020271388	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.418
0.077680086	2.602	MSVOA_L	VL038602.D	0.46	03142022T18:52:00	VL038645.D	0.285
0.038387498	2.602	MSVOA_L	VL038602.D	0.39	03142022T18:52:00	VL038645.D	0.383
0.045614143	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.406
0.040023691	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.375
0.030997782	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.283
0.043058052	2.602	MSVOA_L	VL038602.D	0.34	03142022T18:52:00	VL038645.D	0.356
0.035679067	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.282
0.090091042	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.468
0.055964125	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.392
0.033208182	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.407
0.066831349	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.454
0.029508191	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.408
0.044412039	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.433
0.021886069	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.31
0.052912467	2.602	MSVOA_L	VL038602.D	0.43	03142022T18:52:00	VL038645.D	0.404
0.060470378	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.357
0.004080441	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.026
0.028901845	2.602	MSVOA_L	VL038602.D	0.48	03142022T18:52:00	VL038645.D	0.467
0.033615225	2.602	MSVOA_L	VL038602.D	0.45	03142022T18:52:00	VL038645.D	0.468
0.062847932	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.453
0.031603534	2.602	MSVOA_L	VL038602.D	0.45	03142022T18:52:00	VL038645.D	0.464
0.037391621	2.602	MSVOA_L	VL038602.D	0.47	03142022T18:52:00	VL038645.D	0.439
0.033947938	2.602	MSVOA_L	VL038602.D	0.36	03142022T18:52:00	VL038645.D	0.391
0.024307749	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.317
0.083781636	2.602	MSVOA_L	VL038602.D	0.42	03142022T18:52:00	VL038645.D	0.414
0.02525132	2.602	MSVOA_L	VL038602.D	0.39	03142022T18:52:00	VL038645.D	0.359
0.080544708	2.602	MSVOA_L	VL038602.D	0.49	03142022T18:52:00	VL038645.D	0.43
0.033569579	2.602	MSVOA_L	VL038602.D	0.5	03142022T18:52:00	VL038645.D	0.472
0.030038864	2.602	MSVOA_L	VL038602.D	0.39	03142022T18:52:00	VL038645.D	0.412
0.044355712	2.602	MSVOA_L	VL038602.D	0.33	03142022T18:52:00	VL038645.D	0.357
0.038739676	2.602	MSVOA_L	VL038602.D	0.35	03142022T18:52:00	VL038645.D	0.349
0.032348107	2.602	MSVOA_L	VL038602.D	0.48	03142022T18:52:00	VL038645.D	0.5
0.04092432	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.411
0.026260236	2.602	MSVOA_L	VL038602.D	0.47	03142022T18:52:00	VL038645.D	0.44
0.033337666	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.387
0.079104967	2.602	MSVOA_L	VL038602.D	0.76	03142022T18:52:00	VL038645.D	0.762
0.032870389	2.602	MSVOA_L	VL038602.D	0.31	03142022T18:52:00	VL038645.D	0.334
0.045944169	2.602	MSVOA_L	VL038602.D	0.41	03142022T18:52:00	VL038645.D	0.418
0.070735069	2.602	MSVOA_L	VL038602.D	0.49	03142022T18:52:00	VL038645.D	0.479
0.02880907	2.602	MSVOA_L	VL038602.D	0.25	03142022T18:52:00	VL038645.D	0.289
0.057070096	2.602	MSVOA_L	VL038602.D	0.14	03142022T18:52:00	VL038645.D	0.074
0.031520826	2.602	MSVOA_L	VL038602.D	0.35	03142022T18:52:00	VL038645.D	0.366
0.029937087	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.365
0.045207623	2.602	MSVOA_L	VL038602.D	0.36	03142022T18:52:00	VL038645.D	0.368
0.032938073	2.602	MSVOA_L	VL038602.D	0.35	03142022T18:52:00	VL038645.D	0.366
0.05526165	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.443
0.035804795	2.602	MSVOA_L	VL038602.D	0.38	03142022T18:52:00	VL038645.D	0.392
0.042061661	2.602	MSVOA_L	VL038602.D	0.26	03142022T18:52:00	VL038645.D	0.275
0.022957116	2.602	MSVOA_L	VL038602.D	0.27	03142022T18:52:00	VL038645.D	0.29
0.072141412	2.602	MSVOA_L	VL038602.D	0.44	03142022T18:52:00	VL038645.D	0.459
0.035649626	2.602	MSVOA_L	VL038602.D	0.37	03142022T18:52:00	VL038645.D	0.405
0.005558402	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.031
0.047138758	2.602	MSVOA_L	VL038602.D	0.3	03142022T18:52:00	VL038645.D	0.324
0.042	2.602	MSVOA_L	VL038602.D	0.31	03142022T18:52:00	VL038645.D	0.332
0.07120086	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.396
0.006363634	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.035
0.060517215	2.602	MSVOA_L	VL038602.D	0.46	03142022T18:52:00	VL038645.D	0.472
0.058559372	2.602	MSVOA_L	VL038602.D	0.4	03142022T18:52:00	VL038645.D	0.389
0.00522135	2.602	MSVOA_L	VL038599.D	0.03	03142022T17:01:00	VL038658.D	0.031
0.105747594	2.998	MSVOA_L	VL040210.D	0.438	02222023T00:52:00	VL040220.D	0.43

- A
- B
- C
- D**
- E
- F
- G
- H

Concentration5	Adate5	Datafile6	Concentration6	Adate6	Datafile7	Concentration7	Adate7
0.433	08182022T21:04:00	VL039600.D	0.463	08262022T12:33:00	VL039709.D	0.402	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.031	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.03	08182022T22:24:00	VL039586.D	0.029	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.404	08182022T21:04:00	VL039600.D	0.424	08262022T12:33:00	VL039709.D	0.412	10102022T23:30:00
0.448	08182022T21:04:00	VL039600.D	0.454	08262022T12:33:00	VL039709.D	0.438	10102022T23:30:00
0.426	08182022T21:04:00	VL039600.D	0.451	08262022T12:33:00	VL039709.D	0.445	10102022T23:30:00
0.394	08182022T21:04:00	VL039600.D	0.387	08262022T12:33:00	VL039709.D	0.427	10102022T23:30:00
0.314	08182022T21:04:00	VL039600.D	0.367	08262022T12:33:00	VL039709.D	0.373	10102022T23:30:00
0.342	08182022T21:04:00	VL039600.D	0.389	08262022T12:33:00	VL039709.D	0.389	10102022T23:30:00
0.361	08182022T21:04:00	VL039600.D	0.402	08262022T12:33:00	VL039709.D	0.375	10102022T23:30:00
0.408	08182022T21:04:00	VL039600.D	0.431	08262022T12:33:00	VL039709.D	0.448	10102022T23:30:00
0.411	08182022T21:04:00	VL039600.D	0.455	08262022T12:33:00	VL039709.D	0.418	10102022T23:30:00
0.444	08182022T21:04:00	VL039600.D	0.429	08262022T12:33:00	VL039709.D	0.373	10102022T23:30:00
0.331	08182022T21:04:00	VL039600.D	0.357	08262022T12:33:00	VL039709.D	0.342	10102022T23:30:00
0.405	08182022T21:04:00	VL039600.D	0.449	08262022T12:33:00	VL039709.D	0.422	10102022T23:30:00
0.368	08182022T21:04:00	VL039600.D	0.414	08262022T12:33:00	VL039709.D	0.399	10102022T23:30:00
0.404	08182022T21:04:00	VL039600.D	0.384	08262022T12:33:00	VL039709.D	0.439	10102022T23:30:00
0.373	08182022T21:04:00	VL039600.D	0.341	08262022T12:33:00	VL039709.D	0.418	10102022T23:30:00
0.348	08182022T21:04:00	VL039600.D	0.379	08262022T12:33:00	VL039709.D	0.392	10102022T23:30:00
0.314	08182022T21:04:00	VL039600.D	0.451	08262022T12:33:00	VL039709.D	0.354	10102022T23:30:00
0.312	08182022T21:04:00	VL039600.D	0.359	08262022T12:33:00	VL039709.D	0.348	10102022T23:30:00
0.27	08182022T21:04:00	VL039600.D	0.294	08262022T12:33:00	VL039709.D	0.326	10102022T23:30:00
0.284	08182022T21:04:00	VL039600.D	0.348	08262022T12:33:00	VL039709.D	0.355	10102022T23:30:00
0.322	08182022T21:04:00	VL039600.D	0.335	08262022T12:33:00	VL039709.D	0.32	10102022T23:30:00
0.363	08182022T21:04:00	VL039600.D	0.484	08262022T12:33:00	VL039709.D	0.5	10102022T23:30:00
0.418	08182022T21:04:00	VL039600.D	0.459	08262022T12:33:00	VL039709.D	0.379	10102022T23:30:00
0.345	08182022T21:04:00	VL039600.D	0.376	08262022T12:33:00	VL039709.D	0.398	10102022T23:30:00
0.394	08182022T21:04:00	VL039600.D	0.336	08262022T12:33:00	VL039709.D	0.356	10102022T23:30:00
0.382	08182022T21:04:00	VL039600.D	0.367	08262022T12:33:00	VL039709.D	0.377	10102022T23:30:00
0.408	08182022T21:04:00	VL039600.D	0.441	08262022T12:33:00	VL039709.D	0.388	10102022T23:30:00
0.305	08182022T21:04:00	VL039600.D	0.303	08262022T12:33:00	VL039709.D	0.293	10102022T23:30:00
0.364	08182022T21:04:00	VL039600.D	0.494	08262022T12:33:00	VL039709.D	0.464	10102022T23:30:00
0.399	08182022T21:04:00	VL039600.D	0.268	08262022T12:33:00	VL039709.D	0.332	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.027	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.431	08182022T21:04:00	VL039600.D	0.466	08262022T12:33:00	VL039709.D	0.46	10102022T23:30:00
0.456	08182022T21:04:00	VL039600.D	0.504	08262022T12:33:00	VL039709.D	0.446	10102022T23:30:00
0.462	08182022T21:04:00	VL039600.D	0.504	08262022T12:33:00	VL039709.D	0.531	10102022T23:30:00
0.423	08182022T21:04:00	VL039600.D	0.46	08262022T12:33:00	VL039709.D	0.411	10102022T23:30:00
0.392	08182022T21:04:00	VL039600.D	0.472	08262022T12:33:00	VL039709.D	0.485	10102022T23:30:00
0.36	08182022T21:04:00	VL039600.D	0.407	08262022T12:33:00	VL039709.D	0.438	10102022T23:30:00
0.277	08182022T21:04:00	VL039600.D	0.291	08262022T12:33:00	VL039709.D	0.284	10102022T23:30:00
0.376	08182022T21:04:00	VL039600.D	0.395	08262022T12:33:00	VL039709.D	0.383	10102022T23:30:00
0.339	08182022T21:04:00	VL039600.D	0.335	08262022T12:33:00	VL039709.D	0.301	10102022T23:30:00
0.62	08182022T21:04:00	VL039600.D	0.643	08262022T12:33:00	VL039709.D	0.694	10102022T23:30:00
0.454	08182022T21:04:00	VL039600.D	0.508	08262022T12:33:00	VL039709.D	0.473	10102022T23:30:00
0.341	08182022T21:04:00	VL039600.D	0.432	08262022T12:33:00	VL039709.D	0.395	10102022T23:30:00
0.297	08182022T21:04:00	VL039600.D	0.338	08262022T12:33:00	VL039709.D	0.334	10102022T23:30:00
0.347	08182022T21:04:00	VL039600.D	0.382	08262022T12:33:00	VL039709.D	0.407	10102022T23:30:00
0.447	08182022T21:04:00	VL039600.D	0.502	08262022T12:33:00	VL039709.D	0.468	10102022T23:30:00
0.385	08182022T21:04:00	VL039600.D	0.405	08262022T12:33:00	VL039709.D	0.435	10102022T23:30:00
0.458	08182022T21:04:00	VL039600.D	0.434	08262022T12:33:00	VL039709.D	0.481	10102022T23:30:00
0.361	08182022T21:04:00	VL039600.D	0.388	08262022T12:33:00	VL039709.D	0.405	10102022T23:30:00
0.673	08182022T21:04:00	VL039600.D	0.719	08262022T12:33:00	VL039709.D	0.688	10102022T23:30:00
0.3	08182022T21:04:00	VL039600.D	0.323	08262022T12:33:00	VL039709.D	0.346	10102022T23:30:00
0.337	08182022T21:04:00	VL039600.D	0.355	08262022T12:33:00	VL039709.D	0.396	10102022T23:30:00
0.536	08182022T21:04:00	VL039600.D	0.634	08262022T12:33:00	VL039709.D	0.499	10102022T23:30:00
0.28	08182022T21:04:00	VL039600.D	0.302	08262022T12:33:00	VL039709.D	0.291	10102022T23:30:00
0.158	08182022T21:04:00	VL039600.D	0.094	08262022T12:33:00	VL039709.D	0.149	10102022T23:30:00
0.321	08182022T21:04:00	VL039600.D	0.354	08262022T12:33:00	VL039709.D	0.358	10102022T23:30:00
0.332	08182022T21:04:00	VL039600.D	0.338	08262022T12:33:00	VL039709.D	0.347	10102022T23:30:00
0.315	08182022T21:04:00	VL039600.D	0.347	08262022T12:33:00	VL039709.D	0.354	10102022T23:30:00
0.316	08182022T21:04:00	VL039600.D	0.362	08262022T12:33:00	VL039709.D	0.358	10102022T23:30:00
0.499	08182022T21:04:00	VL039600.D	0.578	08262022T12:33:00	VL039709.D	0.533	10102022T23:30:00
0.359	08182022T21:04:00	VL039600.D	0.382	08262022T12:33:00	VL039709.D	0.371	10102022T23:30:00
0.23	08182022T21:04:00	VL039600.D	0.257	08262022T12:33:00	VL039709.D	0.278	10102022T23:30:00
0.252	08182022T21:04:00	VL039600.D	0.252	08262022T12:33:00	VL039709.D	0.273	10102022T23:30:00
0.384	08182022T21:04:00	VL039600.D	0.382	08262022T12:33:00	VL039709.D	0.454	10102022T23:30:00
0.345	08182022T21:04:00	VL039600.D	0.399	08262022T12:33:00	VL039709.D	0.371	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.038	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.299	08182022T21:04:00	VL039600.D	0.344	08262022T12:33:00	VL039709.D	0.389	10102022T23:30:00
0.288	08182022T21:04:00	VL039600.D	0.296	08262022T12:33:00	VL039709.D	0.33	10102022T23:30:00
0.375	08182022T21:04:00	VL039600.D	0.466	08262022T12:33:00	VL039709.D	0.437	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.031	08252022T20:23:00	VL039711.D	0.04	10112022T00:45:00
0.436	08182022T21:04:00	VL039600.D	0.47	08262022T12:33:00	VL039709.D	0.385	10102022T23:30:00
0.42	08182022T21:04:00	VL039600.D	0.483	08262022T12:33:00	VL039709.D	0.363	10102022T23:30:00
0.03	08182022T22:24:00	VL039586.D	0.029	08252022T20:23:00	VL039711.D	0.03	10112022T00:45:00
0.636	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.607	10042023T19:33:00

Datafile8	Concentration8	Adate8	Datafile9	Concentration9	Adate9	Datafile10	Concentration10
VL040134.D	0.42	01302023T15:54:00	VL040210.D	0.383	02222023T00:52:00	VL040220.D	0.42
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.038	02232023T14:26:00	VL040223.D	0.033
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.032	02232023T14:26:00	VL040223.D	0.027
VL040134.D	0.41	01302023T15:54:00	VL040210.D	0.395	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.476	02222023T00:52:00	VL040220.D	0.47
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.436	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.452	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.32	01302023T15:54:00	VL040210.D	0.299	02222023T00:52:00	VL040220.D	0.32
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.375	02222023T00:52:00	VL040220.D	0.37
VL040134.D	0.39	01302023T15:54:00	VL040210.D	0.38	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.378	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.434	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.41	01302023T15:54:00	VL040210.D	0.432	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.344	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.43	01302023T15:54:00	VL040210.D	0.396	02222023T00:52:00	VL040220.D	0.41
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.395	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.402	02222023T00:52:00	VL040220.D	0.4
VL040134.D	0.41	01302023T15:54:00	VL040210.D	0.387	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.39	01302023T15:54:00	VL040210.D	0.352	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.416	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.34	02222023T00:52:00	VL040220.D	0.36
VL040134.D	0.32	01302023T15:54:00	VL040210.D	0.318	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.321	02222023T00:52:00	VL040220.D	0.36
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.337	02222023T00:52:00	VL040220.D	0.37
VL040134.D	0.45	01302023T15:54:00	VL040210.D	0.482	02222023T00:52:00	VL040220.D	0.5
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.411	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.359	02222023T00:52:00	VL040220.D	0.39
VL040134.D	0.29	01302023T15:54:00	VL040210.D	0.311	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.378	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.413	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.32	01302023T15:54:00	VL040210.D	0.32	02222023T00:52:00	VL040220.D	0.33
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.408	02222023T00:52:00	VL040220.D	0.49
VL040134.D	0.31	01302023T15:54:00	VL040210.D	0.391	02222023T00:52:00	VL040220.D	0.37
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.033	02232023T14:26:00	VL040223.D	0.037
VL040134.D	0.44	01302023T15:54:00	VL040210.D	0.414	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.43	01302023T15:54:00	VL040210.D	0.474	02222023T00:52:00	VL040220.D	0.46
VL040134.D	0.37	01302023T15:54:00	VL040210.D	0.416	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.47	01302023T15:54:00	VL040210.D	0.445	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.46	01302023T15:54:00	VL040210.D	0.443	02222023T00:52:00	VL040220.D	0.47
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.435	02222023T00:52:00	VL040220.D	0.38
VL040134.D	0.33	01302023T15:54:00	VL040210.D	0.284	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.49	01302023T15:54:00	VL040210.D	0.521	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.343	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.48	01302023T15:54:00	VL040210.D	0.497	02222023T00:52:00	VL040220.D	0.49
VL040134.D	0.43	01302023T15:54:00	VL040210.D	0.431	02222023T00:52:00	VL040220.D	0.46
VL040134.D	0.39	01302023T15:54:00	VL040210.D	0.398	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.309	02222023T00:52:00	VL040220.D	0.33
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.336	02222023T00:52:00	VL040220.D	0.37
VL040134.D	0.43	01302023T15:54:00	VL040210.D	0.421	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.41	01302023T15:54:00	VL040210.D	0.4	02222023T00:52:00	VL040220.D	0.41
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.479	02222023T00:52:00	VL040220.D	0.48
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.361	02222023T00:52:00	VL040220.D	0.4
VL040134.D	0.71	01302023T15:54:00	VL040210.D	0.699	02222023T00:52:00	VL040220.D	0.75
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.314	02222023T00:52:00	VL040220.D	0.35
VL040134.D	0.38	01302023T15:54:00	VL040210.D	0.421	02222023T00:52:00	VL040220.D	0.41
VL040134.D	0.44	01302023T15:54:00	VL040210.D	0.558	02222023T00:52:00	VL040220.D	0.57
VL040134.D	0.27	01302023T15:54:00	VL040210.D	0.246	02222023T00:52:00	VL040220.D	0.27
VL040134.D	0.19	01302023T15:54:00	VL040210.D	0.116	02222023T00:52:00	VL040220.D	0.21
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.349	02222023T00:52:00	VL040220.D	0.35
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.354	02222023T00:52:00	VL040220.D	0.36
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.331	02222023T00:52:00	VL040220.D	0.36
VL040134.D	0.33	01302023T15:54:00	VL040210.D	0.329	02222023T00:52:00	VL040220.D	0.34
VL040134.D	0.42	01302023T15:54:00	VL040210.D	0.432	02222023T00:52:00	VL040220.D	0.43
VL040134.D	0.36	01302023T15:54:00	VL040210.D	0.352	02222023T00:52:00	VL040220.D	0.37
VL040134.D	0.29	01302023T15:54:00	VL040210.D	0.262	02222023T00:52:00	VL040220.D	0.27
VL040134.D	0.3	01302023T15:54:00	VL040210.D	0.292	02222023T00:52:00	VL040220.D	0.3
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.395	02222023T00:52:00	VL040220.D	0.44
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.349	02222023T00:52:00	VL040220.D	0.36
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.029	02232023T14:26:00	VL040223.D	0.031
VL040134.D	0.42	01302023T15:54:00	VL040210.D	0.368	02222023T00:52:00	VL040220.D	0.42
VL040134.D	0.34	01302023T15:54:00	VL040210.D	0.297	02222023T00:52:00	VL040220.D	0.33
VL040134.D	0.35	01302023T15:54:00	VL040210.D	0.397	02222023T00:52:00	VL040220.D	0.42
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.027	02232023T14:26:00	VL040223.D	0.028
VL040134.D	0.4	01302023T15:54:00	VL040210.D	0.451	02222023T00:52:00	VL040220.D	0.45
VL040134.D	0.39	01302023T15:54:00	VL040210.D	0.438	02222023T00:52:00	VL040220.D	0.4
VL040136.D	0.03	01302023T17:08:00	VL040222.D	0.033	02232023T14:26:00	VL040223.D	0.038
VL040805.D	0.675	10252023T18:12:00					

Adate10	Datafile11	Concentration11	Adate11	Datafile12	Concentration12	Adate12	Datafile13
02232023T13:12:00	VL040329.D	0.45	04102023T17:39:00	VL040349.D	0.434	04122023T23:23:00	VL040559.D
02232023T15:04:00	VL040332.D	0.028	04102023T19:33:00	VL040346.D	0.027	04122023T19:50:00	VL040576.D
02232023T15:04:00	VL040332.D	0.023	04102023T19:33:00	VL040346.D	0.03	04122023T19:50:00	VL040576.D
02232023T13:12:00	VL040329.D	0.41	04102023T17:39:00	VL040349.D	0.403	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.42	04102023T17:39:00	VL040349.D	0.429	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.43	04102023T17:39:00	VL040349.D	0.4	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.44	04102023T17:39:00	VL040349.D	0.419	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.3	04102023T17:39:00	VL040349.D	0.302	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.34	04102023T17:39:00	VL040349.D	0.362	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.37	04102023T17:39:00	VL040349.D	0.398	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.38	04102023T17:39:00	VL040349.D	0.402	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.45	04102023T17:39:00	VL040349.D	0.427	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.41	04102023T17:39:00	VL040349.D	0.385	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.33	04102023T17:39:00	VL040349.D	0.311	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.38	04102023T17:39:00	VL040349.D	0.448	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.38	04102023T17:39:00	VL040349.D	0.392	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.39	04102023T17:39:00	VL040349.D	0.402	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.33	04102023T17:39:00	VL040349.D	0.33	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.35	04102023T17:39:00	VL040349.D	0.356	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.45	04102023T17:39:00	VL040349.D	0.451	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.32	04102023T17:39:00	VL040349.D	0.31	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.28	04102023T17:39:00	VL040349.D	0.29	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.32	04102023T17:39:00	VL040349.D	0.303	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.33	04102023T17:39:00	VL040349.D	0.339	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.58	04102023T17:39:00	VL040349.D	0.582	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.39	04102023T17:39:00	VL040349.D	0.402	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.37	04102023T17:39:00	VL040349.D	0.365	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.37	04102023T17:39:00	VL040349.D	0.421	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.38	04102023T17:39:00	VL040349.D	0.397	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.41	04102023T17:39:00	VL040349.D	0.415	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.31	04102023T17:39:00	VL040349.D	0.337	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.44	04102023T17:39:00	VL040349.D	0.51	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.34	04102023T17:39:00	VL040349.D	0.342	04122023T23:23:00	VL040559.D
02232023T15:04:00	VL040332.D	0.028	04102023T19:33:00	VL040346.D	0.025	04122023T19:50:00	VL040576.D
02232023T13:12:00	VL040329.D	0.45	04102023T17:39:00	VL040349.D	0.433	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.45	04102023T17:39:00	VL040349.D	0.453	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.4	04102023T17:39:00	VL040349.D	0.487	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.43	04102023T17:39:00	VL040349.D	0.456	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.47	04102023T17:39:00	VL040349.D	0.455	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.38	04102023T17:39:00	VL040349.D	0.376	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.29	04102023T17:39:00	VL040349.D	0.287	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.29	04102023T17:39:00	VL040349.D	0.35	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.35	04102023T17:39:00	VL040349.D	0.339	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.51	04102023T17:39:00	VL040349.D	0.483	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.46	04102023T17:39:00	VL040349.D	0.467	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.39	04102023T17:39:00	VL040349.D	0.394	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.32	04102023T17:39:00	VL040349.D	0.31	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.33	04102023T17:39:00	VL040349.D	0.336	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.45	04102023T17:39:00	VL040349.D	0.441	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.37	04102023T17:39:00	VL040349.D	0.365	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.47	04102023T17:39:00	VL040349.D	0.442	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.38	04102023T17:39:00	VL040349.D	0.363	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.66	04102023T17:39:00	VL040349.D	0.659	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.31	04102023T17:39:00	VL040349.D	0.305	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.41	04102023T17:39:00	VL040349.D	0.362	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.47	04102023T17:39:00	VL040349.D	0.533	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.26	04102023T17:39:00	VL040349.D	0.261	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.21	04102023T17:39:00	VL040349.D	0.19	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.33	04102023T17:39:00	VL040349.D	0.334	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.35	04102023T17:39:00	VL040349.D	0.345	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.31	04102023T17:39:00	VL040349.D	0.327	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.32	04102023T17:39:00	VL040349.D	0.314	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.42	04102023T17:39:00	VL040349.D	0.465	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.35	04102023T17:39:00	VL040349.D	0.335	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.26	04102023T17:39:00	VL040349.D	0.257	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.27	04102023T17:39:00	VL040349.D	0.256	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.46	04102023T17:39:00	VL040349.D	0.427	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.34	04102023T17:39:00	VL040349.D	0.327	04122023T23:23:00	VL040559.D
02232023T15:04:00	VL040332.D	0.022	04102023T19:33:00	VL040346.D	0.029	04122023T19:50:00	VL040576.D
02232023T13:12:00	VL040329.D	0.3	04102023T17:39:00	VL040349.D	0.347	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.29	04102023T17:39:00	VL040349.D	0.291	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.37	04102023T17:39:00	VL040349.D	0.391	04122023T23:23:00	VL040559.D
02232023T15:04:00	VL040332.D	0.029	04102023T19:33:00	VL040346.D	0.028	04122023T19:50:00	VL040576.D
02232023T13:12:00	VL040329.D	0.45	04102023T17:39:00	VL040349.D	0.47	04122023T23:23:00	VL040559.D
02232023T13:12:00	VL040329.D	0.37	04102023T17:39:00	VL040349.D	0.58	04122023T23:23:00	VL040559.D
02232023T15:04:00	VL040332.D	0.024	04102023T19:33:00	VL040346.D	0.029	04122023T19:50:00	VL040576.D

- A
- B
- C
- D
- E
- F
- G
- H

Concentration13	Adate13	Datafile14	Concentration14	Adate14	Datafile15	Concentration15	Adate15
0.48	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.488	10042023T19:33:00
0.03	07312023T17:18:00	VL040581.D	0.031	07312023T20:41:00	VL040772.D	0.035	10122023T01:48:00
0.031	07312023T17:18:00	VL040581.D	0.026	07312023T20:41:00	VL040772.D	0.029	10122023T01:48:00
0.459	07272023T03:27:00	VL040579.D	0.36	07312023T19:20:00	VL040747.D	0.405	10042023T19:33:00
0.573	07272023T03:27:00	VL040579.D	0.5	07312023T19:20:00	VL040747.D	0.557	10042023T19:33:00
0.535	07272023T03:27:00	VL040579.D	0.43	07312023T19:20:00	VL040747.D	0.527	10042023T19:33:00
0.527	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.481	10042023T19:33:00
0.33	07272023T03:27:00	VL040579.D	0.23	07312023T19:20:00	VL040747.D	0.305	10042023T19:33:00
0.389	07272023T03:27:00	VL040579.D	0.3	07312023T19:20:00	VL040747.D	0.361	10042023T19:33:00
0.425	07272023T03:27:00	VL040579.D	0.32	07312023T19:20:00	VL040747.D	0.359	10042023T19:33:00
0.442	07272023T03:27:00	VL040579.D	0.36	07312023T19:20:00	VL040747.D	0.45	10042023T19:33:00
0.519	07272023T03:27:00	VL040579.D	0.41	07312023T19:20:00	VL040747.D	0.426	10042023T19:33:00
0.452	07272023T03:27:00	VL040579.D	0.36	07312023T19:20:00	VL040747.D	0.42	10042023T19:33:00
0.33	07272023T03:27:00	VL040579.D	0.26	07312023T19:20:00	VL040747.D	0.321	10042023T19:33:00
0.412	07272023T03:27:00	VL040579.D	0.41	07312023T19:20:00	VL040747.D	0.466	10042023T19:33:00
0.388	07272023T03:27:00	VL040579.D	0.33	07312023T19:20:00	VL040747.D	0.412	10042023T19:33:00
0.438	07272023T03:27:00	VL040579.D	0.36	07312023T19:20:00	VL040747.D	0.382	10042023T19:33:00
0.373	07272023T03:27:00	VL040579.D	0.29	07312023T19:20:00	VL040747.D	0.257	10042023T19:33:00
0.398	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.335	10042023T19:33:00
0.448	07272023T03:27:00	VL040579.D	0.37	07312023T19:20:00	VL040747.D	0.443	10042023T19:33:00
0.35	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.325	10042023T19:33:00
0.303	07272023T03:27:00	VL040579.D	0.22	07312023T19:20:00	VL040747.D	0.283	10042023T19:33:00
0.32	07272023T03:27:00	VL040579.D	0.23	07312023T19:20:00	VL040747.D	0.317	10042023T19:33:00
0.348	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.302	10042023T19:33:00
0.674	07272023T03:27:00	VL040579.D	0.48	07312023T19:20:00	VL040747.D	0.65	10042023T19:33:00
0.436	07272023T03:27:00	VL040579.D	0.37	07312023T19:20:00	VL040747.D	0.511	10042023T19:33:00
0.397	07272023T03:27:00	VL040579.D	0.3	07312023T19:20:00	VL040747.D	0.386	10042023T19:33:00
0.425	07272023T03:27:00	VL040579.D	0.34	07312023T19:20:00	VL040747.D	0.307	10042023T19:33:00
0.467	07272023T03:27:00	VL040579.D	0.4	07312023T19:20:00	VL040747.D	0.409	10042023T19:33:00
0.508	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.506	10042023T19:33:00
0.352	07272023T03:27:00	VL040579.D	0.33	07312023T19:20:00	VL040747.D	0.349	10042023T19:33:00
0.543	07272023T03:27:00	VL040579.D	0.46	07312023T19:20:00	VL040747.D	0.496	10042023T19:33:00
0.46	07272023T03:27:00	VL040579.D	0.42	07312023T19:20:00	VL040747.D	0.477	10042023T19:33:00
0.036	07312023T17:18:00	VL040581.D	0.033	07312023T20:41:00	VL040772.D	0.031	10122023T01:48:00
0.481	07272023T03:27:00	VL040579.D	0.38	07312023T19:20:00	VL040747.D	0.489	10042023T19:33:00
0.526	07272023T03:27:00	VL040579.D	0.45	07312023T19:20:00	VL040747.D	0.55	10042023T19:33:00
0.516	07272023T03:27:00	VL040579.D	0.43	07312023T19:20:00	VL040747.D	0.576	10042023T19:33:00
0.521	07272023T03:27:00	VL040579.D	0.41	07312023T19:20:00	VL040747.D	0.501	10042023T19:33:00
0.486	07272023T03:27:00	VL040579.D	0.39	07312023T19:20:00	VL040747.D	0.544	10042023T19:33:00
0.405	07272023T03:27:00	VL040579.D	0.32	07312023T19:20:00	VL040747.D	0.431	10042023T19:33:00
0.295	07272023T03:27:00	VL040579.D	0.25	07312023T19:20:00	VL040747.D	0.298	10042023T19:33:00
0.301	07272023T03:27:00	VL040579.D	0.21	07312023T19:20:00	VL040747.D	0.317	10042023T19:33:00
0.403	07272023T03:27:00	VL040579.D	0.35	07312023T19:20:00	VL040747.D	0.381	10042023T19:33:00
0.54	07272023T03:27:00	VL040579.D	0.47	07312023T19:20:00	VL040747.D	0.562	10042023T19:33:00
0.516	07272023T03:27:00	VL040579.D	0.46	07312023T19:20:00	VL040747.D	0.509	10042023T19:33:00
0.424	07272023T03:27:00	VL040579.D	0.34	07312023T19:20:00	VL040747.D	0.441	10042023T19:33:00
0.32	07272023T03:27:00	VL040579.D	0.21	07312023T19:20:00	VL040747.D	0.299	10042023T19:33:00
0.341	07272023T03:27:00	VL040579.D	0.25	07312023T19:20:00	VL040747.D	0.355	10042023T19:33:00
0.466	07272023T03:27:00	VL040579.D	0.37	07312023T19:20:00	VL040747.D	0.479	10042023T19:33:00
0.38	07272023T03:27:00	VL040579.D	0.31	07312023T19:20:00	VL040747.D	0.465	10042023T19:33:00
0.451	07272023T03:27:00	VL040579.D	0.43	07312023T19:20:00	VL040747.D	0.426	10042023T19:33:00
0.379	07272023T03:27:00	VL040579.D	0.29	07312023T19:20:00	VL040747.D	0.386	10042023T19:33:00
0.735	07272023T03:27:00	VL040579.D	0.54	07312023T19:20:00	VL040747.D	0.682	10042023T19:33:00
0.283	07272023T03:27:00	VL040579.D	0.23	07312023T19:20:00	VL040747.D	0.324	10042023T19:33:00
0.468	07272023T03:27:00	VL040579.D	0.38	07312023T19:20:00	VL040747.D	0.472	10042023T19:33:00
0.619	07272023T03:27:00	VL040579.D	0.62	07312023T19:20:00	VL040747.D	0.625	10042023T19:33:00
0.274	07272023T03:27:00	VL040579.D	0.19	07312023T19:20:00	VL040747.D	0.258	10042023T19:33:00
0.107	07272023T03:27:00	VL040579.D	0.06	07312023T19:20:00	VL040747.D	0.133	10042023T19:33:00
0.361	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.343	10042023T19:33:00
0.355	07272023T03:27:00	VL040579.D	0.27	07312023T19:20:00	VL040747.D	0.369	10042023T19:33:00
0.355	07272023T03:27:00	VL040579.D	0.26	07312023T19:20:00	VL040747.D	0.301	10042023T19:33:00
0.357	07272023T03:27:00	VL040579.D	0.26	07312023T19:20:00	VL040747.D	0.313	10042023T19:33:00
0.436	07272023T03:27:00	VL040579.D	0.34	07312023T19:20:00	VL040747.D	0.471	10042023T19:33:00
0.393	07272023T03:27:00	VL040579.D	0.29	07312023T19:20:00	VL040747.D	0.354	10042023T19:33:00
0.242	07272023T03:27:00	VL040579.D	0.17	07312023T19:20:00	VL040747.D	0.266	10042023T19:33:00
0.283	07272023T03:27:00	VL040579.D	0.23	07312023T19:20:00	VL040747.D	0.312	10042023T19:33:00
0.539	07272023T03:27:00	VL040579.D	0.43	07312023T19:20:00	VL040747.D	0.394	10042023T19:33:00
0.366	07272023T03:27:00	VL040579.D	0.28	07312023T19:20:00	VL040747.D	0.381	10042023T19:33:00
0.025	07312023T17:18:00	VL040581.D	0.028	07312023T20:41:00	VL040772.D	0.044	10122023T01:48:00
0.353	07272023T03:27:00	VL040579.D	0.26	07312023T19:20:00	VL040747.D	0.351	10042023T19:33:00
0.285	07272023T03:27:00	VL040579.D	0.2	07312023T19:20:00	VL040747.D	0.288	10042023T19:33:00
0.477	07272023T03:27:00	VL040579.D	0.4	07312023T19:20:00	VL040747.D	0.507	10042023T19:33:00
0.027	07312023T17:18:00	VL040581.D	0.027	07312023T20:41:00	VL040772.D	0.039	10122023T01:48:00
0.589	07272023T03:27:00	VL040579.D	0.52	07312023T19:20:00	VL040747.D	0.561	10042023T19:33:00
0.487	07272023T03:27:00	VL040579.D	0.46	07312023T19:20:00	VL040747.D	0.456	10042023T19:33:00
0.03	07312023T17:18:00	VL040581.D	0.033	07312023T20:41:00	VL040772.D	0.041	10122023T01:48:00

Datafile16	Concentration16	Adate16	Department	AVG Concentration	Recovery
VL040805.D	0.428	10252023T18:12:00	VOC	0.45	110
VL040787.D	0.04	10232023T20:32:00	VOC	0.032	110
VL040787.D	0.04	10232023T20:32:00	VOC	0.03	100
VL040805.D	0.471	10252023T18:12:00	VOC	0.42	110
VL040805.D	0.476	10252023T18:12:00	VOC	0.48	120
VL040805.D	0.427	10252023T18:12:00	VOC	0.46	110
VL040805.D	0.47	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.288	10252023T18:12:00	VOC	0.31	79
VL040805.D	0.435	10252023T18:12:00	VOC	0.37	93
VL040805.D	0.433	10252023T18:12:00	VOC	0.38	95
VL040805.D	0.405	10252023T18:12:00	VOC	0.41	100
VL040805.D	0.501	10252023T18:12:00	VOC	0.44	110
VL040805.D	0.456	10252023T18:12:00	VOC	0.42	100
VL040805.D	0.403	10252023T18:12:00	VOC	0.35	87
VL040805.D	0.455	10252023T18:12:00	VOC	0.44	110
VL040805.D	0.412	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.414	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.486	10252023T18:12:00	VOC	0.38	96
VL040805.D	0.434	10252023T18:12:00	VOC	0.37	93
VL040805.D	0.467	10252023T18:12:00	VOC	0.42	100
VL040805.D	0.42	10252023T18:12:00	VOC	0.35	88
VL040805.D	0.353	10252023T18:12:00	VOC	0.3	75
VL040805.D	0.397	10252023T18:12:00	VOC	0.34	84
VL040805.D	0.403	10252023T18:12:00	VOC	0.34	84
VL040805.D	0.576	10252023T18:12:00	VOC	0.53	130
VL040805.D	0.422	10252023T18:12:00	VOC	0.43	110
VL040805.D	0.445	10252023T18:12:00	VOC	0.38	96
VL040805.D	0.205	10252023T18:12:00	VOC	0.36	91
VL040805.D	0.458	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.436	10252023T18:12:00	VOC	0.44	110
VL040805.D	0.34	10252023T18:12:00	VOC	0.33	82
VL040805.D	0.508	10252023T18:12:00	VOC	0.47	120
VL040805.D	0.361	10252023T18:12:00	VOC	0.38	95
VL040787.D	0.04	10232023T20:32:00	VOC	0.031	100
VL040805.D	0.454	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.517	10252023T18:12:00	VOC	0.48	120
VL040805.D	0.435	10252023T18:12:00	VOC	0.47	120
VL040805.D	0.487	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.458	10252023T18:12:00	VOC	0.46	120
VL040805.D	0.435	10252023T18:12:00	VOC	0.4	99
VL040805.D	0.335	10252023T18:12:00	VOC	0.3	75
VL040805.D	0.416	10252023T18:12:00	VOC	0.39	96
VL040805.D	0.384	10252023T18:12:00	VOC	0.36	89
VL040805.D	0.516	10252023T18:12:00	VOC	0.54	140
VL040805.D	0.505	10252023T18:12:00	VOC	0.48	120
VL040805.D	0.404	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.406	10252023T18:12:00	VOC	0.33	82
VL040805.D	0.404	10252023T18:12:00	VOC	0.35	88
VL040805.D	0.446	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.452	10252023T18:12:00	VOC	0.4	100
VL040805.D	0.442	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.435	10252023T18:12:00	VOC	0.38	96
VL040805.D	0.888	10252023T18:12:00	VOC	0.72	180
VL040805.D	0.348	10252023T18:12:00	VOC	0.32	80
VL040805.D	0.353	10252023T18:12:00	VOC	0.41	100
VL040805.D	0.619	10252023T18:12:00	VOC	0.56	140
VL040805.D	0.226	10252023T18:12:00	VOC	0.27	66
VL040805.D	0	10252023T18:12:00	VOC	0.14	35
VL040805.D	0.415	10252023T18:12:00	VOC	0.35	88
VL040805.D	0.389	10252023T18:12:00	VOC	0.35	89
VL040805.D	0.439	10252023T18:12:00	VOC	0.35	87
VL040805.D	0.381	10252023T18:12:00	VOC	0.34	85
VL040805.D	0.467	10252023T18:12:00	VOC	0.46	110
VL040805.D	0.449	10252023T18:12:00	VOC	0.37	93
VL040805.D	0.35	10252023T18:12:00	VOC	0.27	67
VL040805.D	0.291	10252023T18:12:00	VOC	0.28	70
VL040805.D	0.37	10252023T18:12:00	VOC	0.45	110
VL040805.D	0.427	10252023T18:12:00	VOC	0.37	92
VL040787.D	0.04	10232023T20:32:00	VOC	0.032	110
VL040805.D	0.416	10252023T18:12:00	VOC	0.35	88
VL040805.D	0.38	10252023T18:12:00	VOC	0.31	78
VL040805.D	0.435	10252023T18:12:00	VOC	0.43	110
VL040787.D	0.05	10232023T20:32:00	VOC	0.033	110
VL040805.D	0.506	10252023T18:12:00	VOC	0.48	120
VL040805.D	0.341	10252023T18:12:00	VOC	0.43	110
VL040787.D	0.04	10232023T20:32:00	VOC	0.033	110
			VOC	0.544	136

A
B
C
D
E
F
G
H



SAMPLE DATA

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	565-SG-1	SDG No.:	Q1586
Lab Sample ID:	Q1586-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042154.D	4		03/17/25 12:51	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.84	4.15	U	4.15	9.89	ug/m3
74-87-3	Chloromethane	0.39	0.81	U	0.81	4.13	ug/m3
75-01-4	Vinyl Chloride	0.060	0.15	U	0.15	0.31	ug/m3
74-83-9	Bromomethane	0.56	2.17	U	2.17	7.77	ug/m3
75-00-3	Chloroethane	0.68	1.79	U	1.79	5.28	ug/m3
109-99-9	Tetrahydrofuran	0.52	1.53	U	1.53	5.90	ug/m3
75-69-4	Trichlorofluoromethane	0.64	3.60	U	3.60	11.2	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.64	4.91	U	4.91	15.3	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.35	2.45	U	2.45	14.0	ug/m3
75-65-0	tert-Butyl alcohol	0.76	2.30	U	2.30	6.06	ug/m3
142-82-5	Heptane	570	2340	E	1.80	8.20	ug/m3
75-35-4	1,1-Dichloroethene	0.56	2.22	U	2.22	7.93	ug/m3
67-64-1	Acetone	120	285	E	2.28	4.75	ug/m3
75-15-0	Carbon Disulfide	0.64	1.99	U	1.99	6.23	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.48	1.73	U	1.73	7.21	ug/m3
75-09-2	Methylene Chloride	0.76	2.64	U	2.64	6.95	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.76	3.01	U	3.01	7.93	ug/m3
75-34-3	1,1-Dichloroethane	0.52	2.10	U	2.10	8.09	ug/m3
110-82-7	Cyclohexane	720	2480	E	3.03	6.88	ug/m3
78-93-3	2-Butanone	33.5	98.8		1.42	5.90	ug/m3
56-23-5	Carbon Tetrachloride	0.040	0.25	U	0.25	0.75	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.36	1.43	U	1.43	7.93	ug/m3
67-66-3	Chloroform	0.33	1.61	U	1.61	9.77	ug/m3
71-55-6	1,1,1-Trichloroethane	0.040	0.22	U	0.22	0.65	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.40	1.87	U	1.87	9.34	ug/m3
71-43-2	Benzene	290	926	E	1.12	6.39	ug/m3
107-06-2	1,2-Dichloroethane	0.36	1.46	U	1.46	8.09	ug/m3
79-01-6	Trichloroethene	0.39	2.10		0.38	0.64	ug/m3
78-87-5	1,2-Dichloropropane	0.44	2.03	U	2.03	9.24	ug/m3
75-27-4	Bromodichloromethane	0.31	2.08	U	2.08	13.4	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.37	1.52	U	1.52	8.20	ug/m3
108-88-3	Toluene	19.8	74.6		1.66	7.54	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.24	1.09	U	1.09	9.08	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.26	1.18	U	1.18	9.08	ug/m3
79-00-5	1,1,2-Trichloroethane	0.28	1.53	U	1.53	10.9	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	565-SG-1	SDG No.:	Q1586
Lab Sample ID:	Q1586-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042154.D	4		03/17/25 12:51	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.26	2.22	U	2.22	17.0	ug/m3
106-93-4	1,2-Dibromoethane	0.29	2.23	U	2.23	3.07	ug/m3
127-18-4	Tetrachloroethene	17.9	121		0.41	0.81	ug/m3
108-90-7	Chlorobenzene	0.30	1.38	U	1.38	9.21	ug/m3
100-41-4	Ethyl Benzene	10.7	46.5		2.08	8.69	ug/m3
179601-23-1	m/p-Xylene	39.0	169		3.65	17.4	ug/m3
95-47-6	o-Xylene	11.6	50.4		2.08	8.69	ug/m3
100-42-5	Styrene	0.44	1.87	U	1.87	8.52	ug/m3
75-25-2	Bromoform	0.23	2.38	U	2.38	20.7	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.040	0.27	U	0.27	0.82	ug/m3
95-49-8	2-Chlorotoluene	0.44	2.28	U	2.28	10.4	ug/m3
108-67-8	1,3,5-Trimethylbenzene	2.10	10.3		2.16	9.83	ug/m3
95-63-6	1,2,4-Trimethylbenzene	3.60	17.7		1.52	9.83	ug/m3
541-73-1	1,3-Dichlorobenzene	0.32	1.92	U	1.92	12.0	ug/m3
106-46-7	1,4-Dichlorobenzene	0.21	1.26	U	1.26	12.0	ug/m3
95-50-1	1,2-Dichlorobenzene	0.30	1.80	U	1.80	12.0	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.34	2.52	U	2.52	14.8	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.34	3.63	U	3.63	21.3	ug/m3
106-99-0	1,3-Butadiene	0.52	1.15	U	1.15	4.42	ug/m3
91-20-3	Naphthalene	0.30	1.57	U	1.57	2.10	ug/m3
622-96-8	4-Ethyltoluene	1.20	5.90	J	2.36	9.83	ug/m3
110-54-3	Hexane	970	3420	E	1.55	7.05	ug/m3
107-05-1	Allyl Chloride	0.60	1.88	U	1.88	6.26	ug/m3
123-91-1	1,4-Dioxane	0.84	3.03	U	3.03	7.21	ug/m3
80-62-6	Methyl Methacrylate	0.34	1.39	U	1.39	8.19	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	9.60			70 (65) - 130 (135)	96%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	192000		2.797			
540-36-3	1,4-Difluorobenzene	526000		3.978			
3114-55-4	Chlorobenzene-d5	471000		8.901			

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	565-SG-1	SDG No.:	Q1586
Lab Sample ID:	Q1586-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042154.D	4		03/17/25 12:51	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	565-SG-1DL	SDG No.:	Q1586
Lab Sample ID:	Q1586-01DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042155.D	40		03/17/25 13:37	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	8.40	41.5	UD	41.5	98.9	ug/m3
74-87-3	Chloromethane	3.90	8.05	UD	8.05	41.3	ug/m3
75-01-4	Vinyl Chloride	0.56	1.43	UD	1.43	3.07	ug/m3
74-83-9	Bromomethane	5.60	21.7	UD	21.7	77.7	ug/m3
75-00-3	Chloroethane	6.80	17.9	UD	17.9	52.8	ug/m3
109-99-9	Tetrahydrofuran	5.20	15.3	UD	15.3	59.0	ug/m3
75-69-4	Trichlorofluoromethane	6.40	36.0	UD	36.0	112	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	6.40	49.0	UD	49.0	153	ug/m3
76-14-2	Dichlorotetrafluoroethane	3.50	24.5	UD	24.5	140	ug/m3
75-65-0	tert-Butyl alcohol	7.60	23.0	UD	23.0	60.6	ug/m3
142-82-5	Heptane	830	3400	ED	18.0	82.0	ug/m3
75-35-4	1,1-Dichloroethene	5.60	22.2	UD	22.2	79.3	ug/m3
67-64-1	Acetone	9.60	22.8	UD	22.8	47.5	ug/m3
75-15-0	Carbon Disulfide	6.40	19.9	UD	19.9	62.3	ug/m3
1634-04-4	Methyl tert-Butyl Ether	4.80	17.3	UD	17.3	72.1	ug/m3
75-09-2	Methylene Chloride	7.60	26.4	UD	26.4	69.5	ug/m3
156-60-5	trans-1,2-Dichloroethene	7.60	30.1	UD	30.1	79.3	ug/m3
75-34-3	1,1-Dichloroethane	5.20	21.1	UD	21.1	81.0	ug/m3
110-82-7	Cyclohexane	1100	3790	ED	30.3	68.8	ug/m3
78-93-3	2-Butanone	39.7	117	D	14.2	59.0	ug/m3
56-23-5	Carbon Tetrachloride	0.44	2.77	UD	2.77	7.55	ug/m3
156-59-2	cis-1,2-Dichloroethene	3.60	14.3	UD	14.3	79.3	ug/m3
67-66-3	Chloroform	3.30	16.1	UD	16.1	97.7	ug/m3
71-55-6	1,1,1-Trichloroethane	0.38	2.07	UD	2.07	6.55	ug/m3
540-84-1	2,2,4-Trimethylpentane	4.00	18.7	UD	18.7	93.4	ug/m3
71-43-2	Benzene	440	1410	D	11.2	63.9	ug/m3
107-06-2	1,2-Dichloroethane	3.60	14.6	UD	14.6	81.0	ug/m3
79-01-6	Trichloroethene	0.68	3.65	UD	3.65	6.45	ug/m3
78-87-5	1,2-Dichloropropane	4.40	20.3	UD	20.3	92.4	ug/m3
75-27-4	Bromodichloromethane	3.10	20.8	UD	20.8	134	ug/m3
108-10-1	4-Methyl-2-Pentanone	3.70	15.2	UD	15.2	82.0	ug/m3
108-88-3	Toluene	19.8	74.6	JD	16.6	75.4	ug/m3
10061-02-6	t-1,3-Dichloropropene	2.40	10.9	UD	10.9	90.8	ug/m3
10061-01-5	cis-1,3-Dichloropropene	2.60	11.8	UD	11.8	90.8	ug/m3
79-00-5	1,1,2-Trichloroethane	2.80	15.3	UD	15.3	109	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	565-SG-1DL	SDG No.:	Q1586
Lab Sample ID:	Q1586-01DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042155.D	40		03/17/25 13:37	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	2.60	22.1	UD	22.1	170	ug/m3
106-93-4	1,2-Dibromoethane	2.90	22.3	UD	22.3	30.7	ug/m3
127-18-4	Tetrachloroethene	21.0	142	D	4.07	8.14	ug/m3
108-90-7	Chlorobenzene	3.00	13.8	UD	13.8	92.1	ug/m3
100-41-4	Ethyl Benzene	8.80	38.2	JD	20.9	86.9	ug/m3
179601-23-1	m/p-Xylene	34.0	148	JD	36.5	174	ug/m3
95-47-6	o-Xylene	10.3	44.7	JD	20.9	86.9	ug/m3
100-42-5	Styrene	4.40	18.7	UD	18.7	85.2	ug/m3
75-25-2	Bromoform	2.30	23.8	UD	23.8	207	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.44	3.02	UD	3.02	8.24	ug/m3
95-49-8	2-Chlorotoluene	4.40	22.8	UD	22.8	104	ug/m3
108-67-8	1,3,5-Trimethylbenzene	4.40	21.6	UD	21.6	98.3	ug/m3
95-63-6	1,2,4-Trimethylbenzene	3.10	15.2	UD	15.2	98.3	ug/m3
541-73-1	1,3-Dichlorobenzene	3.20	19.2	UD	19.2	120	ug/m3
106-46-7	1,4-Dichlorobenzene	2.10	12.6	UD	12.6	120	ug/m3
95-50-1	1,2-Dichlorobenzene	3.00	18.0	UD	18.0	120	ug/m3
120-82-1	1,2,4-Trichlorobenzene	3.40	25.2	UD	25.2	148	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	3.40	36.3	UD	36.3	213	ug/m3
106-99-0	1,3-Butadiene	5.20	11.5	UD	11.5	44.3	ug/m3
91-20-3	Naphthalene	3.00	15.7	UD	15.7	21.0	ug/m3
622-96-8	4-Ethyltoluene	4.80	23.6	UD	23.6	98.3	ug/m3
110-54-3	Hexane	4600	16200	ED	15.5	70.5	ug/m3
107-05-1	Allyl Chloride	6.00	18.8	UD	18.8	62.6	ug/m3
123-91-1	1,4-Dioxane	8.40	30.3	UD	30.3	72.1	ug/m3
80-62-6	Methyl Methacrylate	3.40	13.9	UD	13.9	81.9	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	9.50			70 (65) - 130 (135)	95%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	177000		2.8			
540-36-3	1,4-Difluorobenzene	461000		3.978			
3114-55-4	Chlorobenzene-d5	407000		8.901			

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	565-SG-1DL	SDG No.:	Q1586
Lab Sample ID:	Q1586-01DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042155.D	40		03/17/25 13:37	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	565-SG-1DL2	SDG No.:	Q1586
Lab Sample ID:	Q1586-01DL2	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042161.D	600		03/17/25 17:01	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	130	643	UD	643	1480	ug/m3
74-87-3	Chloromethane	58.8	121	UD	121	620	ug/m3
75-01-4	Vinyl Chloride	8.40	21.5	UD	21.5	46.0	ug/m3
74-83-9	Bromomethane	84.0	326	UD	326	1160	ug/m3
75-00-3	Chloroethane	100	264	UD	264	792	ug/m3
109-99-9	Tetrahydrofuran	78.0	230	UD	230	885	ug/m3
75-69-4	Trichlorofluoromethane	96.0	539	UD	539	1690	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	96.0	736	UD	736	2300	ug/m3
76-14-2	Dichlorotetrafluoroethane	52.8	369	UD	369	2100	ug/m3
75-65-0	tert-Butyl alcohol	110	333	UD	333	909	ug/m3
142-82-5	Heptane	590	2420	D	270	1230	ug/m3
75-35-4	1,1-Dichloroethene	84.0	333	UD	333	1190	ug/m3
67-64-1	Acetone	140	333	UD	333	713	ug/m3
75-15-0	Carbon Disulfide	96.0	299	UD	299	934	ug/m3
1634-04-4	Methyl tert-Butyl Ether	72.0	260	UD	260	1080	ug/m3
75-09-2	Methylene Chloride	110	382	UD	382	1040	ug/m3
156-60-5	trans-1,2-Dichloroethene	110	436	UD	436	1190	ug/m3
75-34-3	1,1-Dichloroethane	78.0	316	UD	316	1210	ug/m3
110-82-7	Cyclohexane	780	2680	D	447	1030	ug/m3
78-93-3	2-Butanone	72.0	212	UD	212	885	ug/m3
56-23-5	Carbon Tetrachloride	6.60	41.5	UD	41.5	113	ug/m3
156-59-2	cis-1,2-Dichloroethene	53.4	212	UD	212	1190	ug/m3
67-66-3	Chloroform	49.8	243	UD	243	1470	ug/m3
71-55-6	1,1,1-Trichloroethane	5.70	31.1	UD	31.1	98.2	ug/m3
540-84-1	2,2,4-Trimethylpentane	60.0	280	UD	280	1400	ug/m3
71-43-2	Benzene	420	1340	D	167	958	ug/m3
107-06-2	1,2-Dichloroethane	54.6	221	UD	221	1210	ug/m3
79-01-6	Trichloroethene	10.2	54.8	UD	54.8	96.7	ug/m3
78-87-5	1,2-Dichloropropane	66.0	305	UD	305	1390	ug/m3
75-27-4	Bromodichloromethane	46.2	310	UD	310	2010	ug/m3
108-10-1	4-Methyl-2-Pentanone	55.8	229	UD	229	1230	ug/m3
108-88-3	Toluene	66.0	249	UD	249	1130	ug/m3
10061-02-6	t-1,3-Dichloropropene	36.0	163	UD	163	1360	ug/m3
10061-01-5	cis-1,3-Dichloropropene	38.4	174	UD	174	1360	ug/m3
79-00-5	1,1,2-Trichloroethane	42.6	232	UD	232	1640	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	565-SG-1DL2	SDG No.:	Q1586
Lab Sample ID:	Q1586-01DL2	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042161.D	600		03/17/25 17:01	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	39.6	337	UD	337	2560	ug/m3
106-93-4	1,2-Dibromoethane	43.2	332	UD	332	461	ug/m3
127-18-4	Tetrachloroethene	9.00	61.0	UD	61.0	122	ug/m3
108-90-7	Chlorobenzene	45.6	210	UD	210	1380	ug/m3
100-41-4	Ethyl Benzene	72.0	313	UD	313	1300	ug/m3
179601-23-1	m/p-Xylene	130	565	UD	565	2610	ug/m3
95-47-6	o-Xylene	72.0	313	UD	313	1300	ug/m3
100-42-5	Styrene	66.0	281	UD	281	1280	ug/m3
75-25-2	Bromoform	34.2	354	UD	354	3100	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	6.60	45.3	UD	45.3	124	ug/m3
95-49-8	2-Chlorotoluene	66.0	342	UD	342	1550	ug/m3
108-67-8	1,3,5-Trimethylbenzene	66.0	324	UD	324	1470	ug/m3
95-63-6	1,2,4-Trimethylbenzene	46.8	230	UD	230	1470	ug/m3
541-73-1	1,3-Dichlorobenzene	48.0	289	UD	289	1800	ug/m3
106-46-7	1,4-Dichlorobenzene	31.8	191	UD	191	1800	ug/m3
95-50-1	1,2-Dichlorobenzene	45.0	271	UD	271	1800	ug/m3
120-82-1	1,2,4-Trichlorobenzene	51.0	379	UD	379	2230	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	51.0	544	UD	544	3200	ug/m3
106-99-0	1,3-Butadiene	78.0	173	UD	173	664	ug/m3
91-20-3	Naphthalene	45.0	236	UD	236	315	ug/m3
622-96-8	4-Ethyltoluene	72.0	354	UD	354	1470	ug/m3
110-54-3	Hexane	6200	21900	D	233	1060	ug/m3
107-05-1	Allyl Chloride	90.0	282	UD	282	939	ug/m3
123-91-1	1,4-Dioxane	130	469	UD	469	1080	ug/m3
80-62-6	Methyl Methacrylate	51.6	211	UD	211	1230	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	9.90			70 (65) - 130 (135)	99%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	159000		2.797			
540-36-3	1,4-Difluorobenzene	390000		3.972			
3114-55-4	Chlorobenzene-d5	317000		8.895			

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	565-SG-1DL2	SDG No.:	Q1586
Lab Sample ID:	Q1586-01DL2	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042161.D	600		03/17/25 17:01	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	557-SG-1	SDG No.:	Q1586
Lab Sample ID:	Q1586-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042159.D	1		03/17/25 15:53	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.54	2.67		1.04	2.47	ug/m3
74-87-3	Chloromethane	0.10	0.21	U	0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.14	0.54	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.17	0.45	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.13	0.38	U	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.31	1.74	J	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.16	1.23	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.090	0.63	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.24	0.73	J	0.58	1.52	ug/m3
142-82-5	Heptane	0.11	0.45	U	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
67-64-1	Acetone	8.50	20.2		0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	1.50	4.67		0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.12	0.43	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	0.50	1.74		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.19	0.75	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	0.68	2.01		0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.080	0.50		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.090	0.36	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.78	3.81		0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.10	0.47	U	0.47	2.34	ug/m3
71-43-2	Benzene	0.11	0.35	J	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.11	0.51	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.080	0.54	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.090	0.37	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.11	0.41	U	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.070	0.38	U	0.38	2.73	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	557-SG-1	SDG No.:	Q1586
Lab Sample ID:	Q1586-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042159.D	1		03/17/25 15:53	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.070	0.60	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.070	0.54	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	3.10	21.0		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.12	0.52	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.21	0.91	U	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.12	0.52	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.11	0.47	U	0.47	2.13	ug/m3
75-25-2	Bromoform	0.060	0.62	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.010	0.070	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.11	0.57	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.11	0.54	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.080	0.39	U	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.050	0.30	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.090	0.67	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.090	0.96	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.13	0.29	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.080	0.42	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.12	0.59	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.29	1.02	J	0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.15	0.47	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.21	0.76	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.090	0.37	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.2			70 (65) - 130 (135)	102%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	161000		2.797			
540-36-3	1,4-Difluorobenzene	390000		3.971			
3114-55-4	Chlorobenzene-d5	319000		8.898			

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	557-SG-1	SDG No.:	Q1586
Lab Sample ID:	Q1586-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042159.D	1		03/17/25 15:53	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	557-SG-2	SDG No.:	Q1586
Lab Sample ID:	Q1586-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042158.D	1		03/17/25 15:19	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.56	2.77		1.04	2.47	ug/m3
74-87-3	Chloromethane	0.29	0.60	J	0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.14	0.54	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.17	0.45	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.13	0.38	U	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.30	1.69	J	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.16	1.23	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.090	0.63	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.57	1.73		0.58	1.52	ug/m3
142-82-5	Heptane	0.28	1.15	J	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
67-64-1	Acetone	15.2	36.1	E	0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	2.40	7.47		0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.12	0.43	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	0.60	2.08		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.19	0.75	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	1.30	3.83		0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.090	0.57		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.14	0.56	J	0.36	1.98	ug/m3
67-66-3	Chloroform	8.50	41.5		0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.10	0.47	U	0.47	2.34	ug/m3
71-43-2	Benzene	3.50	11.2		0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.47	2.53		0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.11	0.51	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.32	2.14	J	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.090	0.37	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.20	0.75	J	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.070	0.38	U	0.38	2.73	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	557-SG-2	SDG No.:	Q1586
Lab Sample ID:	Q1586-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042158.D	1		03/17/25 15:19	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.070	0.60	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.070	0.54	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	13.1	88.8		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.12	0.52	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.21	0.91	U	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.12	0.52	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.11	0.47	U	0.47	2.13	ug/m3
75-25-2	Bromoform	0.060	0.62	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.010	0.070	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.11	0.57	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.11	0.54	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.34	1.67	J	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.050	0.30	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.090	0.67	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.090	0.96	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.13	0.29	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.77	4.04		0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.12	0.59	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.70	2.47		0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.15	0.47	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.21	0.76	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.090	0.37	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.3			70 (65) - 130 (135)	103%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	163000		2.8			
540-36-3	1,4-Difluorobenzene	382000		3.975			
3114-55-4	Chlorobenzene-d5	335000		8.898			

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	557-SG-2	SDG No.:	Q1586
Lab Sample ID:	Q1586-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042158.D	1		03/17/25 15:19	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	557-SG-2DL	SDG No.:	Q1586
Lab Sample ID:	Q1586-03DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042157.D	10		03/17/25 14:41	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	2.10	10.4	UD	10.4	24.7	ug/m3
74-87-3	Chloromethane	0.98	2.02	UD	2.02	10.3	ug/m3
75-01-4	Vinyl Chloride	0.14	0.36	UD	0.36	0.77	ug/m3
74-83-9	Bromomethane	1.40	5.44	UD	5.44	19.4	ug/m3
75-00-3	Chloroethane	1.70	4.49	UD	4.49	13.2	ug/m3
109-99-9	Tetrahydrofuran	1.30	3.83	UD	3.83	14.8	ug/m3
75-69-4	Trichlorofluoromethane	1.60	8.99	UD	8.99	28.1	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	1.60	12.3	UD	12.3	38.3	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.88	6.15	UD	6.15	35.0	ug/m3
75-65-0	tert-Butyl alcohol	1.90	5.76	UD	5.76	15.2	ug/m3
142-82-5	Heptane	1.10	4.51	UD	4.51	20.5	ug/m3
75-35-4	1,1-Dichloroethene	1.40	5.55	UD	5.55	19.8	ug/m3
67-64-1	Acetone	16.0	38.0	D	5.70	11.9	ug/m3
75-15-0	Carbon Disulfide	1.60	4.98	UD	4.98	15.6	ug/m3
1634-04-4	Methyl tert-Butyl Ether	1.20	4.33	UD	4.33	18.0	ug/m3
75-09-2	Methylene Chloride	1.90	6.60	UD	6.60	17.4	ug/m3
156-60-5	trans-1,2-Dichloroethene	1.90	7.53	UD	7.53	19.8	ug/m3
75-34-3	1,1-Dichloroethane	1.30	5.26	UD	5.26	20.2	ug/m3
110-82-7	Cyclohexane	2.20	7.57	UD	7.57	17.2	ug/m3
78-93-3	2-Butanone	1.60	4.72	JD	3.54	14.8	ug/m3
56-23-5	Carbon Tetrachloride	0.11	0.69	UD	0.69	1.89	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.89	3.53	UD	3.53	19.8	ug/m3
67-66-3	Chloroform	8.50	41.5	D	4.05	24.4	ug/m3
71-55-6	1,1,1-Trichloroethane	0.10	0.55	UD	0.55	1.64	ug/m3
540-84-1	2,2,4-Trimethylpentane	1.00	4.67	UD	4.67	23.4	ug/m3
71-43-2	Benzene	3.10	9.90	JD	2.78	16.0	ug/m3
107-06-2	1,2-Dichloroethane	0.91	3.68	UD	3.68	20.2	ug/m3
79-01-6	Trichloroethene	0.17	0.91	UD	0.91	1.61	ug/m3
78-87-5	1,2-Dichloropropane	1.10	5.08	UD	5.08	23.1	ug/m3
75-27-4	Bromodichloromethane	0.77	5.16	UD	5.16	33.5	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.93	3.81	UD	3.81	20.5	ug/m3
108-88-3	Toluene	1.10	4.15	UD	4.15	18.8	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.60	2.72	UD	2.72	22.7	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.64	2.91	UD	2.91	22.7	ug/m3
79-00-5	1,1,2-Trichloroethane	0.71	3.87	UD	3.87	27.3	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	557-SG-2DL	SDG No.:	Q1586
Lab Sample ID:	Q1586-03DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042157.D	10		03/17/25 14:41	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.66	5.62	UD	5.62	42.6	ug/m3
106-93-4	1,2-Dibromoethane	0.72	5.53	UD	5.53	7.69	ug/m3
127-18-4	Tetrachloroethene	11.7	79.3	D	1.02	2.03	ug/m3
108-90-7	Chlorobenzene	0.76	3.50	UD	3.50	23.0	ug/m3
100-41-4	Ethyl Benzene	1.20	5.21	UD	5.21	21.7	ug/m3
179601-23-1	m/p-Xylene	2.10	9.12	UD	9.12	43.4	ug/m3
95-47-6	o-Xylene	1.20	5.21	UD	5.21	21.7	ug/m3
100-42-5	Styrene	1.10	4.68	UD	4.68	21.3	ug/m3
75-25-2	Bromoform	0.57	5.89	UD	5.89	51.7	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.11	0.76	UD	0.76	2.06	ug/m3
95-49-8	2-Chlorotoluene	1.10	5.70	UD	5.70	25.9	ug/m3
108-67-8	1,3,5-Trimethylbenzene	1.10	5.41	UD	5.41	24.6	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.78	3.83	UD	3.83	24.6	ug/m3
541-73-1	1,3-Dichlorobenzene	0.80	4.81	UD	4.81	30.1	ug/m3
106-46-7	1,4-Dichlorobenzene	0.53	3.19	UD	3.19	30.1	ug/m3
95-50-1	1,2-Dichlorobenzene	0.75	4.51	UD	4.51	30.1	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.85	6.31	UD	6.31	37.1	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.85	9.07	UD	9.07	53.3	ug/m3
106-99-0	1,3-Butadiene	1.30	2.88	UD	2.88	11.1	ug/m3
91-20-3	Naphthalene	0.75	3.93	UD	3.93	5.24	ug/m3
622-96-8	4-Ethyltoluene	1.20	5.90	UD	5.90	24.6	ug/m3
110-54-3	Hexane	1.10	3.88	UD	3.88	17.6	ug/m3
107-05-1	Allyl Chloride	1.50	4.70	UD	4.70	15.7	ug/m3
123-91-1	1,4-Dioxane	2.10	7.57	UD	7.57	18.0	ug/m3
80-62-6	Methyl Methacrylate	0.86	3.52	UD	3.52	20.5	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	9.80			70 (65) - 130 (135)	98%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	166000		2.797			
540-36-3	1,4-Difluorobenzene	386000		3.975			
3114-55-4	Chlorobenzene-d5	319000		8.895			

Report of Analysis

Client:	RTP Environmental	Date Collected:	03/14/25
Project:	Yellow Cab - TO-15	Date Received:	03/14/25
Client Sample ID:	557-SG-2DL	SDG No.:	Q1586
Lab Sample ID:	Q1586-03DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042157.D	10		03/17/25 14:41	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements



QC SUMMARY

Surrogate Summary

SDG No.: **Q1586**

Client: **RTP Environmental**

Analytical Method: **SWTO-15**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1586-01	565-SG-1	1-Bromo-4-Fluorobenzene	10	9.65	96	70 (65)	130 (135)
Q1586-01DL	565-SG-1DL	1-Bromo-4-Fluorobenzene	10	9.51	95	70 (65)	130 (135)
Q1586-01DL2	565-SG-1DL2	1-Bromo-4-Fluorobenzene	10	9.93	99	70 (65)	130 (135)
Q1586-02	557-SG-1	1-Bromo-4-Fluorobenzene	10	10.2	102	70 (65)	130 (135)
Q1586-02DUP	557-SG-1DUP	1-Bromo-4-Fluorobenzene	10	10.1	101	70 (65)	130 (135)
Q1586-03	557-SG-2	1-Bromo-4-Fluorobenzene	10	10.3	103	70 (65)	130 (135)
Q1586-03DL	557-SG-2DL	1-Bromo-4-Fluorobenzene	10	9.84	98	70 (65)	130 (135)
VL0317ABL01	VL0317ABL01	1-Bromo-4-Fluorobenzene	10	9.84	98	70 (65)	130 (135)
VL0317ABS01	VL0317ABS01	1-Bromo-4-Fluorobenzene	10	10.8	108	70 (65)	130 (135)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1586

Client: RTP Environmental

Analytical Method: SWTO-15

Datafile : VL042152.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0317ABS01	Dichlorodifluoromethane	10	10.3	ppbv	103			40 (70)	160 (130)	
	Chloromethane	10	11.0	ppbv	110			40 (70)	160 (130)	
	Vinyl Chloride	10	11.1	ppbv	111			70 (70)	130 (130)	
	Bromomethane	10	11.1	ppbv	111			40 (70)	160 (130)	
	Chloroethane	10	10.7	ppbv	107			40 (70)	160 (130)	
	Tetrahydrofuran	10	8.50	ppbv	85			70 (70)	130 (130)	
	Trichlorofluoromethane	10	11.0	ppbv	110			40 (70)	160 (130)	
	1,1,2-Trichlorotrifluoroethane	10	10.2	ppbv	102			70 (70)	130 (130)	
	Dichlorotetrafluoroethane	10	10.6	ppbv	106			70 (70)	130 (130)	
	tert-Butyl Alcohol	10	9.10	ppbv	91			70 (70)	130 (130)	
	Heptane	10	8.10	ppbv	81			70 (70)	130 (130)	
	1,1-Dichloroethene	10	10.2	ppbv	102			70 (70)	130 (130)	
	Acetone	10	10.3	ppbv	103			40 (70)	160 (130)	
	Carbon disulfide	10	10.5	ppbv	105			40 (70)	160 (130)	
	Methyl tert-butyl Ether	10	8.10	ppbv	81			70 (70)	130 (130)	
	Methylene Chloride	10	12.4	ppbv	124			70 (70)	130 (130)	
	trans-1,2-Dichloroethene	10	9.60	ppbv	96			70 (70)	130 (130)	
	1,1-Dichloroethane	10	10.0	ppbv	100			70 (70)	130 (130)	
	Cyclohexane	10	7.90	ppbv	79			70 (70)	130 (130)	
	2-Butanone	10	9.00	ppbv	90			40 (70)	160 (130)	
	Carbon Tetrachloride	10	9.40	ppbv	94			70 (70)	130 (130)	
	cis-1,2-Dichloroethene	10	7.70	ppbv	77			70 (70)	130 (130)	
	Chloroform	10	8.60	ppbv	86			70 (70)	130 (130)	
	1,1,1-Trichloroethane	10	8.00	ppbv	80			70 (70)	130 (130)	
	2,2,4-Trimethylpentane	10	8.80	ppbv	88			70 (70)	130 (130)	
	Benzene	10	8.70	ppbv	87			70 (70)	130 (130)	
	1,2-Dichloroethane	10	8.70	ppbv	87			70 (70)	130 (130)	
	Trichloroethene	10	8.10	ppbv	81			70 (70)	130 (130)	
	1,2-Dichloropropane	10	8.60	ppbv	86			70 (70)	130 (130)	
	Bromodichloromethane	10	9.00	ppbv	90			70 (70)	130 (130)	
	4-Methyl-2-Pentanone	10	9.30	ppbv	93			40 (70)	160 (130)	
	Toluene	10	8.00	ppbv	80			70 (70)	130 (130)	
	t-1,3-Dichloropropene	10	7.30	ppbv	73			70 (70)	130 (130)	
	cis-1,3-Dichloropropene	10	7.90	ppbv	79			70 (70)	130 (130)	
	1,1,2-Trichloroethane	10	8.90	ppbv	89			70 (70)	130 (130)	
	Dibromochloromethane	10	8.80	ppbv	88			70 (70)	130 (130)	
	1,2-Dibromoethane	10	8.50	ppbv	85			70 (70)	130 (130)	
	Tetrachloroethene	10	8.30	ppbv	83			70 (70)	130 (130)	
	Chlorobenzene	10	9.30	ppbv	93			70 (70)	130 (130)	
	Ethyl Benzene	10	8.50	ppbv	85			70 (70)	130 (130)	
	m/p-Xylene	20	18.3	ppbv	92			70 (70)	130 (130)	
	o-Xylene	10	9.10	ppbv	91			70 (70)	130 (130)	
	Styrene	10	8.30	ppbv	83			70 (70)	130 (130)	
	Bromoform	10	9.20	ppbv	92			70 (70)	130 (130)	
	1,1,2,2-Tetrachloroethane	10	9.90	ppbv	99			70 (70)	130 (130)	
	2-Chlorotoluene	10	8.60	ppbv	86			70 (70)	130 (130)	
	1,3,5-Trimethylbenzene	10	8.90	ppbv	89			70 (70)	130 (130)	
	1,2,4-Trimethylbenzene	10	9.20	ppbv	92			70 (70)	130 (130)	

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1586
Client: RTP Environmental
Analytical Method: SWTO-15 Datafile : VL042152.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0317ABS01	1,3-Dichlorobenzene	10	9.20	ppbv	92			70 (70)	130 (130)	
	1,4-Dichlorobenzene	10	9.20	ppbv	92			70 (70)	130 (130)	
	1,2-Dichlorobenzene	10	9.40	ppbv	94			70 (70)	130 (130)	
	1,2,4-Trichlorobenzene	10	9.20	ppbv	92			70 (70)	130 (130)	
	Hexachloro-1,3-butadiene	10	8.50	ppbv	85			70 (70)	130 (130)	
	Naphthalene	10	11.0	ppbv	110			40 (70)	160 (130)	
	1,3-Butadiene	10	10.6	ppbv	106			70 (70)	130 (130)	
	4-Ethyltoluene	10	9.10	ppbv	91			70 (70)	130 (130)	
	Hexane	10	9.10	ppbv	91			70 (70)	130 (130)	
	Allyl Chloride	10	10.0	ppbv	100			70 (70)	130 (130)	
	1,4-Dioxane	10	8.50	ppbv	85			40 (70)	160 (130)	
	Methyl methacrylate	10	8.00	ppbv	80			70 (70)	130 (130)	

Duplicate Sample Summary

Lab Sample Id :	Q1586-02DUP	Q1586-02
Client Id :	557-SG-1DUP	557-SG-1
DF :	1	1
Datafile :	VL042160.D	VL042159.D
Anal Date & Time :	03/17/2025 16:29	03/17/2025 15:53

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0	0	0
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0	0	0
2-Butanone	0.63	0.68	7.6
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0	0	0
Acetone	8.3	8.5	2.4
Allyl Chloride	0	0	0
Benzene	0.11	0.11	0
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	1.6	1.5	6.5
Carbon Tetrachloride	0.08	0.08	0

Duplicate Sample Summary

Lab Sample Id :	Q1586-02DUP	Q1586-02
Client Id :	557-SG-1DUP	557-SG-1
DF :	1	1
Datafile :	VL042160.D	VL042159.D
Anal Date & Time :	03/17/2025 16:29	03/17/2025 15:53

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0.75	0.78	3.9
Chloromethane	0	0	0
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.54	0.54	0
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	0	0	0
Heptane	0	0	0
Hexachloro-1,3-Butadiene	0	0	0
Hexane	0.24	0.29	18.9
m/p-Xylene	0	0	0
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	0.42	0.5	17.4
Naphthalene	0	0	0
o-Xylene	0	0	0
Styrene	0	0	0
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0.2	0.24	18.2
Tetrachloroethene	3.1	3.1	0
Tetrahydrofuran	0	0	0
Toluene	0	0	0
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.3	0.31	3.3
Vinyl Chloride	0	0	0

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VL0317ABL01

Lab Name: CHEMTECH

Contract: RTPE01

Lab Code: CHEM Case No.: Q1586

SAS No.: Q1586 SDG NO.: Q1586

Lab File ID: VL042151.D

Lab Sample ID: VL0317ABL01

Date Analyzed: 03/17/2025

Time Analyzed: 10:50

GC Column: RTX-1 ID: 0.32 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_L

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VL0317ABS01	VL0317ABS01	VL042152.D	03/17/2025
565-SG-1	Q1586-01	VL042154.D	03/17/2025
565-SG-1DL	Q1586-01DL	VL042155.D	03/17/2025
557-SG-2DL	Q1586-03DL	VL042157.D	03/17/2025
557-SG-2	Q1586-03	VL042158.D	03/17/2025
557-SG-1	Q1586-02	VL042159.D	03/17/2025
557-SG-1DUP	Q1586-02DUP	VL042160.D	03/17/2025
565-SG-1DL2	Q1586-01DL2	VL042161.D	03/17/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: RTPE01
Lab Code: CHEM Case No.: Q1586 SAS No.: Q1586 SDG NO.: Q1586
Lab File ID: VL042040.D BFB Injection Date: 02/25/2025
Instrument ID: MSVOA_L BFB Injection Time: 07:56
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	22.6
75	30.0 - 66.0% of mass 95	55.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.7 (1.1) 1
174	50.0 - 120.0% of mass 95	63.3
175	4.0 - 9.0% of mass 174	4.9 (7.8) 1
176	93.0 - 101.0% of mass 174	60.9 (96.1) 1
177	5.0 - 9.0% of mass 176	3.4 (5.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICCC010	VSTDICCC010	VL042041.D	02/25/2025	08:56
VSTDICC002	VSTDICC002	VL042042.D	02/25/2025	09:29
VSTDICC001	VSTDICC001	VL042043.D	02/25/2025	10:00
VSTDICC0.5	VSTDICC0.5	VL042044.D	02/25/2025	10:31
VSTDICC0.1	VSTDICC0.1	VL042045.D	02/25/2025	11:03
VSTDICC0.03	VSTDICC0.03	VL042046.D	02/25/2025	11:34
VSTDICC015	VSTDICC015	VL042047.D	02/25/2025	12:07

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: RTPE01
Lab Code: CHEM Case No.: Q1586 SAS No.: Q1586 SDG NO.: Q1586
Lab File ID: VL042149.D BFB Injection Date: 03/17/2025
Instrument ID: MSVOA_L BFB Injection Time: 09:09
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	23.3
75	30.0 - 66.0% of mass 95	53.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.9 (1.4) 1
174	50.0 - 120.0% of mass 95	63.5
175	4.0 - 9.0% of mass 174	4.9 (7.7) 1
176	93.0 - 101.0% of mass 174	61 (96) 1
177	5.0 - 9.0% of mass 176	4 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC010	VSTDCCC010	VL042150.D	03/17/2025	09:55
VL0317ABL01	VL0317ABL01	VL042151.D	03/17/2025	10:50
VL0317ABS01	VL0317ABS01	VL042152.D	03/17/2025	11:38
565-SG-1	Q1586-01	VL042154.D	03/17/2025	12:51
565-SG-1DL	Q1586-01DL	VL042155.D	03/17/2025	13:37
557-SG-2DL	Q1586-03DL	VL042157.D	03/17/2025	14:41
557-SG-2	Q1586-03	VL042158.D	03/17/2025	15:19
557-SG-1	Q1586-02	VL042159.D	03/17/2025	15:53
557-SG-1DUP	Q1586-02DUP	VL042160.D	03/17/2025	16:29
565-SG-1DL2	Q1586-01DL2	VL042161.D	03/17/2025	17:01

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RTPE01
Lab Code: CHEM Case No.: Q1586 SAS No.: Q1586 SDG NO.: Q1586
Lab File ID: VL042150.D Date Analyzed: 03/17/2025
Instrument ID: MSVOA_L Time Analyzed: 09:55
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	167493	2.80	395790	3.98	337104	8.90
UPPER LIMIT	234490	3.13	554106	4.31	471946	9.23
LOWER LIMIT	100496	2.47	237474	3.65	202262	8.57
EPA SAMPLE NO.						
565-SG-1	191555	2.80	526330	3.98	471205	8.90
565-SG-1DL	177355	2.80	461268	3.98	407257	8.90
565-SG-1DL2	159357	2.80	390233	3.97	317448	8.90
557-SG-1	161407	2.80	390134	3.97	319342	8.90
557-SG-1DUP	161300	2.80	382035	3.98	319787	8.90
557-SG-2	163274	2.80	382025	3.98	334931	8.90
557-SG-2DL	165500	2.80	386181	3.98	319433	8.90
VL0317ABL01	165240	2.80	385222	3.98	316907	8.90
VL0317ABS01	168331	2.80	405507	3.98	339099	8.90

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
AREA LOWER LIMIT = -40% of internal standard area
RT UPPER LIMIT = +0.33 minutes of internal standard RT
RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	RTP Environmental	Date Collected:	
Project:	Yellow Cab - TO-15	Date Received:	
Client Sample ID:	VL0317ABL01	SDG No.:	Q1586
Lab Sample ID:	VL0317ABL01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042151.D	1		03/17/25 10:50	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.21	1.04	U	1.04	2.47	ug/m3
74-87-3	Chloromethane	0.10	0.21	U	0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.14	0.54	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.17	0.45	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.13	0.38	U	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.16	0.90	U	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.16	1.23	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.090	0.63	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.19	0.58	U	0.58	1.52	ug/m3
142-82-5	Heptane	0.11	0.45	U	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
67-64-1	Acetone	0.24	0.57	U	0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.16	0.50	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.12	0.43	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	0.19	0.66	U	0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.19	0.75	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	0.12	0.35	U	0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.010	0.060	U	0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.090	0.36	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.080	0.39	U	0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.10	0.47	U	0.47	2.34	ug/m3
71-43-2	Benzene	0.090	0.29	U	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.11	0.51	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.080	0.54	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.090	0.37	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.11	0.41	U	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.070	0.38	U	0.38	2.73	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	
Project:	Yellow Cab - TO-15	Date Received:	
Client Sample ID:	VL0317ABL01	SDG No.:	Q1586
Lab Sample ID:	VL0317ABL01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042151.D	1		03/17/25 10:50	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.070	0.60	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.070	0.54	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.020	0.14	U	0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.12	0.52	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.21	0.91	U	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.12	0.52	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.11	0.47	U	0.47	2.13	ug/m3
75-25-2	Bromoform	0.060	0.62	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.010	0.070	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.11	0.57	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.11	0.54	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.080	0.39	U	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.050	0.30	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.090	0.67	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.090	0.96	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.13	0.29	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.080	0.42	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.12	0.59	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.11	0.39	U	0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.15	0.47	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.21	0.76	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.090	0.37	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	9.80			70 (65) - 130 (135)	98%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	165000		2.8			
540-36-3	1,4-Difluorobenzene	385000		3.975			
3114-55-4	Chlorobenzene-d5	317000		8.901			

Report of Analysis

Client:	RTP Environmental	Date Collected:	
Project:	Yellow Cab - TO-15	Date Received:	
Client Sample ID:	VL0317ABL01	SDG No.:	Q1586
Lab Sample ID:	VL0317ABL01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042151.D	1		03/17/25 10:50	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	RTP Environmental	Date Collected:	
Project:	Yellow Cab - TO-15	Date Received:	
Client Sample ID:	VL0317ABS01	SDG No.:	Q1586
Lab Sample ID:	VL0317ABS01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042152.D	1		03/17/25 11:38	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	10.3	50.9		1.04	2.47	ug/m3
74-87-3	Chloromethane	11.0	22.7		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	11.1	28.4		0.030	0.080	ug/m3
74-83-9	Bromomethane	11.1	43.1		0.54	1.94	ug/m3
75-00-3	Chloroethane	10.7	28.2		0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	8.50	25.1		0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	11.0	61.8		0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	10.2	78.2		1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	10.6	74.1		0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	9.10	27.6		0.58	1.52	ug/m3
142-82-5	Heptane	8.10	33.2		0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	10.2	40.4		0.56	1.98	ug/m3
67-64-1	Acetone	10.3	24.5		0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	10.5	32.7		0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	8.10	29.2		0.43	1.80	ug/m3
75-09-2	Methylene Chloride	12.4	43.1		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	9.60	38.1		0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	10.0	40.5		0.53	2.02	ug/m3
110-82-7	Cyclohexane	7.90	27.2		0.76	1.72	ug/m3
78-93-3	2-Butanone	9.00	26.5		0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	9.40	59.1		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	7.70	30.5		0.36	1.98	ug/m3
67-66-3	Chloroform	8.60	42.0		0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	8.00	43.6		0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	8.80	41.1		0.47	2.34	ug/m3
71-43-2	Benzene	8.70	27.8		0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	8.70	35.2		0.36	2.02	ug/m3
79-01-6	Trichloroethene	8.10	43.5		0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	8.60	39.8		0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	9.00	60.3		0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	9.30	38.1		0.37	2.05	ug/m3
108-88-3	Toluene	8.00	30.1		0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	7.30	33.1		0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	7.90	35.9		0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	8.90	48.6		0.38	2.73	ug/m3

Report of Analysis

Client:	RTP Environmental	Date Collected:	
Project:	Yellow Cab - TO-15	Date Received:	
Client Sample ID:	VL0317ABS01	SDG No.:	Q1586
Lab Sample ID:	VL0317ABS01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042152.D	1		03/17/25 11:38	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	8.80	75.0		0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	8.50	65.3		0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	8.30	56.3		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	9.30	42.8		0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	8.50	36.9		0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	18.3	79.5		0.91	4.34	ug/m3
95-47-6	o-Xylene	9.10	39.5		0.52	2.17	ug/m3
100-42-5	Styrene	8.30	35.3		0.47	2.13	ug/m3
75-25-2	Bromoform	9.20	95.1		0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	9.90	68.0		0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	8.60	44.5		0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	8.90	43.8		0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	9.20	45.2		0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	9.20	55.3		0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	9.20	55.3		0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	9.40	56.5		0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	9.20	68.3		0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	8.50	90.7		0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	10.6	23.4		0.29	1.11	ug/m3
91-20-3	Naphthalene	11.0	57.7		0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	9.10	44.7		0.59	2.46	ug/m3
110-54-3	Hexane	9.10	32.1		0.39	1.76	ug/m3
107-05-1	Allyl Chloride	10.0	31.3		0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	8.50	30.6		0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	8.00	32.8		0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.8			70 (65) - 130 (135)	108%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	168000		2.8			
540-36-3	1,4-Difluorobenzene	406000		3.975			
3114-55-4	Chlorobenzene-d5	339000		8.901			

Report of Analysis

Client:	RTP Environmental	Date Collected:	
Project:	Yellow Cab - TO-15	Date Received:	
Client Sample ID:	VL0317ABS01	SDG No.:	Q1586
Lab Sample ID:	VL0317ABS01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042152.D	1		03/17/25 11:38	VL031725

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements



CALIBRATION SUMMARY

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL022525AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Wed Feb 26 01:08:36 2025

Response Via : Initial Calibration

Calibration Files

0.03=VL042046.D 0.1 =VL042045.D 0.5 =VL042044.D 1 =VL042043.D 2 =VL042042.D 10 =VL042041.D 15 =VL042047.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.109	1.084	1.091	0.943	0.973	1.040	7.31
3) Chlorodifluoro...			1.493	1.566	1.535	1.311	1.362	1.453	7.67
4) Chloromethane			0.438	0.398	0.399	0.344	0.344	0.385	10.53
5) T Vinyl Chloride	0.447	0.336	0.376	0.346	0.361	0.337	0.343	0.364	10.89
6) T Bromomethane			0.170	0.184	0.186	0.160	0.165	0.173	6.63
7) Chloroethane			0.151	0.158	0.156	0.135	0.140	0.148	6.83
8) T Dichlorotetra...			0.892	0.910	0.901	0.779	0.791	0.855	7.47
9) T Propene			0.614	0.630	0.636	0.522	0.558	0.592	8.47
10) T Heptane			1.971	1.946	1.960	1.662	1.679	1.844	8.59
11) T Trichlorofluor...			1.056	1.122	1.119	0.944	0.968	1.042	7.98
12) T 1,1,2-Trichlor...			0.919	0.902	0.865	0.697	0.773	0.831	11.28
13) Ethanol			0.053	0.057	0.040	0.020	0.018	0.038#	48.50
14) T Bromoethene			0.325	0.304	0.299	0.264	0.275	0.293	8.41
15) T Acetone			1.084	0.960	0.933	0.714	0.723	0.883	18.17
16) T 1,3-Butadiene			0.476	0.435	0.418	0.358	0.367	0.411	11.96
17) tert-Butyl alc...			1.184	1.185	1.201	0.950	1.004	1.105	10.71
18) T 1,1-Dichloroet...			0.398	0.437	0.386	0.320	0.338	0.376	12.57
19) T Isopropyl Alcohol			0.561	0.571	0.557	0.419	0.429	0.507	15.06
20) T Methylene Chlo...			0.429	0.374	0.329	0.261	0.278	0.334	20.72
21) T Allyl Chloride			0.707	0.717	0.639	0.538	0.590	0.638	11.94
22) T trans-1,2-Dich...			0.481	0.506	0.450	0.358	0.395	0.438	13.90
23) T Vinyl Acetate			0.689	0.614	0.695	0.520	0.586	0.621	11.82
24) T 1,1-Dichloroet...			0.927	0.856	0.863	0.669	0.758	0.815	12.42
25) T Ethyl Acetate			3.833	3.674	3.776	3.174	3.217	3.535	8.91
26) T Hexane			1.596	1.462	1.519	1.282	1.287	1.429	9.83
27) T Carbon Disulfide			1.042	0.972	1.032	0.856	0.942	0.969	7.78
28) T Methyl tert-Bu...			0.650	0.844	0.838	0.522	0.705	0.711	19.02
29) T Chloroform			2.165	2.044	1.971	1.722	1.716	1.923	10.36
30) T Cyclohexane			1.154	1.239	1.255	1.076	1.076	1.160	7.39
31) T cis-1,2-Dichlo...			1.421	1.401	1.450	1.237	1.255	1.353	7.33
32) T 1,1,1-Trichlor...	1.921	1.875	1.883	1.916	1.944	1.771	1.751	1.866	4.05

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.833	0.799	0.812	0.692	0.759	0.779	7.15
35) T Carbon Tetrach...	0.664	0.705	0.675	0.723	0.725	0.655	0.681	0.690	4.09
36) T Benzene			1.079	1.084	1.104	0.969	0.999	1.047	5.66
37) T 1,2-Dichloroet...			0.624	0.604	0.598	0.520	0.556	0.580	7.20
38) T Trichloroethene	0.447	0.462	0.433	0.447	0.447	0.378	0.379	0.427	8.06
39) T 1,2-Dichloropr...			0.421	0.415	0.414	0.357	0.379	0.397	7.02

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\											
Method File : VL022525AIR.M											
40)	T	1,4-Dioxane			0.192	0.196	0.181	0.164	0.173	0.181	7.31
41)	T	Tetrahydrofuran			0.469	0.492	0.473	0.418	0.450	0.460	6.04
42)	T	Bromodichlorom...			0.806	0.791	0.810	0.739	0.765	0.782	3.81
43)		Methyl Methacr...			0.400	0.418	0.426	0.392	0.409	0.409	3.35
44)	T	2,2,4-Trimethy...			1.928	1.935	1.955	1.670	1.750	1.848	6.99
45)	T	t-1,3-Dichloro...			0.414	0.392	0.418	0.393	0.414	0.406	3.12
46)	T	cis-1,3-Dichlo...			0.514	0.523	0.548	0.516	0.526	0.526	2.59
47)	T	1,1,2-Trichlor...			0.426	0.429	0.426	0.361	0.375	0.403	8.12
48)	T	Dibromochlorom...			0.618	0.667	0.667	0.594	0.613	0.632	5.25
49)	T	Bromoform			0.553	0.551	0.588	0.525	0.526	0.549	4.68
50)	T	4-Methyl-2-Pen...			0.996	0.974	0.975	1.014	1.079	1.008	4.27
51)	T	2-Hexanone			0.577	0.570	0.608	0.834	0.879	0.693	21.69
52)	T	Tetrachloroethene	0.322	0.374	0.380	0.383	0.385	0.327	0.337	0.358	7.88
53)	T	Toluene			1.150	1.210	1.284	1.088	1.118	1.170	6.67
54)	T	1,2-Dibromoethane		0.563	0.556	0.553	0.586	0.500	0.514	0.545	5.87
-----ISTD-----											
55)	I	Chlorobenzene-d5									
56)		1,1,1,2-Tetrac...			0.607	0.596	0.595	0.505	0.524	0.565	8.37
57)	T	Chlorobenzene			1.102	1.113	1.121	0.910	0.938	1.037	10.02
58)	T	Ethyl Benzene			1.886	1.906	1.974	1.694	1.749	1.842	6.31
59)	T	m/p-Xylene			1.488	1.550	1.580	1.330	1.380	1.466	7.35
60)	T	o-Xylene			1.572	1.550	1.577	1.317	1.351	1.473	8.68
61)	T	Styrene			0.659	0.719	0.738	0.658	0.690	0.693	5.18
62)		Isopropylbenzene			2.299	2.339	2.399	1.969	2.026	2.207	8.84
63)	T	1,1,2,2-Tetrac...	0.999	0.975	1.026	1.041	1.031	0.848	0.872	0.970	8.13
64)		n-propylbenzene			0.596	0.607	0.632	0.524	0.535	0.579	8.11
65)		tert-Butylbenzene			2.080	2.251	2.222	1.733	1.753	2.008	12.47
66)	T	Benzyl Chloride			0.141	0.159	0.163	0.153	0.150	0.153	5.40
67)		sec-Butylbenzene			2.973	3.229	3.182	2.482	2.494	2.872	12.65
68)	S	1-Bromo-4-Fluo...	0.773	0.760	0.768	0.777	0.765	0.764	0.779	0.769	0.92
69)		p-Isopropyltol...			2.240	2.533	2.575	2.021	2.056	2.285	11.38
70)		n-Butylbenzene			2.332	2.567	2.572	2.070	2.120	2.332	10.20
71)		2-Chlorotoluene			1.758	1.829	1.853	1.514	1.551	1.701	9.30
72)	T	4-Ethyltoluene			1.775	1.944	1.985	1.621	1.670	1.799	8.99
73)	T	1,3,5-Trimethy...			1.592	1.732	1.696	1.381	1.400	1.560	10.47
74)	T	1,2,4-Trimethy...			1.855	1.982	1.990	1.471	1.488	1.757	14.76
75)	T	1,3-Dichlorobe...			1.112	1.110	1.083	0.869	0.882	1.011	12.30
76)	T	1,4-Dichlorobe...			1.120	1.061	1.092	0.861	0.886	1.004	12.07
77)	T	1,2-Dichlorobe...			1.059	1.097	1.063	0.836	0.852	0.981	12.88
78)	T	Hexachloro-1,3...			0.840	0.832	0.844	0.637	0.666	0.764	13.53
79)	T	Naphthalene		0.841	0.978	1.237	1.352	1.371	1.450	1.205	20.15
80)	T	Naphthalene,2-...			0.090	0.142	0.240	0.562	0.615	0.330	73.74
81)	T	1,2,4-Trichlor...			0.589	0.687	0.765	0.644	0.673	0.671	9.58

(#) = Out of Range

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RTPE01

Lab Code: CHEM Case No.: Q1586 SAS No.: Q1586 SDG No.: Q1586

Instrument ID: MSVOA_L Calibration Date/Time: 03/17/2025 09:55

Lab File ID: VL042150.D Init. Calib. Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:56 12:07

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	1.040	1.104		6.15	30
Chloromethane	0.385	0.433		12.47	30
Vinyl Chloride	0.364	0.409		12.36	30
Bromomethane	0.173	0.195		12.72	30
Chloroethane	0.148	0.159		7.43	30
Tetrahydrofuran	0.460	0.400		-13.04	30
Trichlorofluoromethane	1.042	1.203		15.45	30
1,1,2-Trichlorotrifluoroethane	0.831	0.856		3.01	30
Dichlorotetrafluoroethane	0.855	0.953		11.46	30
tert-Butyl alcohol	1.105	1.010		-8.6	30
Heptane	1.844	1.498		-18.76	30
1,1-Dichloroethene	0.376	0.388		3.19	30
Acetone	0.883	0.922		4.42	30
Carbon Disulfide	0.969	1.008		4.03	30
Methyl tert-Butyl Ether	0.711	0.635		-10.69	30
Methylene Chloride	0.334	0.398		19.16	30
trans-1,2-Dichloroethene	0.438	0.423		-3.42	30
1,1-Dichloroethane	0.815	0.857		5.15	30
Cyclohexane	1.160	0.925		-20.26	30
2-Butanone	0.779	0.708		-9.11	30
Carbon Tetrachloride	0.690	0.677		-1.88	30
cis-1,2-Dichloroethene	1.353	1.048		-22.54	30
Chloroform	1.923	1.681		-12.59	30
1,1,1-Trichloroethane	1.866	1.524		-18.33	30
2,2,4-Trimethylpentane	1.848	1.697		-8.17	30
Benzene	1.047	0.936		-10.6	30
1,2-Dichloroethane	0.580	0.528		-8.97	30
Trichloroethene	0.427	0.353		-17.33	30
1,2-Dichloropropane	0.397	0.353		-11.08	30
Bromodichloromethane	0.782	0.708		-9.46	30
4-Methyl-2-Pentanone	1.008	0.944		-6.35	30
Toluene	1.170	0.964		-17.61	30
t-1,3-Dichloropropene	0.406	0.291		-28.33	30
cis-1,3-Dichloropropene	0.526	0.418		-20.53	30
1,1,2-Trichloroethane	0.403	0.367		-8.93	30
Dibromochloromethane	0.632	0.567		-10.28	30
1,2-Dibromoethane	0.545	0.470		-13.76	30
Tetrachloroethene	0.358	0.299		-16.48	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RTPE01

Lab Code: CHEM Case No.: Q1586 SAS No.: Q1586 SDG No.: Q1586

Instrument ID: MSVOA_L Calibration Date/Time: 03/17/2025 09:55

Lab File ID: VL042150.D Init. Calib. Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:56 12:07

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Chlorobenzene	1.037	0.983		-5.21	30
Ethyl Benzene	1.842	1.586		-13.9	30
m/p-Xylene	1.466	1.362		-7.09	30
o-Xylene	1.473	1.372		-6.86	30
Styrene	0.693	0.587		-15.3	30
Bromoform	0.549	0.506		-7.83	30
1,1,2,2-Tetrachloroethane	0.970	0.979		0.93	30
2-Chlorotoluene	1.701	1.514		-10.99	30
1,3,5-Trimethylbenzene	1.560	1.411		-9.55	30
1,2,4-Trimethylbenzene	1.757	1.648		-6.2	30
1,3-Dichlorobenzene	1.011	0.975		-3.56	30
1,4-Dichlorobenzene	1.004	0.943		-6.08	30
1,2-Dichlorobenzene	0.981	0.958		-2.35	30
1,2,4-Trichlorobenzene	0.671	0.625		-6.86	30
Hexachloro-1,3-Butadiene	0.764	0.670		-12.3	30
1,3-Butadiene	0.411	0.437		6.33	30
Naphthalene	1.205	1.391		15.44	30
4-Ethyltoluene	1.799	1.677		-6.78	30
1-Bromo-4-Fluorobenzene	0.769	0.851		10.66	30
Hexane	1.429	1.284		-10.15	30
Allyl Chloride	0.638	0.628		-1.57	30
1,4-Dioxane	181.383	163.107		-10.08	30
Methyl Methacrylate	0.409	0.330		-19.32	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



SHIPPING DOCUMENTS

Client Contact Information				Bottle Order ID : B2502031				Courier :				1 of 3 COCs							
Client ID : RTPE01 Project ID : Air				Sampler Name(s) : 06				Analysis				Matrix							
Customer Name : RTP Environmental Address : 239 US Highway 22 East City : Green Brook State : NJ Zip Code : 08812 Country :				Project Manager : DAVID GABEL				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified											
				Phone Number : 732-968-9600															
				Fax Number :															
				Site Details: Yellow Cab															
Analysis Turnaround Time				Data Package Type : F-11 deliverable															
Standard : 10 business days OR				EDD Type : Per NJ Hazmat															
Rush (Specify): Days																			
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15			Indoor/Ambient Air	Soil Gas
565-56-1	2/14/25	1130	1230	-30	-7	50	50	-30	-7.5	10774	10448	6 L	100	VL041853.D	X			X	
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) <i>[Signature]</i>									
		Ambient	Maximum	Minimum															
Start																			
Stop																			
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.									
		Ambient	Maximum	Minimum															
Start																			
Stop																			
Special Instructions/QC Requirements & Comments :																			
Suspected Contamination: High Medium Low PID Readings:																			
Sampling site (State):																			
Quick Connector required : NO																			
Canisters Shipped by: <i>[Signature]</i>				Date/Time: 3/12/25				Canisters Received by: <i>[Signature]</i>				Date/Time:				B2502031 - 1			
Samples Relinquished by: <i>[Signature]</i>				Date/Time: 3/11/25 1600				Received by: <i>[Signature]</i>				Date/Time: 3-14-25 16:00							
Relinquished by:				Date/Time:				Received by:				Date/Time:							

Client Contact Information				Bottle Order ID : B2502031				Courier : <i>De</i>				<u>2</u> of <u>3</u> COCs								
Client ID : RTPE01 Project ID : Air				Sampler Name(s) : <i>OG</i>				Analysis				Matrix								
Customer Name : RTP Environmental Address : 239 US Highway 22 East City : Green Brook State : NJ Zip Code : 08812 Country :				Project Manager : DAVID GABEL				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified												
				Phone Number : 732-968-9600																
				Fax Number :																
				Site Details: <i>Yellow Cab</i>																
Analysis Turnaround Time				Standard : <u>10 business days</u> OR				Data Package Type : <i>Full deliverable</i>												
Rush (Specify): <u>Days</u>				EDD Type : <i>NJ Hazsite</i>																
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15				Indoor/Ambient Air	Soil Gas
<i>557-36-1</i>	<i>3/14/25</i>	<i>1341</i>	<i>1441</i>	<i>-30</i>	<i>-10</i>	<i>65</i>	<i>65</i>	<i>-30</i>	<i>-8.0</i>	<i>10633</i>	<i>10609</i>	<i>6 L</i>	<i>100</i>	<i>VL041853.D</i>	<i>X</i>				<i>X</i>	
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) <i>Sault</i>										
		Ambient	Maximum	Minimum																
Start																				
Stop																				
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.										
		Ambient	Maximum	Minimum																
Start																				
Stop																				
Special Instructions/QC Requirements & Comments :																				
Suspected Contamination: High Medium Low PID Readings:																				
Sampling site (State):																				
Quick Connector required : <i>NO</i>																				
Canisters Shipped by: <i>Sam</i>					Date/Time: <i>03/12/25</i>					Canisters Received by:					Date/Time:					
Samples Relinquished by: <i>David Gabel</i>					Date/Time: <i>2/14/25 1600</i>					Received by: <i>OG</i>					Date/Time: <i>3-14-25 10:00</i>					
Relinquished by:					Date/Time:					Received by:					Date/Time:					
B2502031 - 3																				

Client Contact Information				Bottle Order ID : B2502031				Courier :				3 of 3 COCs					
Client ID : RTPE01 Project ID : Air				Sampler Name(s) : DG				Analysis				Matrix					
Customer Name : RTP Environmental Address : 239 US Highway 22 East City : Green Brook State : NJ Zip Code : 08812 Country :				Project Manager : DAVID GABEL				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified									
				Phone Number : 732-968-9600													
				Fax Number :													
				Site Details: Yellow Cab													
Analysis Turnaround Time				Standard : 10 business days OR				Data Package Type : Full deliverable									
Rush (Specify): Days				EDD Type : NJ Hazmat													
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas
557-56-2	3/14/25	1340	1440	-30	-5	65	65	-30	-5.1	10188	10605	6 L	100	VL041853.D	X		X
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) 							
	Ambient	Maximum	Minimum														
Start																	
Stop																	
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
	Ambient	Maximum	Minimum														
Start																	
Stop																	
Special Instructions/QC Requirements & Comments :																	
Suspected Contamination: High Medium Low PID Readings:																	
Sampling site (State):																	
Quick Connector required : NO																	
Canisters Shipped by: Sam				Date/Time: 03/12/25				Canisters Received by:				Date/Time:					
Samples Relinquished by: 3/14/25 1600				Date/Time: 3/14/25				Received by: DR				Date/Time: 3/14-25 16:00					
Relinquished by:				Date/Time:				Received by:				Date/Time:					

From: David Gabel <gabel@rtpenv.com>
Sent: Monday, March 17, 2025 11:18 AM
To: Data-EWR@alliancetg.com
Cc: YAZMEEN@CHEMTECH.NET
Subject: RE: Login Summary Details For Project Yellow Cab - TO-15-Q1586.

Hi Yazmeen, can you change the sample IDs. The 56 should be the letters "SG" for Soil Gas.

Sample IDs should be 565-SG-1, 557-SG-1, and 557-SG-2. Thanks

Dave

David Gabel

Sr. Project Manager, LSRP
RTP Environmental Associates, Inc.
239 U.S. Highway 22 East, Green Brook, NJ 08812
Office: 732-968-9600 | Cell: 732-996-3716
Email: gabel@rtpenv.com | Website: www.rtpenv.com



From: Data-EWR@alliancetg.com <Data-EWR@alliancetg.com>
Sent: Monday, March 17, 2025 11:12 AM
To: David Gabel <gabel@rtpenv.com>
Cc: YAZMEEN@CHEMTECH.NET
Subject: Login Summary Details For Project Yellow Cab - TO-15-Q1586.



To David Gabel;

Please see the attached Login Summary for the following project, or download the file using your login credentials from the link below.

Order ID	: Q1586
Project ID	: Yellow Cab - TO-15
Download File	: https://chemtech.net/secureLogin.aspx
Order Date	: 3/17/2025 9:31:17 AM

CHEMTECH's Project Manager : YAZMEEN GOMEZ , YAZMEEN@CHEMTECH.NET , 908-357-0579 Ext :3149

CHEMTECH's Sales Executive : Jordan Hedvat , jordan@chemtech.net , 908-728-3144 Ext :

Thank you for the opportunity to provide you with our services. For any questions please feel free to contact your project manager.

Click Here for our short online customer Survey <http://chemtech.net/ClientSurvey.aspx>.

Thank you,

Alliance Technical Group LLC.

Notice: The information transmitted in this e-mail message and in any attachments is intended Solely for the attention of the named addressee(s) and may contain confidential and/or privileged material. Any review, retransmission, dissemination or other use of, or taking of any action in reliance upon, this information by persons or entities other than the intended recipient is strictly prohibited and may be unlawful. If you have received this transmission in error, please notify us immediately by return e-mail, and permanently delete this transmission, including attachments if any, from any computer.

Laboratory Certification



Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

Internal Chain of Custody

Instructions: Use 1 form for each 20 samples of aliquot

Laboratory Person Breaking Field Seal on Sample Shuttle & Accepting Responsibility for Sample			
Laboratory: <u>Chemtech</u> GORGE Field Sample Seal No.: <u>Q1586</u> Case No.: <u>Yellow Cab - TO-15</u>	Location: <u>284 Sheffield Street, Mountainside, NJ 7092</u> Title: <u>Sample Custodian</u> Date Broken: <u>3/14/2025</u> Military Time Seal Broken: <u>16:00:00</u> Analytical Parameter/Fraction: <u>TO-15</u>		


Sample No.	Aliquot/Extract No.	Sample No.	Aliquot/Extract No.
Q1586-01	565-56-1 ^{SG}		
Q1586-02	557-56-1 ^{SG}		
Q1586-03	557-56-2 ^{SG}		

Date	Time	Relinquished By	Received By	Purpose of Change of Custody
3/17	1015	Signature 	Signature 	
		Printed Name <u>GORGE N.</u>	Printed Name <u>Sam Scanlon</u>	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	

Distribution: White - Original (Sent With Report)

Yellow - Contractor Archive

Pink - Sample Custodian - Interim Copy

AIR SAMPLE PRESSURE & DILUTION LOGBOOKAnalyst Signature: Supervisor Signature: 

METHOD: TO-15

Pressure Gauge ID: A255971

Date	Sample Number	Canister #	Initial Pressure psia	Initial Pressure Hg	Final Pressure psia	Final Pressure Hg	Dilution Factor	Comment
03/15/25	Q1482-04	10721	1.0		30		30x	sy
03/06/25	Q1506-01	10054	12.6	-4.3				sy
"	Q1506-02	10329	12.2	-5.1				sy
03/07/25	Q1532-01	10303	12.4	-4.7				↓
↓	Q1532-02	10590	13.4	-2.7				
↓	Q1532-03	10279	12.6	-4.3				
↓	Q1532-04	10295	12.4	-4.7				
03/14/25	Q1533-01	10296	11.6	-6.3				sy
↓	Q1533-02	10321	11.2	-7.1				↓
↓	Q1533-03	10410	12.2	-5.1				
↓	Q1533-04	10266	12.4	-4.7				
↓	Q1533-05	10153	12.8	-3.9				
03/14/25	Q1586-01	10448	11.0	-7.5				sy
↓	Q1586-02	10609	10.8	-8.0				↓
↓	Q1586-03	10605	12.2	-5.1				

10448

V-2

284 Sheffield Street, Mountainside, NJ 07092 P: (908) 789-8922

CHEMTECH

DL 565-56-1

Client Sample ID #: 565-56-1
Client Name: RTP
Project Name: Yellow Cab
Date: 3/14/25 Time: 1230
Analysis: TU-15
Comments: _____
CHEMTECH'S SAMPLE ID: _____
Date Received: _____ Expiration Date: _____ Date of: _____

Storage Location: F11
Sample: Q1586-01
Cust #: 565-56-1

B250203/

60901

C-8

284 Sheffield Street, Mountainside, NJ 07092 P: (908) 789-8922

CHEMTECH

Client Sample ID #: 557-56-1
Client Name: RTP
Project Name: Yellow Cab
Date: 3/14/25 Time: 1441
Analysis: TU-15
Comments: _____
CHEMTECH'S SAMPLE ID: _____
Date Received: _____ Expiration Date: _____ Date of: _____

Storage Location: F11
Sample: Q1586-02
Cust #: 557-56-1

B250203/

CHEMTECH

284 Sheffield Street, Mountainside, NJ 07092 P: (908) 789-8922

Client Sample ID #: 557-56-2
Client Name: RTP
Project Name: Yellow Cab
Date: 3/14/25 Time: 1440
Analysis: TU-15
Comments: _____

CHEMTECH'S SAMPLE ID: _____
Date Received: _____ Expiration Date: _____ Date of: _____

Storage Location: F11
Sample: Q1586-03
Cust #: 557-56-2

8922

B250203/