

Duplicate Sample Summary

Lab Sample Id :	Q1586-02DUP	Q1586-02
Client Id :	557-SG-1DUP	557-SG-1
DF :	1	1
Datafile :	VL042160.D	VL042159.D
Anal Date & Time :	03/17/2025 16:29	03/17/2025 15:53

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0	0	0
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0	0	0
2-Butanone	0.63	0.68	7.6
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0	0	0
Acetone	8.3	8.5	2.4
Allyl Chloride	0	0	0
Benzene	0.11	0.11	0
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	1.6	1.5	6.5
Carbon Tetrachloride	0.08	0.08	0

Duplicate Sample Summary

Lab Sample Id :	Q1586-02DUP	Q1586-02
Client Id :	557-SG-1DUP	557-SG-1
DF :	1	1
Datafile :	VL042160.D	VL042159.D
Anal Date & Time :	03/17/2025 16:29	03/17/2025 15:53

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0.75	0.78	3.9
Chloromethane	0	0	0
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.54	0.54	0
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	0	0	0
Heptane	0	0	0
Hexachloro-1,3-Butadiene	0	0	0
Hexane	0.24	0.29	18.9
m/p-Xylene	0	0	0
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	0.42	0.5	17.4
Naphthalene	0	0	0
o-Xylene	0	0	0
Styrene	0	0	0
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0.2	0.24	18.2
Tetrachloroethene	3.1	3.1	0
Tetrahydrofuran	0	0	0
Toluene	0	0	0
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.3	0.31	3.3
Vinyl Chloride	0	0	0