

### Initial and Continuing Calibration Verification

**Client:** Weston Solutions

**SDG No.:** Q1588

**Project:** Ft Meade Tipton Airfield Parcel RI - PO 0111169

**RunNo.:** LB135085

| Analyte           |      | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|-------------------|------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID:<br>TOC | ICV1 | mg/L  | 9.7    | 10         | 97            | 90-110                    | 03/11/2025       |
| Sample ID:<br>TOC | CCV1 | mg/L  | 10.2   | 10         | 102           | 90-110                    | 03/19/2025       |
| Sample ID:<br>TOC | CCV2 | mg/L  | 10.8   | 10         | 108           | 90-110                    | 03/19/2025       |