

Matrix Spike Summary

Client:	Weston Solutions	SDG No.:	Q1588
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Sample ID:	Q1588-01
Client ID:	TAPFTA-MW01D-031425-00-T1MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/L	75-125	22.5	OR	10.4		10	1	121		03/19/2025

Matrix Spike Summary

Client:	Weston Solutions	SDG No.:	Q1588
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Sample ID:	Q1588-01
Client ID:	TAPFTA-MW01D-031425-00-T1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/L	75-125	21.6	OR	10.4		10	1	112		03/19/2025