

Laboratory Control Sample Summary

Client:	Weston Solutions	SDG No.:	Q1588
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Run No.:	LB135085

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135085BSW							
TOC	mg/L	10	10.5		105	1	90-110	03/19/2025