

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135071

Review By	Mohan	Review On	3/20/2025 11:18:38 AM
Supervise By	jaswal	Supervise On	3/20/2025 1:26:45 PM

STD. NAME	STD REF.#
ICAL Standard	MP84901,MP84902,MP84903,MP84904,MP84905,MP84906
ICV Standard	MP84907
CCV Standard	MP84909
ICSA Standard	
CRI Standard	MP84911
LCS Standard	
Chk Standard	MP84908,MP84910,MP84912,MP84898

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	03/18/25 13:17		Mohan	OK
2	S0.2	S0.2	CAL2	03/18/25 13:19		Mohan	OK
3	S2.5	S2.5	CAL3	03/18/25 13:22		Mohan	OK
4	S5	S5	CAL4	03/18/25 13:32		Mohan	OK
5	S7.5	S7.5	CAL5	03/18/25 13:34		Mohan	OK
6	S10	S10	CAL6	03/18/25 13:39		Mohan	OK
7	ICV47	ICV47	ICV	03/18/25 13:42		Mohan	OK
8	ICB47	ICB47	ICB	03/18/25 13:44		Mohan	OK
9	CCV23	CCV23	CCV	03/18/25 13:47		Mohan	OK
10	CCB23	CCB23	CCB	03/18/25 13:49		Mohan	OK
11	CRA	CRA	CRDL	03/18/25 13:51		Mohan	OK
12	HighStd	HighStd	HIGH STD	03/18/25 13:53		Mohan	OK
13	ChkStd	ChkStd	SAM	03/18/25 13:58		Mohan	OK
14	PB167196BL	PB167196BL	MB	03/18/25 14:01		Mohan	OK
15	PB167196BS	PB167196BS	LCS	03/18/25 14:03		Mohan	OK
16	Q1553-03	FRAC-TANK-FMI120	SAM	03/18/25 14:05		Mohan	OK
17	Q1553-03DUP	FRAC-TANK-FMI120	DUP	03/18/25 14:07		Mohan	OK
18	Q1553-03MS	FRAC-TANK-FMI120	MS	03/18/25 14:10		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135071

Review By	Mohan	Review On	3/20/2025 11:18:38 AM
Supervise By	jaswal	Supervise On	3/20/2025 1:26:45 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84901,MP84902,MP84903,MP84904,MP84905,MP84906		
ICV Standard	MP84907		
CCV Standard	MP84909		
ICSA Standard			
CRI Standard	MP84911		
LCS Standard			
Chk Standard	MP84908,MP84910,MP84912,MP84898		

19	Q1553-03MSD	FRAC-TANK-FMI120M	MSD	03/18/25 14:12		Mohan	OK
20	Q1563-01	437	SAM	03/18/25 14:14		Mohan	OK
21	CCV24	CCV24	CCV	03/18/25 14:17		Mohan	OK
22	CCB24	CCB24	CCB	03/18/25 14:19		Mohan	OK
23	Q1563-02	FERNOT-WATER	SAM	03/18/25 14:21		Mohan	OK
24	Q1576-01	MW14	SAM	03/18/25 14:23	High	Mohan	Dilution
25	Q1577-03	MOO-25-0072	SAM	03/18/25 14:26		Mohan	OK
26	Q1588-01	TAPFTA-MW01D-0314	SAM	03/18/25 14:28		Mohan	OK
27	Q1590-04	3799	SAM	03/18/25 14:30		Mohan	OK
28	Q1553-03L	FRAC-TANK-FMI120L	SD	03/18/25 14:44		Mohan	OK
29	Q1553-03A	FRAC-TANK-FMI120A	PS	03/18/25 14:47		Mohan	OK
30	Q1576-01DL	MW14DL	SAM	03/18/25 14:49	Report 2X	Mohan	Confirms
31	CCV25	CCV25	CCV	03/18/25 14:54		Mohan	OK
32	CCB25	CCB25	CCB	03/18/25 14:56		Mohan	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB135118

Review By	JANVI	Review On	3/26/2025 11:00:41 AM
Supervise By	Mohan	Supervise On	3/26/2025 2:17:51 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84682,MP84684,MP84683,MP84688,MP84687,MP84686,MP84685,MP84681,MP84689		
ICV Standard	MP84690		
CCV Standard	MP84697		
ICSA Standard	MP84692,MP84693		
CRI Standard	MP84683		
LCS Standard			
Chk Standard	MP84695,MP84696		

Sr#	SampleID	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE	TUNE	03/20/25 15:19		Jaswal	OK
2	S0	S0	CAL1	03/20/25 15:48		Jaswal	OK
3	S2	S2	CAL3	03/20/25 15:55		Jaswal	OK
4	S3	S3	CAL4	03/20/25 15:58		Jaswal	OK
5	S4	S4	CAL5	03/20/25 16:04		Jaswal	OK
6	S5	S5	CAL6	03/20/25 16:07		Jaswal	OK
7	S6	S6	CAL7	03/20/25 16:10		Jaswal	OK
8	S7	S7	CAL8	03/20/25 16:13		Jaswal	OK
9	S8	S8	CAL9	03/20/25 16:19		Jaswal	OK
10	ICV01	ICV01	ICV	03/20/25 17:16		Jaswal	OK
11	LLICV01	LLICV01	LLICV	03/20/25 17:19		Jaswal	OK
12	ICB01	ICB01	ICB	03/20/25 17:27		Jaswal	OK
13	ICSA01	ICSA01	ICSA	03/20/25 17:40		Jaswal	OK
14	ICSAB01	ICSAB01	ICSAB	03/20/25 17:43		Jaswal	OK
15	CCV01	CCV01	CCV	03/20/25 17:49		Jaswal	OK
16	CCB01	CCB01	CCB	03/20/25 17:56		Jaswal	OK
17	CRI	CRI	CRDL	03/20/25 17:59		Jaswal	OK
18	PB167152BL	PB167152BL	MB	03/20/25 18:02		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB135118

Review By	JANVI	Review On	3/26/2025 11:00:41 AM
Supervise By	Mohan	Supervise On	3/26/2025 2:17:51 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84682,MP84684,MP84683,MP84688,MP84687,MP84686,MP84685,MP84681,MP84689		
ICV Standard	MP84690		
CCV Standard	MP84697		
ICSA Standard	MP84692,MP84693		
CRI Standard	MP84683		
LCS Standard			
Chk Standard	MP84695,MP84696		

19	PB167152BS	PB167152BS	LCS	03/20/25 18:08		Jaswal	OK
20	Q1563-01DL	437DL	SAM	03/20/25 18:12	Na high	Jaswal	Dilution
21	Q1563-02DL	FERNOT-WATERDL	SAM	03/20/25 18:15		Jaswal	OK
22	Q1577-03DL	MOO-25-0072DL	SAM	03/20/25 18:22	All INT_STD Fail	Jaswal	Dilution
23	Q1577-03DL2	MOO-25-0072DL2	SAM	03/20/25 18:29	50X For All INT_STD	Jaswal	Confirms
24	Q1563-01DL2	437DL2	SAM	03/20/25 18:33	50X for Na	Jaswal	Confirms
25	CCV02	CCV02	CCV	03/20/25 18:36		Jaswal	OK
26	CCB02	CCB02	CCB	03/20/25 18:45		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB135143

Review By	Janvi	Review On	3/24/2025 10:32:11 AM
Supervise By	jaswal	Supervise On	3/24/2025 12:27:15 PM

STD. NAME	STD REF.#
ICAL Standard	MP84682,MP84684,MP84683,MP84688,MP84687,MP84686,MP84685,MP84681,MP84689
ICV Standard	MP84690
CCV Standard	MP84697
ICSA Standard	MP84692,MP84693
CRI Standard	MP84683
LCS Standard	
Chk Standard	MP84695,MP84696

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE	TUNE	03/23/25 14:11		Jaswal	OK
2	S0	S0	CAL1	03/23/25 14:44		Jaswal	OK
3	S2	S2	CAL3	03/23/25 14:50		Jaswal	OK
4	S3	S3	CAL4	03/23/25 14:54		Jaswal	OK
5	S4	S4	CAL5	03/23/25 14:57		Jaswal	OK
6	S5	S5	CAL6	03/23/25 15:00		Jaswal	OK
7	S6	S6	CAL7	03/23/25 15:03		Jaswal	OK
8	S7	S7	CAL8	03/23/25 15:05		Jaswal	OK
9	S8	S8	CAL9	03/23/25 15:08		Jaswal	OK
10	ICV01	ICV01	ICV	03/23/25 15:35		Jaswal	OK
11	LLICV	LLICV	LLICV	03/23/25 15:47		Jaswal	OK
12	ICB01	ICB01	ICB	03/23/25 15:50		Jaswal	OK
13	ICSA01	ICSA01	ICSA	03/23/25 15:54		Jaswal	OK
14	ICSAB01	ICSAB01	ICSAB	03/23/25 15:57		Jaswal	OK
15	CCV01	CCV01	CCV	03/23/25 16:00		Jaswal	OK
16	CCB01	CCB01	CCB	03/23/25 16:04		Jaswal	OK
17	CRI	CRI	CRDL	03/23/25 16:11		Jaswal	OK
18	PB167118BL	PB167118BL	MB	03/23/25 16:18		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB135143

Review By	Janvi	Review On	3/24/2025 10:32:11 AM
Supervise By	jaswal	Supervise On	3/24/2025 12:27:15 PM

STD. NAME	STD REF.#
ICAL Standard	MP84682,MP84684,MP84683,MP84688,MP84687,MP84686,MP84685,MP84681,MP84689
ICV Standard	MP84690
CCV Standard	MP84697
ICSA Standard	MP84692,MP84693
CRI Standard	MP84683
LCS Standard	
Chk Standard	MP84695,MP84696

19	PB167118BS	PB167118BS	LCS	03/23/25 16:21		Jaswal	OK
20	Q1539-01	TAPIAL3-MW03D-031	SAM	03/23/25 16:24		Jaswal	OK
21	Q1539-02	TAPFTA-MW01I-0310	SAM	03/23/25 16:27		Jaswal	OK
22	Q1539-02DUP	TAPFTA-MW01I-0310	DUP	03/23/25 16:30		Jaswal	OK
23	Q1539-02L	TAPFTA-MW01I-0310	SD	03/23/25 16:34		Jaswal	OK
24	Q1539-02MS	TAPFTA-MW01I-0310	MS	03/23/25 16:37		Jaswal	OK
25	Q1539-02MSD	TAPFTA-MW01I-0310	MSD	03/23/25 16:40		Jaswal	OK
26	Q1539-02A	TAPFTA-MW01I-0310	PS	03/23/25 16:43		Jaswal	OK
27	CCV02	CCV02	CCV	03/23/25 16:45		Jaswal	OK
28	CCB02	CCB02	CCB	03/23/25 16:53		Jaswal	OK
29	Q1563-02DL	FERNOT-WATERDL	SAM	03/23/25 16:56	Not use	Jaswal	Not Ok
30	Q1563-02DUPDL	FERNOT-WATERDUP	DUP	03/23/25 16:59	Not use	Jaswal	Not Ok
31	Q1563-02LDL	FERNOT-WATERLDL	SD	03/23/25 17:03	Not use	Jaswal	Not Ok
32	Q1563-02MSDL	FERNOT-WATERMSDL	MS	03/23/25 17:06	Not use	Jaswal	Not Ok
33	Q1563-02MSDDL	FERNOT-WATERMSDDL	MSD	03/23/25 17:09	Not use	Jaswal	Not Ok
34	Q1563-02ADL	FERNOT-WATERADL	PS	03/23/25 17:12	Not use	Jaswal	Not Ok
35	Q1588-01	TAPFTA-MW01D-031	SAM	03/23/25 17:15	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
36	Q1588-01DUP	TAPFTA-MW01D-031	DUP	03/23/25 17:18	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
37	Q1588-01L	TAPFTA-MW01D-031	SD	03/23/25 17:22	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
38	Q1588-01MS	TAPFTA-MW01D-031	MS	03/23/25 17:25	INT_STD 89Y(1&2) Fail	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB135143

Review By	Janvi	Review On	3/24/2025 10:32:11 AM
Supervise By	jaswal	Supervise On	3/24/2025 12:27:15 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84682,MP84684,MP84683,MP84688,MP84687,MP84686,MP84685,MP84681,MP84689		
ICV Standard	MP84690		
CCV Standard	MP84697		
ICSA Standard	MP84692,MP84693		
CRI Standard	MP84683		
LCS Standard			
Chk Standard	MP84695,MP84696		

39	CCV03	CCV03	CCV	03/23/25 17:31		Jaswal	OK
40	CCB03	CCB03	CCB	03/23/25 17:37		Jaswal	OK
41	Q1588-01MSD	TAPFTA-MW01D-031	MSD	03/23/25 17:40	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
42	Q1588-01A	TAPFTA-MW01D-031	PS	03/23/25 17:43	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
43	Q1590-04DL	3799DL	SAM	03/23/25 17:45	K,NA High	Jaswal	Dilution
44	Q1590-04DL2	3799DL2	SAM	03/23/25 17:55	25X For K,NA	Jaswal	Confirms
45	CCV04	CCV04	CCV	03/23/25 18:20		Jaswal	OK
46	CCB04	CCB04	CCB	03/23/25 18:25		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB135164

Review By	JANVI	Review On	3/26/2025 10:54:36 AM
Supervise By	Mohan	Supervise On	3/26/2025 2:18:30 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84682,MP84684,MP84683,MP84688,MP84687,MP84686,MP84685,MP84681,MP84689		
ICV Standard	MP84690		
CCV Standard	MP84697		
ICSA Standard	MP84692,MP84693		
CRI Standard	MP84683		
LCS Standard			
Chk Standard	MP84695,MP84696		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE	TUNE	03/24/25 13:37		Jaswal	OK
2	S0	S0	CAL1	03/24/25 14:08		Jaswal	OK
3	S2	S2	CAL3	03/24/25 14:15		Jaswal	OK
4	S3	S3	CAL4	03/24/25 14:18		Jaswal	OK
5	S4	S4	CAL5	03/24/25 14:21		Jaswal	OK
6	S5	S5	CAL6	03/24/25 14:47		Jaswal	OK
7	S6	S6	CAL7	03/24/25 14:50		Jaswal	OK
8	S7	S7	CAL8	03/24/25 14:53		Jaswal	OK
9	S8	S8	CAL9	03/24/25 14:55		Jaswal	OK
10	ICV01	ICV01	ICV	03/24/25 15:23	Ca Fail	Jaswal	OK
11	LLICV01	LLICV01	LLICV	03/24/25 15:32		Jaswal	OK
12	ICB01	ICB01	ICB	03/24/25 15:36		Jaswal	OK
13	ICSA01	ICSA01	ICSA	03/24/25 15:44		Jaswal	OK
14	ICSAB01	ICSAB01	ICSAB	03/24/25 15:49		Jaswal	OK
15	CCV01	CCV01	CCV	03/24/25 15:55		Jaswal	OK
16	CCB01	CCB01	CCB	03/24/25 15:58		Jaswal	OK
17	CRI	CRI	CRDL	03/24/25 16:01		Jaswal	OK
18	Q1588-01DL	TAPFTA-MW01D-0314	SAM	03/24/25 16:04	5X For INT_STD 89Y(1&2)	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB135164

Review By	JANVI	Review On	3/26/2025 10:54:36 AM
Supervise By	Mohan	Supervise On	3/26/2025 2:18:30 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84682,MP84684,MP84683,MP84688,MP84687,MP84686,MP84685,MP84681,MP84689		
ICV Standard	MP84690		
CCV Standard	MP84697		
ICSA Standard	MP84692,MP84693		
CRI Standard	MP84683		
LCS Standard			
Chk Standard	MP84695,MP84696		

19	Q1588-01DUPDL	TAPFTA-MW01D-031	DUP	03/24/25 16:09	5X For INT_STD 89Y(1&2)	Jaswal	Confirms
20	Q1588-01LDL	TAPFTA-MW01D-031	SD	03/24/25 16:12	25X For INT_STD 89Y(1&2)	Jaswal	Confirms
21	Q1588-01MSDL	TAPFTA-MW01D-031	MS	03/24/25 16:16	5X For INT_STD 89Y(1&2)	Jaswal	Confirms
22	Q1588-01MSDDL	TAPFTA-MW01D-031	MSD	03/24/25 16:19	5X For INT_STD 89Y(1&2)	Jaswal	Confirms
23	Q1588-01ADL	TAPFTA-MW01D-031	PS	03/24/25 16:22	5X For INT_STD 89Y(1&2)	Jaswal	Confirms
24	CCV02	CCV02	CCV	03/24/25 16:25		Jaswal	OK
25	CCB02	CCB02	CCB	03/24/25 16:28		Jaswal	OK