

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions **SDG No.:** Q1588
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1588 **SAS No.:** Q1588

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167152BS							
Aluminum	ug/L	10000	9560		96	84 - 117	P
Antimony	ug/L	500	513		103	85 - 117	P
Arsenic	ug/L	500	501		100	84 - 116	P
Barium	ug/L	2500	2500		100	86 - 114	P
Beryllium	ug/L	500	493		99	83 - 121	P
Cadmium	ug/L	500	506		101	87 - 115	P
Calcium	ug/L	50000	50600		101	87 - 118	P
Chromium	ug/L	500	504		101	85 - 116	P
Cobalt	ug/L	500	509		102	86 - 115	P
Copper	ug/L	5000	5100		102	85 - 118	P
Iron	ug/L	25000	26700		107	87 - 118	P
Lead	ug/L	2500	2490		100	88 - 115	P
Magnesium	ug/L	50000	50000		100	83 - 118	P
Manganese	ug/L	5000	5060		101	87 - 115	P
Nickel	ug/L	500	503		101	85 - 117	P
Potassium	ug/L	25000	25100		100	87 - 115	P
Selenium	ug/L	500	506		101	80 - 120	P
Silver	ug/L	500	509		102	85 - 116	P
Sodium	ug/L	50000	50300		101	85 - 117	P
Thallium	ug/L	500	494		99	82 - 116	P
Vanadium	ug/L	500	467		93	86 - 115	P
Zinc	ug/L	5000	5080		102	83 - 119	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions **SDG No.:** Q1588
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1588 **SAS No.:** Q1588

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167196BS Mercury	ug/L	4.0	3.88		97	82 - 119	CV