

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions

Contract: WEST04

Lab Code: CHEM Case no.: Q1588

Sas No.: Q1588 SDG No.: Q1588

Instrument ID: P7

Start Date : 03/20/2025

Run Number: LB135118

End Date : 03/20/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1548	100		100							
S2	S2	1555	101		100							
S3	S3	1558	105		104							
S4	S4	1604	101		100							
S5	S5	1607	102		102							
S6	S6	1610	104		102							
S7	S7	1613	106		102							
S8	S8	1619	105		95							
ICV01	ICV01	1716	108		104							
LLICV01	LLICV01	1719	106		103							
ICB01	ICB01	1727	104		103							
ICSA01	ICSA01	1740	110		106							
ICSAB01	ICSAB01	1743	108		104							
CCV01	CCV01	1749	112		102							
CCB01	CCB01	1756	111		110							
CRI	CRI	1759	111		112							
PB167152BL	PB167152BL	1802	109		108							
PB167152BS	PB167152BS	1808	110		108							
ZZZZZZ	ZZZZZZ	1812										
ZZZZZZ	ZZZZZZ	1815										
ZZZZZZ	ZZZZZZ	1822										
ZZZZZZ	ZZZZZZ	1829										
ZZZZZZ	ZZZZZZ	1833										
CCV02	CCV02	1836	113		105							
CCB02	CCB02	1845	103		105							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions

Contract: WEST04

Lab Code: CHEM Case no.: Q1588

Sas No.: Q1588 SDG No.: Q1588

Instrument ID: P7

Start Date : 03/20/2025

Run Number: LB135118

End Date : 03/20/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1548	100									
S2	S2	1555	104									
S3	S3	1558	104									
S4	S4	1604	102									
S5	S5	1607	100									
S6	S6	1610	102									
S7	S7	1613	101									
S8	S8	1619	93									
ICV01	ICV01	1716	109									
LLICV01	LLICV01	1719	106									
ICB01	ICB01	1727	106									
ICSA01	ICSA01	1740	105									
ICSAB01	ICSAB01	1743	106									
CCV01	CCV01	1749	102									
CCB01	CCB01	1756	113									
CRI	CRI	1759	111									
PB167152BL	PB167152BL	1802	110									
PB167152BS	PB167152BS	1808	109									
ZZZZZZ	ZZZZZZ	1812										
ZZZZZZ	ZZZZZZ	1815										
ZZZZZZ	ZZZZZZ	1822										
ZZZZZZ	ZZZZZZ	1829										
ZZZZZZ	ZZZZZZ	1833										
CCV02	CCV02	1836	102									
CCB02	CCB02	1845	106									

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions
Lab Code: CHEM Case no.: Q1588
Instrument ID: P7
Run Number: LB135143

Contract: WEST04
Sas No.: Q1588 SDG No.: Q1588
Start Date : 03/23/2025
End Date : 03/23/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1444	100		100							
S2	S2	1450	100		100							
S3	S3	1454	100		102							
S4	S4	1457	99		100							
S5	S5	1500	98		100							
S6	S6	1503	101		102							
S7	S7	1505	101		101							
S8	S8	1508	99		90							
ICV01	ICV01	1535	101		101							
LLICV	LLICV	1547	101		101							
ICB01	ICB01	1550	99		98							
ICSA01	ICSA01	1554	100		95							
ICSAB01	ICSAB01	1557	105		103							
CCV01	CCV01	1600	101		96							
CCB01	CCB01	1604	103		103							
CRI	CRI	1611	100		102							
ZZZZZZ	ZZZZZZ	1618										
ZZZZZZ	ZZZZZZ	1621										
ZZZZZZ	ZZZZZZ	1624										
ZZZZZZ	ZZZZZZ	1627										
ZZZZZZ	ZZZZZZ	1630										
ZZZZZZ	ZZZZZZ	1634										
ZZZZZZ	ZZZZZZ	1637										
ZZZZZZ	ZZZZZZ	1640										
ZZZZZZ	ZZZZZZ	1643										
CCV02	CCV02	1645	103		98							
CCB02	CCB02	1653	102		104							
ZZZZZZ	ZZZZZZ	1656										
Q1563-02DUP	FERNOT-WATEF	1659	105		104							
Q1563-02L	FERNOT-WATEF	1703	105		107							
Q1563-02MS	FERNOT-WATEF	1706	106		105							
Q1563-02MSD	FERNOT-WATEF	1709	105		104							
Q1563-02A	FERNOT-WATEF	1712	102		101							
Q1588-01	TAPFTA-MW01I	1715	113		113							

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions
Lab Code: CHEM Case no.: Q1588
Instrument ID: P7
Run Number: LB135143

Contract: WEST04
Sas No.: Q1588 SDG No.: Q1588
Start Date : 03/23/2025
End Date : 03/23/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
Q1588-01DUP	TAPFTA-MW01I	1718	112		112							
Q1588-01L	TAPFTA-MW01I	1722	106		105							
Q1588-01MS	TAPFTA-MW01I	1725	104		103							
CCV03	CCV03	1731	102		97							
CCB03	CCB03	1737	101		102							
Q1588-01MSD	TAPFTA-MW01I	1740	104		102							
Q1588-01A	TAPFTA-MW01I	1743	106		103							
ZZZZZZ	ZZZZZZ	1745										
ZZZZZZ	ZZZZZZ	1755										
CCV04	CCV04	1820	103		96							
CCB04	CCB04	1825	101		102							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions
Lab Code: CHEM Case no.: Q1588
Instrument ID: P7
Run Number: LB135143

Contract: WEST04
Sas No.: Q1588 SDG No.: Q1588
Start Date : 03/23/2025
End Date : 03/23/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1444	100									
S2	S2	1450	100									
S3	S3	1454	101									
S4	S4	1457	100									
S5	S5	1500	100									
S6	S6	1503	100									
S7	S7	1505	97									
S8	S8	1508	89									
ICV01	ICV01	1535	101									
LLICV	LLICV	1547	102									
ICB01	ICB01	1550	99									
ICSA01	ICSA01	1554	96									
ICSAB01	ICSAB01	1557	100									
CCV01	CCV01	1600	93									
CCB01	CCB01	1604	104									
CRI	CRI	1611	101									
ZZZZZZ	ZZZZZZ	1618										
ZZZZZZ	ZZZZZZ	1621										
ZZZZZZ	ZZZZZZ	1624										
ZZZZZZ	ZZZZZZ	1627										
ZZZZZZ	ZZZZZZ	1630										
ZZZZZZ	ZZZZZZ	1634										
ZZZZZZ	ZZZZZZ	1637										
ZZZZZZ	ZZZZZZ	1640										
ZZZZZZ	ZZZZZZ	1643										
CCV02	CCV02	1645	96									
CCB02	CCB02	1653	103									
ZZZZZZ	ZZZZZZ	1656										
Q1563-02DUP	FERNOT-WATEF	1659	103									
Q1563-02L	FERNOT-WATEF	1703	107									
Q1563-02MS	FERNOT-WATEF	1706	104									
Q1563-02MSD	FERNOT-WATEF	1709	104									
Q1563-02A	FERNOT-WATEF	1712	101									
Q1588-01	TAPFTA-MW01I	1715	109									

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions
Lab Code: CHEM Case no.: Q1588
Instrument ID: P7
Run Number: LB135143

Contract: WEST04
Sas No.: Q1588 SDG No.: Q1588
Start Date : 03/23/2025
End Date : 03/23/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
Q1588-01DUP	TAPFTA-MW01I	1718	112									
Q1588-01L	TAPFTA-MW01I	1722	104									
Q1588-01MS	TAPFTA-MW01I	1725	102									
CCV03	CCV03	1731	94									
CCB03	CCB03	1737	100									
Q1588-01MSD	TAPFTA-MW01I	1740	102									
Q1588-01A	TAPFTA-MW01I	1743	104									
ZZZZZZ	ZZZZZZ	1745										
ZZZZZZ	ZZZZZZ	1755										
CCV04	CCV04	1820	95									
CCB04	CCB04	1825	104									

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions

Contract: WEST04

Lab Code: CHEM Case no.: Q1588

Sas No.: Q1588 SDG No.: Q1588

Instrument ID: P7

Start Date : 03/24/2025

Run Number: LB135164

End Date : 03/24/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1408	100		100							
S2	S2	1415	100		101							
S3	S3	1418	99		99							
S4	S4	1421	98		100							
S5	S5	1447	101		103							
S6	S6	1450	99		100							
S7	S7	1453	99		98							
S8	S8	1455	98		88							
ICV01	ICV01	1523	99		101							
LLICV01	LLICV01	1532	97		98							
ICB01	ICB01	1536	97		98							
ICSA01	ICSA01	1544	100		95							
ICSAB01	ICSAB01	1549	102		98							
CCV01	CCV01	1555	98		94							
CCB01	CCB01	1558	99		100							
CRI	CRI	1601	100		101							
Q1588-01	TAPFTA-MW01I	1604	103		102							
Q1588-01DUP	TAPFTA-MW01I	1609	109		109							
Q1588-01L	TAPFTA-MW01I	1612	98		97							
Q1588-01MS	TAPFTA-MW01I	1616	98		99							
Q1588-01MSD	TAPFTA-MW01I	1619	97		98							
Q1588-01A	TAPFTA-MW01I	1622	99		100							
CCV02	CCV02	1625	99		94							
CCB02	CCB02	1628	100		101							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions
Lab Code: CHEM Case no.: Q1588
Instrument ID: P7
Run Number: LB135164

Contract: WEST04
Sas No.: Q1588 SDG No.: Q1588
Start Date : 03/24/2025
End Date : 03/24/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1408	100									
S2	S2	1415	100									
S3	S3	1418	96									
S4	S4	1421	95									
S5	S5	1447	101									
S6	S6	1450	97									
S7	S7	1453	94									
S8	S8	1455	85									
ICV01	ICV01	1523	101									
LLICV01	LLICV01	1532	98									
ICB01	ICB01	1536	96									
ICSA01	ICSA01	1544	96									
ICSAB01	ICSAB01	1549	96									
CCV01	CCV01	1555	89									
CCB01	CCB01	1558	97									
CRI	CRI	1601	100									
Q1588-01	TAPFTA-MW01I	1604	101									
Q1588-01DUP	TAPFTA-MW01I	1609	104									
Q1588-01L	TAPFTA-MW01I	1612	98									
Q1588-01MS	TAPFTA-MW01I	1616	95									
Q1588-01MSD	TAPFTA-MW01I	1619	96									
Q1588-01A	TAPFTA-MW01I	1622	97									
CCV02	CCV02	1625	90									
CCB02	CCB02	1628	98									