



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Elegant Jewelers Mfg. Co. Inc.

Project Name: Waste Water 2025

Project # N/A

Chemtech Project # Q1595

Test Name: Cyanide,Cyanide-Amenable

A. Number of Samples and Date of Receipt:

2 Water samples were received on 03/19/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Cyanide, Cyanide-Amenable, Mercury, Metals Group1 and Metals ICP-Group1. This data package contains results for Cyanide,Cyanide-Amenable.

C. Analytical Techniques:

The analysis of Cyanide-Amenable was based on method SM4500-CN B,G Cyanide-Amenable and The analysis of Cyanide was based on method SM4500-CN C,E.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

Samples were received on 03/19/2025, 10:50 and composited in the Lab on 03/19/2025, 11:40.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____