

Initial and Continuing Calibration Verification

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q1595

Project: Waste Water 2025

RunNo.: LB135127

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.088	0.099	89	85-115	03/21/2025
Sample ID: Cyanide	CCV1	mg/L	0.23	0.25	92	90-110	03/21/2025
Sample ID: Cyanide	CCV2	mg/L	0.25	0.25	100	90-110	03/21/2025

Initial and Continuing Calibration Verification

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q1595

Project: Waste Water 2025

RunNo.: LB135127

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------