

### Initial and Continuing Calibration Blank Summary

**Client:** Elegant Jewelers Mfg. Co. Inc.

**SDG No.:** Q1595

**Project:** Waste Water 2025

**RunNo.:** LB135127

| Analyte                                  | Units | Result   | Acceptance Limits | Conc Qual | MDL    | RDL   | Analysis Date |
|------------------------------------------|-------|----------|-------------------|-----------|--------|-------|---------------|
| Sample ID: <b>ICB1</b><br><b>Cyanide</b> | mg/L  | < 0.0025 | 0.0025            | U         | 0.0012 | 0.005 | 03/21/2025    |
| Sample ID: <b>CCB1</b><br><b>Cyanide</b> | mg/L  | < 0.0025 | 0.0025            | U         | 0.0012 | 0.005 | 03/21/2025    |
| Sample ID: <b>CCB2</b><br><b>Cyanide</b> | mg/L  | < 0.0025 | 0.0025            | U         | 0.0012 | 0.005 | 03/21/2025    |

### Initial and Continuing Calibration Blank Summary

**Client:** Elegant Jewelers Mfg. Co. Inc.

**SDG No.:** Q1595

**Project:** Waste Water 2025

**RunNo.:** LB135127

| Analyte | Units | Result | Acceptance<br>Limits | Conc<br>Qual | MDL | RDL | Analysis<br>Date |
|---------|-------|--------|----------------------|--------------|-----|-----|------------------|
|---------|-------|--------|----------------------|--------------|-----|-----|------------------|