



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Elegant Jewelers Mfg. Co. Inc.</u>				SDG No.:	<u>Q1595</u>			
Contract:	<u>ELEG01</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1595</u>	SAS No.:	<u>Q1595</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB60	Mercury	0.20	+/-0.20	U	0.20	CV	03/25/2025	11:38	LB135174

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q1595

Contract: ELEG01

Lab Code: CHEM

Case No.: Q1595

SAS No.: Q1595

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB57	Mercury	0.20	+/-0.20	U	0.20	CV	03/25/2025	11:43	LB135174
CCB58	Mercury	0.20	+/-0.20	U	0.20	CV	03/25/2025	12:16	LB135174
CCB59	Mercury	0.20	+/-0.20	U	0.20	CV	03/25/2025	12:40	LB135174

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q1595

Contract: ELEG01

Lab Code: CHEM

Case No.: Q1595

SAS No.: Q1595

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	10:49	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	10:49	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	10:49	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	10:49	LB135158
	Molybdenum	200	+/-200	U	200	P	03/24/2025	10:49	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	10:49	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	10:49	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	10:49	LB135158



- 3a -

Client:	Elegant Jewelers Mfg. Co. Inc.				SDG No.:	Q1595			
Contract:	ELEG01	Lab Code:	CHEM	Case No.:	Q1595	SAS No.: Q1595			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	11:26	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	11:26	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	11:26	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	11:26	LB135158
	Molybdenum	200	+/-200	U	200	P	03/24/2025	11:26	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	11:26	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	11:26	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	11:26	LB135158
CCB02	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	12:17	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	12:17	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	12:17	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	12:17	LB135158
	Molybdenum	200	+/-200	U	200	P	03/24/2025	12:17	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	12:17	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	12:17	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	12:17	LB135158
CCB03	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	13:11	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	13:11	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	13:11	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	13:11	LB135158
	Molybdenum	200	+/-200	U	200	P	03/24/2025	13:11	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	13:11	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	13:11	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	13:11	LB135158
CCB04	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	14:11	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	14:11	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	14:11	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	14:11	LB135158
	Molybdenum	200	+/-200	U	200	P	03/24/2025	14:11	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	14:11	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	14:11	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	14:11	LB135158
CCB05	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	15:34	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	15:34	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	15:34	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	15:34	LB135158
	Molybdenum	200	+/-200	U	200	P	03/24/2025	15:34	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	15:34	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	15:34	LB135158

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Elegant Jewelers Mfg. Co. Inc.</u>		SDG No.: <u>Q1595</u>							
Contract: <u>ELEG01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1595</u>	SAS No.: <u>Q1595</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	15:34	LB135158
CCB06	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	16:43	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	16:43	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	16:43	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	16:43	LB135158
	Molybdenum	200	+/-200	U	200	P	03/24/2025	16:43	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	16:43	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	16:43	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	16:43	LB135158
CCB07	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	17:33	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	17:33	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	17:33	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	17:33	LB135158
	Molybdenum	200	+/-200	U	200	P	03/24/2025	17:33	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	17:33	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	17:33	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	17:33	LB135158
CCB08	Cadmium	6.00	+/-6.00	U	6.00	P	03/24/2025	18:08	LB135158
	Chromium	10.0	+/-10.0	U	10.0	P	03/24/2025	18:08	LB135158
	Copper	20.0	+/-20.0	U	20.0	P	03/24/2025	18:08	LB135158
	Lead	12.0	+/-12.0	U	12.0	P	03/24/2025	18:08	LB135158
	Molybdenum	200	+/-200	U	200	P	03/24/2025	18:08	LB135158
	Nickel	40.0	+/-40.0	U	40.0	P	03/24/2025	18:08	LB135158
	Silver	10.0	+/-10.0	U	10.0	P	03/24/2025	18:08	LB135158
	Zinc	40.0	+/-40.0	U	40.0	P	03/24/2025	18:08	LB135158

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q1595

Contract: ELEG01

Lab Code: CHEM

Case No.: Q1595

SAS No.: Q1595

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------