

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q1595

Contract: ELEG01

Lab Code: CHEM

Case No.: Q1595

SAS No.: Q1595

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167266BS							
Cadmium	ug/L	100	88.3		88	85 - 115	P
Chromium	ug/L	200	191		96	85 - 115	P
Copper	ug/L	150	150		100	85 - 115	P
Lead	ug/L	500	445		89	85 - 115	P
Molybdenum	ug/L	200	200		100	85 - 115	P
Nickel	ug/L	250	227		91	85 - 115	P
Silver	ug/L	37.5	36.4		97	85 - 115	P
Zinc	ug/L	100	98.5		98	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Elegant Jewelers Mfg. Co. Inc.

SDG No.: Q1595

Contract: ELEG01

Lab Code: CHEM

Case No.: Q1595

SAS No.: Q1595

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167308BS Mercury	ug/L	4.0	3.99		100	85 - 115	CV