

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: ATCE02

Lab Code: CHEM **Case No.:** Q1596 **SAS No.:** Q1596 **SDG NO.:** Q1596

Instrument ID: ECD_P **Date(s) Analyzed:** 03/11/2025 03/11/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.73	4.63	4.83	19940100
		2	4.81	4.71	4.91	14608700
		3	4.89	4.79	4.99	44364600
		4	0.00			0
		5	0.00			0
Aroclor-1262	500	1	8.11	8.01	8.21	81718600
		2	8.43	8.33	8.53	164714000
		3	8.75	8.65	8.85	112164000
		4	8.83	8.73	8.93	83803800
		5	9.49	9.39	9.59	58896200

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: ATCE02

Lab Code: CHEM **Case No.:** Q1596 **SAS No.:** Q1596 **SDG NO.:** Q1596

Instrument ID: ECD_P **Date(s) Analyzed:** 03/11/2025 03/11/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.04	3.94	4.14	14882500
		2	4.13	4.03	4.23	11429100
		3	4.20	4.10	4.30	33421800
		4	0.00			0
		5	0.00			0
Aroclor-1262	500	1	6.96	6.86	7.06	83664000
		2	7.22	7.12	7.32	68729800
		3	7.74	7.64	7.84	65986200
		4	7.81	7.71	7.91	104435000
		5	8.31	8.21	8.41	51550000