

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1596 SAS No.: Q1596 SDG NO.: Q1596

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/19/2025

Lab Sample No.: AR1660CCC500 Data File : PP070712.D Time Analyzed: 19:41

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.671	5.577	5.777	467.020	500.000	-6.6
Aroclor-1016-2	5.693	5.599	5.799	515.660	500.000	3.1
Aroclor-1016-3	5.755	5.661	5.861	484.090	500.000	-3.2
Aroclor-1016-4	5.853	5.758	5.958	479.530	500.000	-4.1
Aroclor-1016-5	6.145	6.051	6.251	478.290	500.000	-4.3
Aroclor-1260-1	7.264	7.171	7.371	586.340	500.000	17.3
Aroclor-1260-2	7.517	7.425	7.625	555.220	500.000	11.0
Aroclor-1260-3	7.875	7.783	7.983	546.040	500.000	9.2
Aroclor-1260-4	8.100	8.008	8.208	510.970	500.000	2.2
Aroclor-1260-5	8.420	8.328	8.528	518.000	500.000	3.6
Decachlorobiphenyl	10.237	10.152	10.352	50.150	50.000	0.3
Tetrachloro-m-xylene	4.520	4.424	4.624	49.290	50.000	-1.4

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1596 SAS No.: Q1596 SDG NO.: Q1596

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/19/2025

Lab Sample No.: AR1660CCC500 Data File : PP070712.D Time Analyzed: 19:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.909	4.817	5.017	491.650	500.000	-1.7
Aroclor-1016-2	4.927	4.836	5.036	480.150	500.000	-4.0
Aroclor-1016-3	5.105	5.013	5.213	498.120	500.000	-0.4
Aroclor-1016-4	5.146	5.054	5.254	508.430	500.000	1.7
Aroclor-1016-5	5.361	5.270	5.470	508.520	500.000	1.7
Aroclor-1260-1	6.397	6.307	6.507	494.760	500.000	-1.0
Aroclor-1260-2	6.585	6.495	6.695	481.760	500.000	-3.6
Aroclor-1260-3	6.739	6.649	6.849	475.800	500.000	-4.8
Aroclor-1260-4	7.211	7.121	7.321	454.430	500.000	-9.1
Aroclor-1260-5	7.451	7.362	7.562	450.700	500.000	-9.9
Decachlorobiphenyl	8.867	8.781	8.981	47.340	50.000	-5.3
Tetrachloro-m-xylene	3.823	3.729	3.929	46.490	50.000	-7.0