

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1596 SAS No.: Q1596 SDG NO.: Q1596

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP070729.D Time Analyzed: 07:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.673	5.577	5.777	481.450	500.000	-3.7
Aroclor-1016-2	5.694	5.599	5.799	528.830	500.000	5.8
Aroclor-1016-3	5.757	5.661	5.861	490.570	500.000	-1.9
Aroclor-1016-4	5.854	5.758	5.958	478.450	500.000	-4.3
Aroclor-1016-5	6.147	6.051	6.251	505.320	500.000	1.1
Aroclor-1260-1	7.265	7.171	7.371	601.640	500.000	20.3
Aroclor-1260-2	7.519	7.425	7.625	596.150	500.000	19.2
Aroclor-1260-3	7.878	7.783	7.983	587.660	500.000	17.5
Aroclor-1260-4	8.102	8.008	8.208	547.610	500.000	9.5
Aroclor-1260-5	8.422	8.328	8.528	542.300	500.000	8.5
Decachlorobiphenyl	10.240	10.152	10.352	52.930	50.000	5.9
Tetrachloro-m-xylene	4.521	4.424	4.624	50.850	50.000	1.7

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1596 SAS No.: Q1596 SDG NO.: Q1596

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP070729.D Time Analyzed: 07:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.909	4.817	5.017	510.230	500.000	2.0
Aroclor-1016-2	4.928	4.836	5.036	498.960	500.000	-0.2
Aroclor-1016-3	5.105	5.013	5.213	509.350	500.000	1.9
Aroclor-1016-4	5.146	5.054	5.254	505.150	500.000	1.0
Aroclor-1016-5	5.361	5.270	5.470	505.250	500.000	1.1
Aroclor-1260-1	6.397	6.307	6.507	507.590	500.000	1.5
Aroclor-1260-2	6.586	6.495	6.695	497.080	500.000	-0.6
Aroclor-1260-3	6.739	6.649	6.849	489.760	500.000	-2.0
Aroclor-1260-4	7.210	7.121	7.321	479.880	500.000	-4.0
Aroclor-1260-5	7.451	7.362	7.562	483.070	500.000	-3.4
Decachlorobiphenyl	8.867	8.781	8.981	49.320	50.000	-1.4
Tetrachloro-m-xylene	3.823	3.729	3.929	50.740	50.000	1.5